

## ANALYTICAL REPORT

Eurofins TestAmerica, Canton  
4101 Shuffel Street NW  
North Canton, OH 44720  
Tel: (330)497-9396

Laboratory Job ID: 240-114396-1

Client Project/Site: Ford LTP Livonia MI - E203631  
Revision: 1

For:  
ARCADIS U.S., Inc.  
28550 Cabot Drive  
Suite 500  
Novi, Michigan 48377

Attn: Kristoffer Hinskey



---

Authorized for release by:  
8/8/2019 9:21:41 AM

Michael DelMonico, Project Manager I  
(330)497-9396  
[michael.delmonico@testamericainc.com](mailto:michael.delmonico@testamericainc.com)

### LINKS

Review your project  
results through  
**TotalAccess**

Have a Question?



Visit us at:  
[www.testamericainc.com](http://www.testamericainc.com)

*This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.*

*Results relate only to the items tested and the sample(s) as received by the laboratory.*



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Method Summary . . . . .	5
Sample Summary . . . . .	6
Detection Summary . . . . .	7
Client Sample Results . . . . .	8
Surrogate Summary . . . . .	10
QC Sample Results . . . . .	11
QC Association Summary . . . . .	14
Lab Chronicle . . . . .	15
Certification Summary . . . . .	16
Chain of Custody . . . . .	17

# Definitions/Glossary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## Qualifiers

### GC/MS VOA

Qualifier	Qualifier Description
F2	MS/MSD RPD exceeds control limits
U	Indicates the analyte was analyzed for but not detected.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
▫	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

# Case Narrative

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

**Job ID: 240-114396-1**

**Laboratory: Eurofins TestAmerica, Canton**

**Narrative**

## CASE NARRATIVE

**Client: ARCADIS U.S., Inc.**

**Project: Ford LTP Livonia MI - E203631**

**Report Number: 240-114396-1**

### Revision

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

Report revised 8/8/2019 to correct ID for sample 1.

Eurofins TestAmerica, Canton attests to the validity of the laboratory data generated by Eurofins TestAmerica facilities reported herein. All analyses performed by Eurofins TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the application methods. Eurofins TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

The test results in this report meet all NELAP requirements for parameters for which accreditation is required or available. Any exceptions to NELAP requirements are noted in this report. Pursuant to NELAP, this report may not be reproduced, except in full, without the written approval of the laboratory.

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of Eurofins TestAmerica and its client.

#### **RECEIPT**

The samples were received on 6/14/2019 8:15 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperatures of the 2 coolers at receipt time were 1.0° C and 1.9° C.

#### **VOLATILE ORGANIC COMPOUNDS (GCMS)**

Samples SUMP-34380CAPITOL-01\_061219 (240-114396-1) and TRIP BLANK (240-114396-2) were analyzed for volatile organic compounds (GCMS) in accordance with EPA SW-846 Method 8260B. The samples were analyzed on 06/24/2019.

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

#### **VOLATILE ORGANIC COMPOUNDS (GCMS SIM)**

Sample SUMP-34380CAPITOL-01\_061219 (240-114396-1) was analyzed for volatile organic compounds (GCMS SIM) in accordance with EPA SW-846 Method 8260B SIM. The sample was analyzed on 06/18/2019.

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

# Method Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
8260B SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
5030B	Purge and Trap	SW846	TAL CAN

**Protocol References:**

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

**Laboratory References:**

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396



# Sample Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

---

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Asset ID
240-114396-1	SUMP-34380CAPITOL-01_061219	Water	06/12/19 18:15	06/14/19 08:15	
240-114396-2	TRIP BLANK	Water	06/12/19 00:00	06/14/19 08:15	

---

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

# Detection Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

**Client Sample ID: SUMP-34380CAPITOL-01\_061219**

**Lab Sample ID: 240-114396-1**

No Detections.

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-114396-2**

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

This Detection Summary does not include radiochemical test results.

Eurofins TestAmerica, Canton

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

**Client Sample ID: SUMP-34380CAPITOL-01\_061219**

**Lab Sample ID: 240-114396-1**

**Date Collected: 06/12/19 18:15**

**Matrix: Water**

**Date Received: 06/14/19 08:15**

**Method: 8260B SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			06/18/19 15:31	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	108		63 - 125		06/18/19 15:31	1

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:30	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/24/19 20:30	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/24/19 20:30	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:30	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/24/19 20:30	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/24/19 20:30	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	118		70 - 121		06/24/19 20:30	1
4-Bromofluorobenzene (Surr)	101		59 - 120		06/24/19 20:30	1
Toluene-d8 (Surr)	95		70 - 123		06/24/19 20:30	1
Dibromofluoromethane (Surr)	114		75 - 128		06/24/19 20:30	1



# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-114396-2**

**Date Collected: 06/12/19 00:00**

**Matrix: Water**

**Date Received: 06/14/19 08:15**

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:55	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/24/19 20:55	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/24/19 20:55	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:55	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/24/19 20:55	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/24/19 20:55	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	118		70 - 121		06/24/19 20:55	1
4-Bromofluorobenzene (Surr)	89		59 - 120		06/24/19 20:55	1
Toluene-d8 (Surr)	91		70 - 123		06/24/19 20:55	1
Dibromofluoromethane (Surr)	123		75 - 128		06/24/19 20:55	1

# Surrogate Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA (70-121)	BFB (59-120)	TOL (70-123)	DBFM (75-128)
240-114396-1	SUMP-34380CAPITOL-01_0612	118	101	95	114
240-114396-2	TRIP BLANK	118	89	91	123
240-114521-K-3 MS	Matrix Spike	111	92	97	109
240-114521-L-3 MSD	Matrix Spike Duplicate	113	104	90	113
LCS 240-387956/4	Lab Control Sample	106	92	87	106
MB 240-387956/7	Method Blank	110	85	88	115

#### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)  
BFB = 4-Bromofluorobenzene (Surr)  
TOL = Toluene-d8 (Surr)  
DBFM = Dibromofluoromethane (Surr)

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA (63-125)
240-114396-1	SUMP-34380CAPITOL-01_0612	108
240-114490-A-4 MS	Matrix Spike	108
240-114490-A-4 MSD	Matrix Spike Duplicate	110
LCS 240-386776/4	Lab Control Sample	105
MB 240-386776/5	Method Blank	109

#### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

# QC Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 240-387956/7**  
**Matrix: Water**  
**Analysis Batch: 387956**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 18:14	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/24/19 18:14	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/24/19 18:14	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 18:14	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/24/19 18:14	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/24/19 18:14	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	110		70 - 121		06/24/19 18:14	1
4-Bromofluorobenzene (Surr)	85		59 - 120		06/24/19 18:14	1
Toluene-d8 (Surr)	88		70 - 123		06/24/19 18:14	1
Dibromofluoromethane (Surr)	115		75 - 128		06/24/19 18:14	1

**Lab Sample ID: LCS 240-387956/4**  
**Matrix: Water**  
**Analysis Batch: 387956**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	10.0	11.0		ug/L		110	65 - 139
cis-1,2-Dichloroethene	10.0	9.86		ug/L		99	76 - 128
Tetrachloroethene	10.0	11.9		ug/L		119	74 - 130
trans-1,2-Dichloroethene	10.0	9.51		ug/L		95	78 - 133
Trichloroethene	10.0	12.2		ug/L		122	76 - 125
Vinyl chloride	10.0	9.63		ug/L		96	58 - 143

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	106		70 - 121
4-Bromofluorobenzene (Surr)	92		59 - 120
Toluene-d8 (Surr)	87		70 - 123
Dibromofluoromethane (Surr)	106		75 - 128

**Lab Sample ID: 240-114521-K-3 MS**  
**Matrix: Water**  
**Analysis Batch: 387956**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	1.0	U	10.0	8.58		ug/L		86	53 - 140
cis-1,2-Dichloroethene	1.0	U	10.0	8.32		ug/L		83	64 - 130
Tetrachloroethene	1.0	U	10.0	10.5		ug/L		105	51 - 136
trans-1,2-Dichloroethene	1.0	U	10.0	8.48		ug/L		85	68 - 133
Trichloroethene	1.0	U F2	10.0	8.82		ug/L		88	55 - 131
Vinyl chloride	0.67	J	10.0	9.69		ug/L		90	43 - 154

Surrogate	MS %Recovery	MS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	111		70 - 121
4-Bromofluorobenzene (Surr)	92		59 - 120
Toluene-d8 (Surr)	97		70 - 123

Eurofins TestAmerica, Canton

# QC Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 240-114521-K-3 MS**  
**Matrix: Water**  
**Analysis Batch: 387956**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

Surrogate	MS %Recovery	MS Qualifier	Limits
Dibromofluoromethane (Surr)	109		75 - 128

**Lab Sample ID: 240-114521-L-3 MSD**  
**Matrix: Water**  
**Analysis Batch: 387956**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,1-Dichloroethene	1.0	U	10.0	9.34		ug/L		93	53 - 140	8	35
cis-1,2-Dichloroethene	1.0	U	10.0	9.01		ug/L		90	64 - 130	8	21
Tetrachloroethene	1.0	U	10.0	9.55		ug/L		96	51 - 136	9	23
trans-1,2-Dichloroethene	1.0	U	10.0	8.87		ug/L		89	68 - 133	4	24
Trichloroethene	1.0	U F2	10.0	11.4	F2	ug/L		114	55 - 131	26	23
Vinyl chloride	0.67	J	10.0	9.27		ug/L		86	43 - 154	4	29

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	113		70 - 121
4-Bromofluorobenzene (Surr)	104		59 - 120
Toluene-d8 (Surr)	90		70 - 123
Dibromofluoromethane (Surr)	113		75 - 128

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 240-386776/5**  
**Matrix: Water**  
**Analysis Batch: 386776**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			06/18/19 12:37	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	109		63 - 125		06/18/19 12:37	1

**Lab Sample ID: LCS 240-386776/4**  
**Matrix: Water**  
**Analysis Batch: 386776**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	10.0	10.9		ug/L		109	59 - 131

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	105		63 - 125

**Lab Sample ID: 240-114490-A-4 MS**  
**Matrix: Water**  
**Analysis Batch: 386776**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	2.0	U	10.0	10.2		ug/L		102	52 - 129

Eurofins TestAmerica, Canton

# QC Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS) (Continued)

<i>Surrogate</i>	<i>MS</i> <i>%Recovery</i>	<i>MS</i> <i>Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	108		63 - 125

Lab Sample ID: 240-114490-A-4 MSD  
 Matrix: Water  
 Analysis Batch: 386776

Client Sample ID: Matrix Spike Duplicate  
 Prep Type: Total/NA

<i>Analyte</i>	<i>Sample</i> <i>Result</i>	<i>Sample</i> <i>Qualifier</i>	<i>Spike</i> <i>Added</i>	<i>MSD</i> <i>Result</i>	<i>MSD</i> <i>Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec.</i> <i>Limits</i>	<i>RPD</i>	<i>RPD</i> <i>Limit</i>
1,4-Dioxane	2.0	U	10.0	10.6		ug/L		106	52 - 129	4	13

<i>Surrogate</i>	<i>MSD</i> <i>%Recovery</i>	<i>MSD</i> <i>Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	110		63 - 125

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

# QC Association Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## GC/MS VOA

### Analysis Batch: 386776

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-114396-1	SUMP-34380CAPITOL-01_061219	Total/NA	Water	8260B SIM	
MB 240-386776/5	Method Blank	Total/NA	Water	8260B SIM	
LCS 240-386776/4	Lab Control Sample	Total/NA	Water	8260B SIM	
240-114490-A-4 MS	Matrix Spike	Total/NA	Water	8260B SIM	
240-114490-A-4 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B SIM	

### Analysis Batch: 387956

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-114396-1	SUMP-34380CAPITOL-01_061219	Total/NA	Water	8260B	
240-114396-2	TRIP BLANK	Total/NA	Water	8260B	
MB 240-387956/7	Method Blank	Total/NA	Water	8260B	
LCS 240-387956/4	Lab Control Sample	Total/NA	Water	8260B	
240-114521-K-3 MS	Matrix Spike	Total/NA	Water	8260B	
240-114521-L-3 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B	

# Lab Chronicle

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

**Client Sample ID: SUMP-34380CAPITOL-01\_061219**

**Lab Sample ID: 240-114396-1**

**Date Collected: 06/12/19 18:15**

**Matrix: Water**

**Date Received: 06/14/19 08:15**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	387956	06/24/19 20:30	LRW	TAL CAN
Total/NA	Analysis	8260B SIM		1	386776	06/18/19 15:31	SAM	TAL CAN

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-114396-2**

**Date Collected: 06/12/19 00:00**

**Matrix: Water**

**Date Received: 06/14/19 08:15**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	387956	06/24/19 20:55	LRW	TAL CAN

**Laboratory References:**

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396

# Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

## Laboratory: Eurofins TestAmerica, Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	EPA Region	Identification Number	Expiration Date
California	State		2927	02-23-20
California	State Program	9	2927	02-23-20
Connecticut	State		PH-0590	12-31-19
Connecticut	State Program	1	PH-0590	12-31-19
Florida	NELAP	4	E87225	06-30-20
Illinois	NELAP	5	200004	07-31-20
Illinois	NELAP		004498	07-31-19
Iowa	State Program	7	421	06-01-21
Kansas	NELAP	7	E-10336	04-30-20
Kansas	NELAP		E-10336	04-30-20
Kentucky (UST)	State Program	4	58	02-23-20
Kentucky (WW)	State		KY98016	12-31-19
Kentucky (WW)	State Program	4	98016	12-31-19
Minnesota	NELAP	5	039-999-348	12-31-19 *
Minnesota	NELAP		OH00048	12-31-19
Minnesota (Petrofund)	State Program	1	3506	07-31-21
Nevada	State Program	9	OH00048	07-31-19
New Jersey	NELAP	2	OH001	06-30-20
New York	NELAP	2	10975	03-31-20
New York	NELAP		10975	03-31-20
Ohio VAP	State		CL0024	06-05-21
Ohio VAP	State Program	5	CL0024	06-05-21
Oregon	NELAP	10	4062	02-23-20
Oregon	NELAP		4062	02-23-20
Pennsylvania	NELAP	3	68-00340	08-31-19 *
Pennsylvania	NELAP		68-00340	08-31-19
Texas	NELAP	6	T104704517-19-11	08-31-20
Texas	NELAP		T104704517-18-10	08-31-19
USDA	Federal		P330-16-00404	12-28-19
Virginia	NELAP	3	460175	09-14-19 *
Virginia	NELAP		010101	09-14-19
Washington	State		C971	01-12-20
Washington	State Program	10	C971	01-12-20 *
West Virginia DEP	State		210	12-31-19
West Virginia DEP	State Program	3	210	12-31-19

\* Accreditation/Certification renewal pending - accreditation/certification considered valid.



# MICHIGAN 190

## Chain of Custody Record



THE LEADER IN ENVIRONMENTAL TESTING

1.8/1.9

TestAmerica Laboratory location: N.Canton — 4101 Shuffel Street NW/ North Canton, OH 44720 / 330-497-9396

<b>Client Contact</b> Company Name: Arcadis Address: 28550 Cabot Drive, Suite 500 City/State/Zip: Novi, MI, 48377 Phone: 248-994-2240 Project Name: Ford LTP Project Number: MI001454.0003 PO # MI001454.0003		<b>Regulatory program:</b> <input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other	
<b>Client Project Manager:</b> Kris Hinskey Telephone: 248-994-2240 Email: kristoffer.hinskey@arcadis.com		<b>Site Contact:</b> Angela DeGrandis Telephone: 734-320-0065	
<b>Lab Contact:</b> Mike DelMonico Telephone: 330-497-9396		TestAmerica Laboratories, Inc. COC No:	
Shipping/Tracking No: 31380 CAPITOL		For lab use only	
Method of Shipment/Carrier:		Walk-in client Lab sampling	
Job/SDG No:		Job/SDG No:	
Sample Identification: SUMP-31380-CAPITOL-01-06019 TRIP BLANK		Sample Specific Notes / Special Instructions: 6 CONTAINERS	
Sample Date: 6/13/19 Sample Time: 1815		Filtered Sample (Y/N): N	
Matrix: Aqueous, Sediment, Solid, Other:		Composite=C / Grab=G: N	
Containers & Preservatives: H2SO4, HNO3, HCl, NaOH, ZnAc, NaOH, Upret, Other:		1,1-DCE 8260B: X cis-1,2-DCE 8260B: X Trans-1,2-DCE 8260B: X PCE 8260B: X TCE 8260B: X Vinyl Chloride 8260B: X 1,4-Dioxane 8260B SIM: X	
Analysis Turnaround Time: 5 Day <input type="checkbox"/> 1 week <input checked="" type="checkbox"/> 2 days <input type="checkbox"/> 1 day <input type="checkbox"/>		TAT if different from below: <input type="checkbox"/> 3 weeks <input type="checkbox"/> 2 weeks <input type="checkbox"/> 1 week <input checked="" type="checkbox"/> 2 days <input type="checkbox"/> 1 day <input type="checkbox"/>	
Possible Hazard Identification: <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Irritable <input type="checkbox"/> In Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month): <input type="checkbox"/> Return to Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months	
Special Instructions/QC Requirements & Comments:			
Submit all results through Cadena at jim.tomalia@cadenalab.com. Cadena #E203531 Level IV Reporting.			
Requisitioned by: [Signature] / Arcadis Date/Time: 6-12-19 1930 Company: Arcadis	Received by: [Signature] Date/Time: 6-13-19 09:28 Company: ARCADIS	Received by: [Signature] Date/Time: 6-13-19 15:05 Company: ETC	Received by: [Signature] Date/Time: 6-13-19 08:51 Company: ETC
Requisitioned by: [Signature] / Arcadis Date/Time: 6-12-19 1930 Company: Arcadis	Received by: [Signature] Date/Time: 6-13-19 09:28 Company: ARCADIS	Received by: [Signature] Date/Time: 6-13-19 15:05 Company: ETC	Received by: [Signature] Date/Time: 6-13-19 08:51 Company: ETC

©2008 TestAmerica Laboratories, Inc. All rights reserved. TestAmerica & Design are trademarks of TestAmerica Laboratories, Inc.



**Eurofins TestAmerica Canton Sample Receipt Form/Narrative  
Canton Facility**

Login # : 114396

Client Areadis Site Name \_\_\_\_\_ Cooler unpacked by: [Signature]  
 Cooler Received on 6-14-19 Opened on 6-19-19

FedEx: 1<sup>st</sup>  Grd Exp UPS FAS Clipper Client Drop Off TestAmerica Courier Other \_\_\_\_\_  
**Receipt After-hours:** Drop-off Date/Time \_\_\_\_\_ Storage Location \_\_\_\_\_

TestAmerica Cooler # TA Foam Box Client Cooler Box Other \_\_\_\_\_  
 Packing material used: Bubble Wrap Foam Plastic Bag None Other \_\_\_\_\_  
 COOLANT: Wet Ice Blue Ice Dry Ice Water None

1. Cooler temperature upon receipt  See Multiple Cooler Form  
 IR GUN# IR-8 (CF +0.1 °C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C  
 IR GUN #36 (CF +0.6°C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C
2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 2 Yes No  
 -Were the seals on the outside of the cooler(s) signed & dated? Yes No NA  
 -Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No  
 -Were tamper/custody seals intact and uncompromised? Yes No NA
3. Shippers' packing slip attached to the cooler(s)? Yes No
4. Did custody papers accompany the sample(s)? Yes No
5. Were the custody papers relinquished & signed in the appropriate place? Yes No
6. Was/were the person(s) who collected the samples clearly identified on the COC? Yes No
7. Did all bottles arrive in good condition (Unbroken)? Yes No
8. Could all bottle labels be reconciled with the COC? Yes No
9. Were correct bottle(s) used for the test(s) indicated? Yes No
10. Sufficient quantity received to perform indicated analyses? Yes No
11. Are these work share samples? Yes No  
 If yes, Questions 12-16 have been checked at the originating laboratory.
12. Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC984738
13. Were VOAs on the COC? Yes No
14. Were air bubbles >6 mm in any VOA vials?  Larger than this. Yes No NA
15. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # B904401VB Yes No
16. Was a LL Hg or Me Hg trip blank present? Yes No

Tests that are not checked for pH by Receiving:  
 VOAs  
 Oil and Grease  
 TOC

Contacted PM \_\_\_\_\_ Date \_\_\_\_\_ by \_\_\_\_\_ via Verbal Voice Mail Other \_\_\_\_\_  
 Concerning \_\_\_\_\_

**17. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES** Samples processed by: Amm  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**18. SAMPLE CONDITION**  
 Sample(s) \_\_\_\_\_ were received after the recommended holding time had expired.  
 Sample(s) \_\_\_\_\_ were received in a broken container.  
 Sample(s) \_\_\_\_\_ were received with bubble >6 mm in diameter. (Notify PM)

**19. SAMPLE PRESERVATION**  
 Sample(s) \_\_\_\_\_ were further preserved in the laboratory.  
 Time preserved: \_\_\_\_\_ Preservative(s) added/Lot number(s): \_\_\_\_\_  
 VOA Sample Preservation - Date/Time VOAs Frozen: \_\_\_\_\_



# DATA VERIFICATION REPORT



August 8, 2019

Kris Hinskey  
Arcadis Inc  
10559 Citation Ave  
Suite 100  
Brighton, MI 48116

CADENA project ID: E203631

Project: Ford Livonia Transmission Project - OFF-SITE - Soil Gas and Groundwater

Project number: MI001454.0002/3/4.00002/2B/3B

Event Specific Scope of Work References: Sample COC

Laboratory: TestAmerica - North Canton

Laboratory submittal: 114396-1

Sample date: 2019-06-12

Report received by CADENA: 2019-06-26

Initial Data Verification completed by CADENA: 2019-06-26

Number of Samples:2

Sample Matrices: Water

Test Categories: GCMS VOC

**Please see attached criteria report or sample result/qualified analytical result summary for qualifier flags assigned to sample data.**

This report was revised to correct the sample name.

The following minor QC exceptions or missing information were noted:

GCMS VOC QC batch MS/MSD recovery outliers were not determined using a client sample so qualification was not required based on these sample-specific QC outliers.

Sample/MS/MSD Surrogate Recovery, Blank/LCS Surrogate Recovery, LCS/LCD Recovery, Blank Contamination and Hold Time Exception were reviewed as part of our verification.

Data verification for the report specified above was completed using the Ford Motor Company Environmental Laboratory Technical Specification, the CADENA Standard Operating Procedure for the Verification of Environmental Analytical Data and the associated analytical methods as references for evaluating the batch QC, sample data and report content. The EPA National Functional Guidelines for validating organic and inorganic data were used as guidance when addressing out of control QC results and the associated data qualifiers.

The definitions of the qualifiers used for this data package are defined in the analytical report. CADENA valid qualifiers are defined in the table below. To view and download a PDF copy of the laboratory analytical report access the CADENA CLMS at <http://clms.cadenaco.com/index.cfm>.

Please contact me if you have any questions.

Sincerely,

Jim Tomalia

Project Scientist

CADENA Inc, 1099 Highland Drive, Suite E, Ann Arbor, MI 48108 517-819-0356

## CADENA Valid Qualifiers

Valid Qualifiers	Description
<	Less than the reported concentration.
>	Greater than the reported concentration.
B	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was greater than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the reported concentration. For Inorganic methods the sample concentration was greater than the RDL and less than 10x the blank concentration and is considered non-detect at the reported concentration.
E	The analyte / Compound reported exceeds the calibration range and is considered estimated.
EMPC	Estimated Minimum Potential Contamination - Dioxin/Furan analyses only.
J	Indicates an estimated value. This flag is used either when estimating a concentration for a tentatively identified compound or when the data indicates the presence of an analyte / compound but the result is less than the sample Quantitation limit, but greater than zero. The flag is also used in data validation to indicate a reported value should be considered estimated due to associated quality assurance deficiencies.
J-	The result is an estimated quantity, but the result may be biased low.
JB	NON-DETECT AT THE CONCENTRATION REPORTED AND ESTIMATED
JH	The sample result is considered estimated and is potentially biased high.
JL	The sample result is considered estimated and is potentially biased low.
JUB	NON-DETECT AT THE REPORTING LIMIT AND ESTIMATED
NJ	Tentatively identified compound with approximated concentration.
R	Indicates the value is considered to be unusable. (Note: The analyte / compound may or may not be present.)
TNTC	Too Numerous to Count - Asbestos and Microbiological Results.
U	Indicates that the analyte / compound was analyzed for, but not detected.
UB	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was less than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the RDL. For Inorganic methods the sample concentration was less than the RDL and less than 10x the blank concentration and is considered non-detect at the RDL.
UJ	The analyte / compound was not detected above the reported sample Quantitation limit. However, the Quantitation limit is considered to be approximate due to associated quality assurance results and may or may not represent the actual limit of Quantitation to accurately and precisely report the analyte in the sample.

## SAMPLING AND ANALYSIS SUMMARY

**CADENA Project ID:** E203631

**Laboratory:** TestAmerica-North Canton

**Laboratory Submittal:** 114396-1

Lab Sample ID	Sample ID	Collection Date (mm/yy/dd)	Collection Time (hh:mm:ss)	Volatile Organics by GCMS	8260B with Single Ion Monitoring	Comment
2401143961	SUMP-34380CAPITOL-01_061219	6/12/2019	6:15:00	X	X	
2401143962	TRIP BLANK	6/12/2019	12:00:00	X		

# Analytical Results Summary

## Reportable Results Only

CADENA Project ID: E203631

Laboratory: TestAmerica - North Canton

Laboratory Submittal: 114396-1

Sample Name: SUMP-34380CAPITOL-01\_061219 TRIP BLANK  
Lab Sample ID: 2401143961 2401143962  
Sample Date: 6/12/2019 6/12/2019

Analyte	Cas No.	Report		Valid		Report		Valid		
		Result	Limit	Units	Qualifier	Result	Limit	Units	Qualifier	
<b>GC/MS VOC</b>										
<u>OSW-8260B</u>										
1,1-Dichloroethene	75-35-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---	
cis-1,2-Dichloroethene	156-59-2	ND	1.0	ug/l	---	ND	1.0	ug/l	---	
Tetrachloroethene	127-18-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---	
trans-1,2-Dichloroethene	156-60-5	ND	1.0	ug/l	---	ND	1.0	ug/l	---	
Trichloroethene	79-01-6	ND	1.0	ug/l	---	ND	1.0	ug/l	---	
Vinyl chloride	75-01-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---	
<u>OSW-8260BBSim</u>										
1,4-Dioxane	123-91-1	ND	2.0	ug/l	---					



# Ford Motor Company – Livonia Transmission Project

## DATA REVIEW

### Livonia, Michigan

Volatile Organic Compounds (VOC) Analysis

SDG #240-114396-1

CADENA Verification Report: 2019-06-26

Analyses Performed By:

TestAmerica  
Canton, Ohio

Report #33481R

Review Level: Tier III

Project: MI001454.0004.00002 (30016346)



## DATA REVIEW

### SUMMARY

This data quality assessment summarizes the review of Sample Delivery Group (SDG) # 240-114396-1 for samples collected in association with the Ford – Livonia, Michigan site. The review was conducted as a Tier III validation in addition to a verification/Tier II validation review performed by CADENA Inc. and included review of level IV laboratory data package completeness. Only elements of a Tier III validation effort (Tier III includes a detailed review of laboratory raw data to check for errors in calculation, calibration review, internal standard review and compound identification) and omitted deviations from the CADENA verification/Tier II report are documented in this report. Only analytical data associated with constituents of concern were reviewed for this validation. Field documentation was not included in this review. Included with this assessment are the validation annotated sample result sheets, and chain of custody. Analyses were performed on the following samples:

SDG	Sample ID	Lab ID	Matrix	Sample Collection Date	Parent Sample	Analysis		
						VOC (Full Scan)	VOC (SIM)	MISC
240-114396-1	SUMP-34380CAPITOL-01_061219	240-114396-1	Water	6/12/2019		X	X	
	TRIP BLANK	240-114396-2	Water	6/12/2019		X		

## DATA REVIEW

### ANALYTICAL DATA PACKAGE DOCUMENTATION

The table below is the evaluation of the data package completeness.

Items Reviewed	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
1. Sample receipt condition		X		X	
2. Requested analyses and sample results		X		X	
3. Master tracking list		X		X	
4. Methods of analysis		X		X	
5. Reporting limits		X		X	
6. Sample collection date		X		X	
7. Laboratory sample received date		X		X	
8. Sample preservation verification (as applicable)		X		X	
9. Sample preparation/extraction/analysis dates		X		X	
10. Fully executed Chain-of-Custody (COC) form		X		X	
11. Narrative summary of Quality Assurance or sample problems provided		X		X	
12. Data Package Completeness and Compliance		X		X	

## DATA REVIEW

### ORGANIC ANALYSIS INTRODUCTION

Analyses were performed according to United States Environmental Protection Agency (USEPA) SW-846 Method 8260B and 8260B SIM. Data were reviewed in accordance with USEPA National Functional Guidelines of October 1999.

The data review process is an evaluation of data on a technical basis rather than a determination of contract compliance. As such, the standards against which the data are being weighed may differ from those specified in the analytical method. It is assumed that the data package represents the best efforts of the laboratory and had already been subjected to adequate and sufficient quality review prior to submission.

During the review process, laboratory qualified and unqualified data are verified against the supporting documentation. Based on this evaluation, qualifier codes may be added, deleted, or modified by the data reviewer. Results are qualified with the following codes in accordance with USEPA National Functional Guidelines:

- Concentration (C) Qualifiers
  - U The analyte was analyzed for but was not detected above the level of the reported sample quantitation limit.
  - B The compound has been found in the sample as well as its associated blank, its presence in the sample may be suspect.
- Quantitation (Q) Qualifiers
  - E The compound was quantitated above the calibration range.
  - D Concentration is based on a diluted sample analysis.
- Validation Qualifiers
  - J The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
  - UJ The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
  - J+ The result is an estimated quantity, but the result may be biased high.
  - J- The result is an estimated quantity, but the result may be biased low.
  - UB Analyte considered non-detect at the listed value due to associated blank contamination.
  - N The analysis indicates the presence of a compound for which there is presumptive evidence to make a tentative identification.
  - R The sample results are rejected.

Two facts should be noted by all data users. First, the "R" flag means that the associated value is unusable. In other words, due to significant quality control (QC) problems, the analysis is invalid and provides no information as to whether the compound is present or not. "R" values should not appear on data tables because they cannot be relied upon, even as a last resort. The second fact to keep in mind is that no compound concentration, even if it has passed all QC tests, is guaranteed to be accurate. Strict QC serves to increase confidence in data but any value potentially contains error.

## DATA REVIEW

### VOLATILE ORGANIC COMPOUND (VOC) ANALYSES

#### 1. Holding Times

The specified holding times for the following methods are presented in the following table.

Method	Matrix	Holding Time	Preservation
SW-846 8260B	Water	14 days from collection to analysis	Cool to < 6 °C; pH < 2 with HCl

All samples were analyzed within the specified holding time criteria.

#### 2. Mass Spectrometer Tuning

Mass spectrometer performance was acceptable and all analyses were performed within a 12-hour tune clock.

System performance and column resolution were acceptable.

#### 3. Calibration

Satisfactory instrument calibration is established to ensure that the instrument is capable of producing acceptable quantitative data. An initial calibration demonstrates that the instrument is capable of acceptable performance at the beginning of an experimental sequence. The continuing calibration verifies that the instrument daily performance is satisfactory.

##### 3.1 Initial Calibration

The method specifies percent relative standard deviation (%RSD) and relative response factor (RRF) limits for select compounds only. A technical review of the data applies limits to all compounds with no exceptions.

All target compounds associated with the initial calibration standards must exhibit a %RSD less than the control limit (20%) or a correlation coefficient greater than 0.99 and an RRF value greater than control limit (0.05).

All compounds associated with the initial calibrations were within the specified control limits.

##### 3.2 Continuing Calibration

All target compounds associated with the continuing calibration standard must exhibit a percent difference (%D) less than the control limit (20%) and RRF value greater than control limit (0.05).

All compounds associated with the calibrations were within the specified control limits, with the exception of the compounds presented in the following table.

Sample Locations	Initial/Continuing	Compound	Criteria
SUMP-34380CAPITOL-01_061219 TRIP BLANK	CCV %D	Trichloroethene	+20.9%
	CCV %D	Tetrachloroethene	+24.8%

## DATA REVIEW

The criteria used to evaluate the initial and continuing calibration are presented in the following table. In the case of a calibration deviation, the sample results are qualified.

Initial/Continuing	Criteria	Sample Result	Qualification
Continuing Calibration	%D >20% (increase in sensitivity)	Non-detect	No Action
		Detect	J
	%D >20% (decrease in sensitivity)	Non-detect	UJ
		Detect	J
	%D >90% (increase/decrease in sensitivity)	Non-detect	R
		Detect	J

### 4. Internal Standard Performance

Internal standard performance criteria insure that the GC/MS sensitivity and response are stable during every sample analysis. The criteria require the internal standard compounds associated with the VOC exhibit area counts that are not greater than two times (+100%) or less than one-half (-50%) of the area counts of the associated continuing calibration standard.

All internal standard responses were within control limits.

### 5. Compound Identification

Compounds are identified on the GC/MS by using the analytes relative retention time and ion spectra.

All detected compounds met the specified criteria.

### 6. System Performance and Overall Assessment

Overall system performance was acceptable. Other than for those deviations specifically mentioned in this review, the overall data quality is within the guidelines specified in the method.

**DATA REVIEW**

**DATA VALIDATION CHECKLIST FOR VOCs**

VOCs: 8260B/8260B-SIM	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
<b>GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS)</b>					
<b>Tier II Validation</b>					
Holding times/Preservation		X		X	
<b>Tier III Validation</b>					
System performance and column resolution		X		X	
Initial calibration %RSDs		X		X	
Continuing calibration RRFs		X		X	
Continuing calibration %Ds		X	X		
Instrument tune and performance check		X		X	
Ion abundance criteria for each instrument used		X		X	
Internal standard		X		X	
Compound identification and quantitation					
A. Reconstructed ion chromatograms		X		X	
B. Quantitation Reports		X		X	
C. RT of sample compounds within the established RT windows		X		X	
D. Transcription/calculation errors present		X		X	
E. Reporting limits adjusted to reflect sample dilutions		X		X	

Notes:

- %RSD Relative standard deviation
- %R Percent recovery
- RPD Relative percent difference
- %D Percent difference

VALIDATION PERFORMED BY: Andrew Korycinski

SIGNATURE:



DATE: July 30, 2019

PEER REVIEW: Dennis Capria

DATE: August 5, 2019





**CHAIN OF CUSTODY  
CORRECTED SAMPLE ANALYSIS DATA  
SHEETS**



**NO CORRECTIONS/QUALIFERS ADDED  
TO SAMPLE ANALYSIS DATA SHEETS**



# MICHIGAN 190

## Chain of Custody Record

**TestAmerica**  
THE LEADER IN ENVIRONMENTAL TESTING

1.8/1.9

TestAmerica Laboratory location: N.Canton — 4101 Shuffel Street NW/ North Canton, OH 44720 / 330-497-9396

<b>Client Contact</b> Company Name: Arcadis Address: 28550 Cabot Drive, Suite 500 City/State/Zip: Novi, MI, 48377 Phone: 248-994-2240 Project Name: Ford LTP Project Number: MI001454.0003 PO # MI001454.0003		<b>Regulatory program:</b> <input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other																													
<b>Client Project Manager:</b> Kris Hinskey Telephone: 248-994-2240 Email: kristoffer.hinskey@arcadis.com		<b>Site Contact:</b> Angela DeGrandis Telephone: 734-320-0065																													
<b>Lab Contact:</b> Mike DelMonico Telephone: 330-497-9396		TestAmerica Laboratories, Inc. COC No: _____ of _____ COCs																													
TAT if different from below <input type="checkbox"/> 3 weeks <input type="checkbox"/> 2 weeks <input checked="" type="checkbox"/> 1 week <input type="checkbox"/> 2 days <input type="checkbox"/> 1 day		Walk-in client Lab sampling Job/SDG No: _____																													
Method of Shipment/Carrier: Shipping/Tracking No:		Sample Specific Notes / Special Instructions:																													
Sample Identification SUMP-34380CAPITOL-01-06019 TRIP BLANK		6 CONTAINERS																													
Sample Date: _____ Sample Time: _____		Filtered Sample (Y/N)																													
Matrix: <table border="1"> <tr> <th>Air</th> <th>Aqueous</th> <th>Sediment</th> <th>Solid</th> <th>Other:</th> </tr> <tr> <td>X</td> <td>X</td> <td></td> <td></td> <td></td> </tr> </table>		Air	Aqueous	Sediment	Solid	Other:	X	X				Composite=C / Grab=G																			
Air	Aqueous	Sediment	Solid	Other:																											
X	X																														
Containers & Preservatives: <table border="1"> <tr> <th>H2SO4</th> <th>HNO3</th> <th>HCl</th> <th>NaOH</th> <th>ZnAc</th> <th>NaOH</th> <th>Other:</th> </tr> <tr> <td>X</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> </table>		H2SO4	HNO3	HCl	NaOH	ZnAc	NaOH	Other:	X							Analyses: <table border="1"> <tr> <th>1,1-DCE 8260B</th> <th>cis-1,2-DCE 8260B</th> <th>Trans-1,2-DCE 8260B</th> <th>PCE 8260B</th> <th>TCE 8260B</th> <th>Vinyl Chloride 8260B</th> <th>1,4-Dioxane 8260B SIM</th> </tr> <tr> <td>X</td> <td>X</td> <td>X</td> <td>X</td> <td>X</td> <td>X</td> <td>X</td> </tr> </table>		1,1-DCE 8260B	cis-1,2-DCE 8260B	Trans-1,2-DCE 8260B	PCE 8260B	TCE 8260B	Vinyl Chloride 8260B	1,4-Dioxane 8260B SIM	X	X	X	X	X	X	X
H2SO4	HNO3	HCl	NaOH	ZnAc	NaOH	Other:																									
X																															
1,1-DCE 8260B	cis-1,2-DCE 8260B	Trans-1,2-DCE 8260B	PCE 8260B	TCE 8260B	Vinyl Chloride 8260B	1,4-Dioxane 8260B SIM																									
X	X	X	X	X	X	X																									
Possible Hazard Identification <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Irritable <input type="checkbox"/> In Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return to Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months																													
Special Instructions/QC Requirements & Comments: Submit all results through Cadena at jim.tomalia@cadenalab.com. Cadena #E203531 Level IV Reporting.																															
Requisitioned by: _____ Relinquished by: _____ Relinquished by: _____	Company: Arcadis Company: ARCADIS Company: ETC	Received by: _____ Received by: _____ Received in Laboratory by: _____	Date/Time: 6-12-19 1930 Date/Time: 6-13-19 0951 Date/Time: 6-14-19 815																												

©2008 TestAmerica Laboratories, Inc. All rights reserved.  
 TestAmerica & Design are trademarks of TestAmerica Laboratories, Inc.



# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-114396-1

**Client Sample ID: SUMP-34380CAPITOL-01\_061219**

**Lab Sample ID: 240-114396-1**

**Date Collected: 06/12/19 18:15**

**Matrix: Water**

**Date Received: 06/14/19 08:15**

**Method: 8260B SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			06/18/19 15:31	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	108		63 - 125					06/18/19 15:31	1

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:30	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/24/19 20:30	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/24/19 20:30	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:30	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/24/19 20:30	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/24/19 20:30	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	118		70 - 121					06/24/19 20:30	1
4-Bromofluorobenzene (Surr)	101		59 - 120					06/24/19 20:30	1
Toluene-d8 (Surr)	95		70 - 123					06/24/19 20:30	1
Dibromofluoromethane (Surr)	114		75 - 128					06/24/19 20:30	1

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-114396-2**

**Date Collected: 06/12/19 00:00**

**Matrix: Water**

**Date Received: 06/14/19 08:15**

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:55	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/24/19 20:55	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/24/19 20:55	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/24/19 20:55	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/24/19 20:55	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/24/19 20:55	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	118		70 - 121					06/24/19 20:55	1
4-Bromofluorobenzene (Surr)	89		59 - 120					06/24/19 20:55	1
Toluene-d8 (Surr)	91		70 - 123					06/24/19 20:55	1
Dibromofluoromethane (Surr)	123		75 - 128					06/24/19 20:55	1

9/26/2019

Mr. Jim Tomalia  
Arcadis U.S., Inc.  
28550 Cabot Dr.  
Suite 500  
Novi MI 48377

Project Name: Ford LTP Off-Site Sampling

Project #: 30016344.0002B

Workorder #: 1906311R1

Dear Mr. Jim Tomalia

The following report includes the data for the above referenced project for sample(s) received on 6/17/2019 at Air Toxics Ltd.

The data and associated QC analyzed by TO-15 are compliant with the project requirements or laboratory criteria with the exception of the deviations noted in the attached case narrative.

Thank you for choosing Eurofins Air Toxics Inc. for your air analysis needs. Eurofins Air Toxics Inc. is committed to providing accurate data of the highest quality. Please feel free to contact the Project Manager: Ausha Scott at 916-985-1000 if you have any questions regarding the data in this report.

Regards,



Ausha Scott  
Project Manager

**WORK ORDER #: 1906311R1**

Work Order Summary

<b>CLIENT:</b>	Mr. Jim Tomalia Arcadis U.S., Inc. 28550 Cabot Dr. Suite 500 Novi, MI 48377	<b>BILL TO:</b>	Accounts Payable Arcadis U.S., Inc. 630 Plaza Drive Suite 600 Highlands Ranch, CO 80129
<b>PHONE:</b>	517-819-0356	<b>P.O. #</b>	30016344.0002B
<b>FAX:</b>		<b>PROJECT #</b>	30016344.0002B Ford LTP Off-Site
<b>DATE RECEIVED:</b>	06/17/2019	<b>CONTACT:</b>	Sampling Ausha Scott
<b>DATE COMPLETED:</b>	06/22/2019		
<b>DATE REISSUED:</b>	09/26/2019		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>	<u>FINAL PRESSURE</u>
01A	SSMP-34380CAPITOL-01_061219	TO-15	6.1 "Hg	15.2 psi
02A	Lab Blank	TO-15	NA	NA
03A	CCV	TO-15	NA	NA
04A	LCS	TO-15	NA	NA
04AA	LCSD	TO-15	NA	NA

CERTIFIED BY:   
 Technical Director

DATE: 09/26/19

Certification numbers: AZ Licensure AZ0775, FL NELAP – E87680, LA NELAP – 02089, NH NELAP - 209218, NJ NELAP - CA016, NY NELAP - 11291, TX NELAP - T104704434-18-13, UT NELAP – CA009332019-11, VA NELAP - 460197, WA NELAP - C935

Name of Accreditation Body: NELAP/ORELAP (Oregon Environmental Laboratory Accreditation Program)

Accreditation number: CA300005-011, Effective date: 10/18/2018, Expiration date: 10/17/2019.

Eurofins Air Toxics Inc.. certifies that the test results contained in this report meet all requirements of the NELAC standards

This report shall not be reproduced, except in full, without the written approval of Eurofins Air Toxics, Inc.

180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630

(916) 985-1000 . (800) 985-5955 . FAX (916) 985-1020

**LABORATORY NARRATIVE**  
**EPA Method TO-15**  
**Arcadis U.S., Inc.**  
**Workorder# 1906311R1**

One 1 Liter Summa Canister (100% Certified) sample was received on June 17, 2019. The laboratory performed analysis via EPA Method TO-15 using GC/MS in the full scan mode.

This workorder was independently validated prior to submittal using 'USEPA National Functional Guidelines' as generally applied to the analysis of volatile organic compounds in air. A rules-based, logic driven, independent validation engine was employed to assess completeness, evaluate pass/fail of relevant project quality control requirements and verification of all quantified amounts.

**Receiving Notes**

There were no receiving discrepancies.

The work order was reissued on 9/26/2019 to change identification of sample SSMP-34380CAPITOL-01\_061219 per the revised Chain of Custody (COC) provided by the client.

**Analytical Notes**

As per client project requirements, the laboratory has reported estimated values for target compound hits that are below the Reporting Limit but greater than the Method Detection Limit. Concentrations that are below the level at which the canister was certified may be false positives.

**Definition of Data Qualifying Flags**

Ten qualifiers may have been used on the data analysis sheets and indicates as follows:

B - Compound present in laboratory blank greater than reporting limit (background subtraction not performed).

J - Estimated value.

E - Exceeds instrument calibration range.

S - Saturated peak.

Q - Exceeds quality control limits.

U - Compound analyzed for but not detected above the reporting limit, LOD, or MDL value. See data page for project specific U-flag definition.

UJ- Non-detected compound associated with low bias in the CCV

N - The identification is based on presumptive evidence.

M - Reported value may be biased due to apparent matrix interferences.

CN - See Case Narrative.

File extensions may have been used on the data analysis sheets and indicates as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	SSMP-34380CAPITOL-01_061219	<b>Date/Time Analyzed:</b>	6/19/19 02:50 PM
<b>Lab ID:</b>	1906311R1-01A	<b>Dilution Factor:</b>	2.55
<b>Date/Time Collected:</b>	6/12/19 06:36 PM	<b>Instrument/Filename:</b>	msdp.i / p061909
<b>Media:</b>	1 Liter Summa Canister (100% Certified)		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	1.9	4.5	5.0	Not Detected
1,4-Dioxane	123-91-1	2.4	13	18	Not Detected
cis-1,2-Dichloroethene	156-59-2	2.3	4.5	5.0	Not Detected
Tetrachloroethene	127-18-4	1.6	7.8	8.6	5.2 J
trans-1,2-Dichloroethene	156-60-5	3.1	4.5	5.0	Not Detected
Trichloroethene	79-01-6	0.90	6.2	6.8	Not Detected
Vinyl Chloride	75-01-4	0.78	2.9	3.2	Not Detected

J = Estimated value.

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	111
4-Bromofluorobenzene	460-00-4	70-130	89
Toluene-d8	2037-26-5	70-130	104



EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	Lab Blank	<b>Date/Time Analyzed:</b>	6/19/19 11:25 AM
<b>Lab ID:</b>	1906311R1-02A	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdp.i / p061906a
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.75	1.8	2.0	Not Detected
1,4-Dioxane	123-91-1	0.95	5.0	7.2	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.90	1.8	2.0	Not Detected
Tetrachloroethene	127-18-4	0.64	3.0	3.4	Not Detected
trans-1,2-Dichloroethene	156-60-5	1.2	1.8	2.0	Not Detected
Trichloroethene	79-01-6	0.35	2.4	2.7	Not Detected
Vinyl Chloride	75-01-4	0.30	1.1	1.3	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	108
4-Bromofluorobenzene	460-00-4	70-130	83
Toluene-d8	2037-26-5	70-130	101

EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	CCV	<b>Date/Time Analyzed:</b>	6/19/19 09:32 AM
<b>Lab ID:</b>	1906311R1-03A	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdp.i / p061902
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
1,1-Dichloroethene	75-35-4	102
1,4-Dioxane	123-91-1	102
cis-1,2-Dichloroethene	156-59-2	100
Tetrachloroethene	127-18-4	101
trans-1,2-Dichloroethene	156-60-5	108
Trichloroethene	79-01-6	102
Vinyl Chloride	75-01-4	130

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	116
4-Bromofluorobenzene	460-00-4	70-130	98
Toluene-d8	2037-26-5	70-130	110

EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	LCS	<b>Date/Time Analyzed:</b>	6/19/19 10:05 AM
<b>Lab ID:</b>	1906311R1-04A	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdp.i / p061903
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
1,1-Dichloroethene	75-35-4	107
1,4-Dioxane	123-91-1	102
cis-1,2-Dichloroethene	156-59-2	115
Tetrachloroethene	127-18-4	99
trans-1,2-Dichloroethene	156-60-5	96
Trichloroethene	79-01-6	104
Vinyl Chloride	75-01-4	130

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	115
4-Bromofluorobenzene	460-00-4	70-130	95
Toluene-d8	2037-26-5	70-130	108

\* % Recovery is calculated using unrounded analytical results.

EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	LCSD	<b>Date/Time Analyzed:</b>	6/19/19 10:31 AM
<b>Lab ID:</b>	1906311R1-04AA	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdp.i / p061904
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
1,1-Dichloroethene	75-35-4	105
1,4-Dioxane	123-91-1	103
cis-1,2-Dichloroethene	156-59-2	116
Tetrachloroethene	127-18-4	100
trans-1,2-Dichloroethene	156-60-5	93
Trichloroethene	79-01-6	104
Vinyl Chloride	75-01-4	129

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	114
4-Bromofluorobenzene	460-00-4	70-130	97
Toluene-d8	2037-26-5	70-130	106

\* % Recovery is calculated using unrounded analytical results.



REVISED REPORT: September 26, 2019  
REVISION SUMMARY: Sample ID's updated per revised COC per client request.

Kris Hinskey  
Arcadis Inc  
10559 Citation Ave  
Suite 100  
Brighton, MI 48116

CADENA project ID: E203631  
Project: Ford Livonia Transmission Project - OFF-SITE - Soil Gas and Groundwater  
Project number: MI001454.0002/3/4.00002/2B/3B  
Client project scope reference: Sample COC only was used to define project analytical requirements.  
Laboratory: Eurofins Air Toxics - Folsom  
Laboratory submittal: 1906311  
Sample date: 2019-06-12  
Report received by CADENA: 2019-09-26  
Initial Data Verification completed by CADENA: 2019-06-23

1 Air sample was analyzed for TO-15 parameters.

There were no significant QC anomalies or exceptions to report.

Data verification for the report specified above was completed using the Ford Motor Company Environmental Laboratory Technical Specification, the CADENA Standard Operating Procedure for the Verification of Environmental Analytical Data and the associated analytical methods as references for evaluating the batch QC, sample data and report content. The EPA National Functional Guidelines for validating organic and inorganic data were used as guidance when addressing out of control QC results and the associated data qualifiers.

Analytical results reported between RDL and MDL are flagged 'J' and considered estimated values.

The definitions of the qualifiers used for this data package are defined in the analytical report. CADENA valid qualifiers are defined in the table below. To view and download a PDF copy of the laboratory analytical report access the CADENA CLMS at <http://clms.cadenaco.com/index.cfm>.

Please contact me if you have any questions.

Sincerely,

Jim Tomalia

Project Scientist

CADENA Inc, 1099 Highland Drive, Suite E, Ann Arbor, MI 48108 517-819-0356

## CADENA Valid Qualifiers

Valid Qualifiers	Description
<	Less than the reported concentration.
>	Greater than the reported concentration.
B	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was greater than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the reported concentration. For Inorganic methods the sample concentration was greater than the RDL and less than 10x the blank concentration and is considered non-detect at the reported concentration.
E	The analyte / Compound reported exceeds the calibration range and is considered estimated.
EMPC	Estimated Minimum Potential Contamination - Dioxin/Furan analyses only.
J	Indicates an estimated value. This flag is used either when estimating a concentration for a tentatively identified compound or when the data indicates the presence of an analyte / compound but the result is less than the sample Quantitation limit, but greater than zero. The flag is also used in data validation to indicate a reported value should be considered estimated due to associated quality assurance deficiencies.
J-	The result is an estimated quantity, but the result may be biased low.
JB	NON-DETECT AT THE CONCENTRATION REPORTED AND ESTIMATED
JH	The sample result is considered estimated and is potentially biased high.
JL	The sample result is considered estimated and is potentially biased low.
JUB	NON-DETECT AT THE REPORTING LIMIT AND ESTIMATED
NJ	Tentatively identified compound with approximated concentration.
R	Indicates the value is considered to be unusable. (Note: The analyte / compound may or may not be present.)
TNTC	Too Numerous to Count - Asbestos and Microbiological Results.
U	Indicates that the analyte / compound was analyzed for, but not detected.
UB	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was less than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the RDL. For Inorganic methods the sample concentration was less than the RDL and less than 10x the blank concentration and is considered non-detect at the RDL.
UJ	The analyte / compound was not detected above the reported sample Quantitation limit. However, the Quantitation limit is considered to be approximate due to associated quality assurance results and may or may not represent the actual limit of Quantitation to accurately and precisely report the analyte in the sample.

# Ford Motor Company – Livonia Transmission Project

## DATA REVIEW

### Livonia, Michigan

Volatile Organic Compounds (VOC) TO-15 Analysis

SDG #1906311R1

CADENA Verification Report: 2019-06-23

Analyses Performed By:  
Eurofins Air Toxics  
Folsom, California

Report #33403R rev1  
Review Level: Tier III  
Project: 30016346.00003 (MI001454.0004.00002)

## DATA REVIEW

### SUMMARY

This data quality assessment summarizes the review of Sample Delivery Group (SDG) # 1906311R1 for samples collected in association with the Ford – Livonia, Michigan site. The review was conducted as a Tier III validation in addition to a verification/Tier II validation review performed by CADENA Inc. and included review of level IV laboratory data package completeness. Only elements of a Tier III validation effort (Tier III includes a detailed review of laboratory raw data to check for errors in calculation, calibration review, internal standard review and compound identification) and omitted deviations from the CADENA verification/Tier II report are documented in this report. Only analytical data associated with constituents of concern were reviewed for this validation. Field documentation was not included in this review. Included with this assessment are the validation annotated sample result sheets, and chain of custody. Analyses were performed on the following samples:

SDG	Sample ID	Lab ID	Matrix	Sample Collection Date	Parent Sample	Analysis		
						TO-15 (Full Scan)	TO-15 (SIM)	MISC
1906311R1	SSMP-34380CAPITOL-01_061219	1906311R1-01A	Air	6/12/2019		X		

Note: This report was revised to incorporate the correct sample address designation of “CAPITOL” (Original report was CAPITAL) into the sample ID.



## DATA REVIEW

### ANALYTICAL DATA PACKAGE DOCUMENTATION

The table below is the evaluation of the data package completeness.

Items Reviewed	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
1. Sample receipt condition		X		X	
2. Requested analyses and sample results		X		X	
3. Master tracking list		X		X	
4. Methods of analysis		X		X	
5. Reporting limits		X		X	
6. Sample collection date		X		X	
7. Laboratory sample received date		X		X	
8. Sample preservation verification (as applicable)		X		X	
9. Sample preparation/extraction/analysis dates		X		X	
10. Fully executed Chain-of-Custody (COC) form		X		X	
11. Narrative summary of Quality Assurance or sample problems provided		X		X	
12. Data Package Completeness and Compliance		X		X	

## DATA REVIEW

### ORGANIC ANALYSIS INTRODUCTION

Analyses were performed according to United States Environmental Protection Agency (USEPA) Method TO-15 (Full Scan). Data were reviewed in accordance with USEPA National Functional Guidelines of October 1999.

The data review process is an evaluation of data on a technical basis rather than a determination of contract compliance. As such, the standards against which the data are being weighed may differ from those specified in the analytical method. It is assumed that the data package represents the best efforts of the laboratory and had already been subjected to adequate and sufficient quality review prior to submission.

During the review process, laboratory qualified and unqualified data are verified against the supporting documentation. Based on this evaluation, qualifier codes may be added, deleted, or modified by the data reviewer. Results are qualified with the following codes in accordance with USEPA National Functional Guidelines:

- Concentration (C) Qualifiers
  - U The analyte was analyzed for but was not detected above the level of the reported sample quantitation limit.
  - B The compound has been found in the sample as well as its associated blank, its presence in the sample may be suspect.
- Quantitation (Q) Qualifiers
  - E The compound was quantitated above the calibration range.
  - D Concentration is based on a diluted sample analysis.
- Validation Qualifiers
  - J The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
  - UJ The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
  - J+ The result is an estimated quantity, but the result may be biased high.
  - J- The result is an estimated quantity, but the result may be biased low.
  - UB Analyte considered non-detect at the listed value due to associated blank contamination.
  - N The analysis indicates the presence of a compound for which there is presumptive evidence to make a tentative identification.
  - R The sample results are rejected.

Two facts should be noted by all data users. First, the "R" flag means that the associated value is unusable. In other words, due to significant quality control (QC) problems, the analysis is invalid and provides no information as to whether the compound is present or not. "R" values should not appear on data tables because they cannot be relied upon, even as a last resort. The second fact to keep in mind is that no compound concentration, even if it has passed all QC tests, is guaranteed to be accurate. Strict QC serves to increase confidence in data but any value potentially contains error.

## DATA REVIEW

### VOLATILE ORGANIC COMPOUND (VOC) ANALYSES

#### 1. Holding Times

The specified holding times for the following methods are presented in the following table.

Method	Matrix	Holding Time	Preservation	Return Canister Pressure
USEPA TO-15	Air	30 days from collection to analysis (Canister)	Ambient Temperature	< -2" Hg

All samples were analyzed within the specified holding time and canister return pressure / vacuum criteria.

#### 2. Mass Spectrometer Tuning

Mass spectrometer performance was acceptable and all analyses were performed within a 12-hour tune clock.

System performance and column resolution were acceptable.

#### 3. Calibration

Satisfactory instrument calibration is established to ensure that the instrument is capable of producing acceptable quantitative data. An initial calibration demonstrates that the instrument is capable of acceptable performance at the beginning of an experimental sequence. The continuing calibration verifies that the instrument daily performance is satisfactory.

##### 3.1 Initial Calibration

The method specifies percent relative standard deviation (%RSD) and relative response factor (RRF) limits for select compounds only. A technical review of the data applies limits to all compounds with no exceptions.

All target compounds associated with the initial calibration standards must exhibit a %RSD less than the control limit (30%) or a correlation coefficient greater than 0.99 and an RRF value greater than control limit (0.05).

All compounds associated with the initial calibrations were within the specified control limits.

##### 3.2 Continuing Calibration

All target compounds associated with the continuing calibration standard must exhibit a percent difference (%D) less than the control limit (30%) and RRF value greater than control limit (0.05).

All compounds associated with the continuing calibrations were within the specified control limits.

#### 4. Internal Standard Performance

Internal standard performance criteria ensure that the GC/MS sensitivity and response are stable during every sample analysis. The criteria requires the internal standard compounds associated with the VOC exhibit area counts that are not greater than 140% or less than 60% of the area counts of the associated continuing calibration standard.

All internal standard responses were within control limits.

## **DATA REVIEW**

### **5. Compound Identification**

Compounds are identified on the GC/MS by using the analytes relative retention time and ion spectra. All identified compounds met the specified criteria.

### **6. Field Duplicate Sample Analysis**

The field duplicate analysis is used to assess the precision of the field sampling procedures and analytical method. A control limit of 35% for air matrices is applied to the RPD between the parent sample and the field duplicate. In the instance when the parent and/or duplicate sample concentrations are not greater than five times the RL, a control limit of one times the RL is applied to the difference between the duplicate sample results.

A field duplicate was not performed on a sample location within this SDG.

### **7. System Performance and Overall Assessment**

Overall system performance was acceptable. Other than for those deviations specifically mentioned in this review, the overall data quality is within the guidelines specified in the method.

**DATA REVIEW**

**DATA VALIDATION CHECKLIST FOR VOCs**

VOCs: TO-15 ( Full Scan)	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
<b>GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS)</b>					
<b>Tier II Validation</b>					
Canister return pressure (<-2"Hg)		X		X	
<b>Tier III Validation</b>					
System performance and column resolution		X		X	
Initial calibration %RSDs		X		X	
Continuing calibration RRFs		X		X	
Continuing calibration %Ds		X		X	
Instrument tune and performance check		X		X	
Ion abundance criteria for each instrument used		X		X	
Internal standard		X		X	
Field Duplicate Sample RPD					X
Compound identification and quantitation					
A. Reconstructed ion chromatograms		X		X	
B. Quantitation Reports		X		X	
C. RT of sample compounds within the established RT windows		X		X	
D. Transcription/calculation errors present		X		X	
E. Reporting limits adjusted to reflect sample dilutions		X		X	

Notes:

%RSD Relative standard deviation

%R Percent recovery

RPD Relative percent difference

%D Percent difference

VALIDATION PERFORMED BY: Joseph C. Houser


SIGNATURE:



DATE: July 11, 2019

PEER REVIEW: Dennis Capria

DATE: July 11, 2019



**CHAIN OF CUSTODY  
CORRECTED SAMPLE ANALYSIS DATA  
SHEETS**



**NO CORRECTIONS/QUALIFIERS ADDED  
TO SAMPLE ANALYSIS DATA SHEETS**





EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	SSMP-34380CAPITOL-01_061219	<b>Date/Time Analyzed:</b>	6/19/19 02:50 PM
<b>Lab ID:</b>	1906311R1-01A	<b>Dilution Factor:</b>	2.55
<b>Date/Time Collected:</b>	6/12/19 06:36 PM	<b>Instrument/Filename:</b>	msdp.i / p061909
<b>Media:</b>	1 Liter Summa Canister (100% Certified)		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	1.9	4.5	5.0	Not Detected
1,4-Dioxane	123-91-1	2.4	13	18	Not Detected
cis-1,2-Dichloroethene	156-59-2	2.3	4.5	5.0	Not Detected
Tetrachloroethene	127-18-4	1.6	7.8	8.6	5.2 J
trans-1,2-Dichloroethene	156-60-5	3.1	4.5	5.0	Not Detected
Trichloroethene	79-01-6	0.90	6.2	6.8	Not Detected
Vinyl Chloride	75-01-4	0.78	2.9	3.2	Not Detected

J = Estimated value.

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	111
4-Bromofluorobenzene	460-00-4	70-130	89
Toluene-d8	2037-26-5	70-130	104

Analysis Request Canister Chain of Custody  
For Laboratory Use  
Only  
Version: 08/11/10

180 Blue Ravine Rd. Suite B, Folsom, CA 95630  
Phone (800) 985-9955; Fax (916) 351-8275  
Client: Aracis PID:

Project Name: Ford CTP Off Site Sampling  
Project Manager: Kili Hrosky P.O.  
Sampler: Shantel Johnson  
Site Name: 24380 GARFIELD

MWD1414 0000 0000

Special Instructions/Notes: Report ONLY: 1,1-DCE, cis-1,2-DCE, trans-1,2-DCE, 1,4-Dioxane, PCE, TCE and VC. Submit results through Cadent at jmh@mhaz@cadent.com. Cadent #E203631. Level IV Reporting

Turnaround Time (Rush surcharges may apply)

5 Day Turnaround Time

Lab ID	Sample Identification	Canister #	Flow Controller #	Start Sampling Information		Stop Sampling Information		Canister Vacuum/Pressure		Lab Use Only		Requested Analytes TG-19 (See Special Instructions/Notes)
				Date	Time	Date	Time	Initial (Hg)	Final (Hg)	Receipt	Final (psig) Gas: N2 / He	
	SSMP-24380CAP-01_091219	3L2834	23795	09/22/19	18:24	09/22/19	18:28	05	6			
Relinquished by: (Signature/Affiliation)			Date	Time	Relinquished by: (Signature/Affiliation)	Date	Time					
Relinquished by: (Signature/Affiliation)			Date	Time	Relinquished by: (Signature/Affiliation)	Date	Time					
Relinquished by: (Signature/Affiliation)			Date	Time	Relinquished by: (Signature/Affiliation)	Date	Time					
Stripper Name		Custody Given Insect?	Yes	No	None							

Sample Transportation Notice: Relinquishing signature on this document indicates that samples are shipped in compliance with all applicable local, State, Federal, and international laws, regulations, and ordinances of any kind. Relinquishing signature also indicates agreement to hold harmless, defend, and indemnify Eurofins Air Toxic against any claim, demand, or action of any kind, related to the collection, handling, or shipping of samples. O O T Hotline (800) 467-4322

Revised COC 9/24/19 -AS

Analysis Request /Canister Chain of Custody

For Laboratory Use

Only

Workorder # :

190631

PID: \_\_\_\_\_

Page 1 of 1

180 Blue Ravine Rd. Suite B, Folsom, CA 95630

Phone (800) 985-5955; Fax (916) 351-8279

Client: Arcadis PID: \_\_\_\_\_

Project Name: Ford LTP Off-Site Sampling

Project Manager: Kris Hinskey P.O# MI001454.0003.00002

Sampler: Shantel Johnson

Site Name: 34380 CAPITOL

Special Instructions/Notes: Report ONLY: 1,1-DCE, cis-1,2-DCE, trans-1,2-DCE, 1,4-Dioxane, PCE, TCE and VC. Submit results through Cadena at jim.tomalia@cadena.com. Cadena #E203631. Level IV Reporting

Turnaround Time (Rush surcharges may apply)

5 Day Turnaround Time

Lab ID	Sample Identification	Canister #	Flow Controller #	Start Sampling Information		Stop Sampling Information		Canister Vacuum/Pressure		Requested Analyses	
				Date	Time	Date	Time	Initial (in Hg)	Final (in Hg)	Receipt	Final (pa/g) Gas: N2 / He
01A	SSMP-34380CAPITAL-01_061219	1L2634	23795	06/12/2019	18:24	06/12/2019	18:36	-29	-6		x
Relinquished by: (Signature/Affiliation)				Date	6-13-19	Time	1200	Relinquished by: (Signature/Affiliation)			
Relinquished by: (Signature/Affiliation)				Date		Time		Relinquished by: (Signature/Affiliation)			
Relinquished by: (Signature/Affiliation)				Date		Time		Relinquished by: (Signature/Affiliation)			

Shipper Name: Fel Sr Custody Seals Intact? Yes Lab Use Only: No

Sample Transportation Notice: Relinquishing signature on this document indicates that samples are shipped in compliance with all applicable local, State, Federal, and international laws, regulations, and ordinances of any kind. Relinquishing signature also indicates agreement to hold harmless, defend, and indemnify Eurofins Air Toxics against any claim, demand, or action, of any kind, related to the collection, handling, of shipping of samples. U.O.T Hotline (800) 467-4922

Custody Seal Intact?

Y N None Temp NA

Fel Sr  
JTL EARL 6/17/19 0904

9/26/2019

Mr. Jim Tomalia  
Arcadis U.S., Inc.  
28550 Cabot Dr.  
Suite 500  
Novi MI 48377

Project Name: Ford LTP Off-Site Sampling  
Project #: 30016344.0002B  
Workorder #: 1906314R1

Dear Mr. Jim Tomalia

The following report includes the data for the above referenced project for sample(s) received on 6/17/2019 at Air Toxics Ltd.

The data and associated QC analyzed by Modified TO-15 are compliant with the project requirements or laboratory criteria with the exception of the deviations noted in the attached case narrative.

Thank you for choosing Eurofins Air Toxics Inc. for your air analysis needs. Eurofins Air Toxics Inc. is committed to providing accurate data of the highest quality. Please feel free to contact the Project Manager: Ausha Scott at 916-985-1000 if you have any questions regarding the data in this report.

Regards,



Ausha Scott  
Project Manager

**WORK ORDER #: 1906314R1**

Work Order Summary

<b>CLIENT:</b>	Mr. Jim Tomalia Arcadis U.S., Inc. 28550 Cabot Dr. Suite 500 Novi, MI 48377	<b>BILL TO:</b>	Accounts Payable Arcadis U.S., Inc. 630 Plaza Drive Suite 600 Highlands Ranch, CO 80129
<b>PHONE:</b>	517-819-0356	<b>P.O. #</b>	30016344.0002B
<b>FAX:</b>		<b>PROJECT #</b>	30016344.0002B Ford LTP Off-Site
<b>DATE RECEIVED:</b>	06/17/2019	<b>CONTACT:</b>	Sampling Ausha Scott
<b>DATE COMPLETED:</b>	06/24/2019		
<b>DATE REISSUED:</b>	09/26/2019		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>	<u>FINAL PRESSURE</u>
01A	AA-34380CAPITOL-01_061119	Modified TO-15	8.5 "Hg	5 psi
02A	IAF-34380CAPITOL-02_061119	Modified TO-15	8.0 "Hg	5 psi
02B	IAF-34380CAPITOL-02_061119	Modified TO-15	8.0 "Hg	5 psi
03A	DUP-34380CAPITOL-01_061119	Modified TO-15	6.5 "Hg	5 psi
04A	IAB-34380CAPITOL-03_061119	Modified TO-15	6.0 "Hg	5 psi
05A	IAG-34380CAPITOL-01_061119	Modified TO-15	8.5 "Hg	5 psi
06A	Lab Blank	Modified TO-15	NA	NA
06B	Lab Blank	Modified TO-15	NA	NA
07A	CCV	Modified TO-15	NA	NA
07B	CCV	Modified TO-15	NA	NA
08A	LCS	Modified TO-15	NA	NA
08AA	LCSD	Modified TO-15	NA	NA
08B	LCS	Modified TO-15	NA	NA
08BB	LCSD	Modified TO-15	NA	NA

CERTIFIED BY:   
 Technical Director

DATE: 09/26/19

Certification numbers: AZ Licensure AZ0775, FL NELAP – E87680, LA NELAP – 02089, NH NELAP - 209218, NJ NELAP - CA016, NY NELAP - 11291, TX NELAP - T104704434-18-13, UT NELAP – CA009332019-11, VA NELAP - 460197, WA NELAP - C935

Name of Accreditation Body: NELAP/ORELAP (Oregon Environmental Laboratory Accreditation Program)  
 Accreditation number: CA300005-011, Effective date: 10/18/2018, Expiration date: 10/17/2019.

Eurofins Air Toxics Inc.. certifies that the test results contained in this report meet all requirements of the NELAC standards

This report shall not be reproduced, except in full, without the written approval of Eurofins Air Toxics, Inc.  
 180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630  
 (916) 985-1000 . (800) 985-5955 . FAX (916) 985-1020

**LABORATORY NARRATIVE**  
**Modified TO-15 Low Level**  
**Arcadis U.S., Inc.**  
**Workorder# 1906314R1**

Five 6 Liter Summa Canister (100% Cert Ambient) samples were received on June 17, 2019. The laboratory performed analysis via modified EPA Method TO-15 using GC/MS in the full scan mode.

This workorder was independently validated prior to submittal using 'USEPA National Functional Guidelines' as generally applied to the analysis of volatile organic compounds in air. A rules-based, logic driven, independent validation engine was employed to assess completeness, evaluate pass/fail of relevant project quality control requirements and verification of all quantified amounts.

Method modifications taken to run these samples are summarized in the table below. Specific project requirements may over-ride the ATL modifications.

<i>Requirement</i>	<i>TO-15</i>	<i>ATL Modifications</i>
Initial Calibration	</=30% RSD with 2 compounds allowed out to < 40% RSD	</=30% RSD with 4 compounds allowed out to < 40% RSD
Blank and standards	Zero Air	UHP Nitrogen provides a higher purity gas matrix than zero air

**Receiving Notes**

There were no receiving discrepancies.

The work order was reissued on 9/26/2019 to change identification of samples AA-34380CAPITOL-01\_061119, IAF-34380CAPITOL-02\_061119, IAF-34380CAPITOL-02\_061119, DUP-34380CAPITOL-01\_061119, IAB-34380CAPITOL-03\_061119 and IAG-34380CAPITOL-01\_061119 per the revised Chain of Custody (COC) provided by the client.

**Analytical Notes**

As per client project requirements, the laboratory has reported estimated values for target compound hits that are below the Reporting Limit but greater than the Method Detection Limit. Concentrations that are below the level at which the canister was certified may be false positives.

The results for sample IAF-34380CAPITOL-02\_061119 in this report was acquired from two separate data files originating from the same analytical run. The two data files have the same base file name and are differentiated with a "sim" extension on the SIM data file.

Dilution was performed on sample IAF-34380CAPITOL-02\_061119 due to the presence of high level non-target species.

**Definition of Data Qualifying Flags**

Eight qualifiers may have been used on the data analysis sheets and indicates as follows:

B - Compound present in laboratory blank greater than reporting limit (background subtraction not performed).

J - Estimated value.

E - Exceeds instrument calibration range.

S - Saturated peak.

Q - Exceeds quality control limits.

U - Compound analyzed for but not detected above the reporting limit, LOD, or MDL value. See data page for project specific U-flag definition.

UJ- Non-detected compound associated with low bias in the CCV

N - The identification is based on presumptive evidence.

File extensions may have been used on the data analysis sheets and indicates as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	AA-34380CAPITOL-01_061119	<b>Date/Time Analyzed:</b>	6/18/19 02:20 PM
<b>Lab ID:</b>	1906314R1-01A	<b>Dilution Factor:</b>	1.87
<b>Date/Time Collected:</b>	6/12/19 07:12 PM	<b>Instrument/Filename:</b>	msdv.i / v061810
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier)		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.36	0.67	0.74	Not Detected
1,4-Dioxane	123-91-1	0.39	0.61	0.67	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.45	0.67	0.74	Not Detected
Tetrachloroethene	127-18-4	0.63	1.1	1.3	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.63	0.67	0.74	Not Detected
Trichloroethene	79-01-6	0.46	0.90	1.0	Not Detected
Vinyl Chloride	75-01-4	0.36	0.43	0.48	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	95
4-Bromofluorobenzene	460-00-4	70-130	98
Toluene-d8	2037-26-5	70-130	101



MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAF-34380CAPITOL-02_061119	<b>Date/Time Analyzed:</b>	6/18/19 02:58 PM
<b>Lab ID:</b>	1906314R1-02A	<b>Dilution Factor:</b>	1.83
<b>Date/Time Collected:</b>	6/12/19 07:05 PM	<b>Instrument/Filename:</b>	msdv.i / v061811
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.35	0.65	0.72	Not Detected
1,4-Dioxane	123-91-1	0.38	0.59	0.66	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.44	0.65	0.72	Not Detected
Tetrachloroethene	127-18-4	0.62	1.1	1.2	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.62	0.65	0.72	Not Detected
Vinyl Chloride	75-01-4	0.35	0.42	0.47	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	94
4-Bromofluorobenzene	460-00-4	70-130	105
Toluene-d8	2037-26-5	70-130	101

MODIFIED EPA METHOD TO-15 GC/MS SIM  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAF-34380CAPITOL-02_061119	<b>Date/Time Analyzed:</b>	6/18/19 02:58 PM
<b>Lab ID:</b>	1906314R1-02B	<b>Dilution Factor:</b>	1.83
<b>Date/Time Collected:</b>	6/12/19 07:05 PM	<b>Instrument/Filename:</b>	msdv.i / v061811sim
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier)		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Trichloroethene	79-01-6	0.13	0.059	0.20	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	95
4-Bromofluorobenzene	460-00-4	70-130	103
Toluene-d8	2037-26-5	70-130	100

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	DUP-34380CAPITOL-01_061119	<b>Date/Time Analyzed:</b>	6/18/19 03:36 PM
<b>Lab ID:</b>	1906314R1-03A	<b>Dilution Factor:</b>	1.71
<b>Date/Time Collected:</b>	6/12/19 07:07 PM	<b>Instrument/Filename:</b>	msdv.i / v061812
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.33	0.61	0.68	Not Detected
1,4-Dioxane	123-91-1	0.36	0.55	0.62	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.41	0.61	0.68	Not Detected
Tetrachloroethene	127-18-4	0.58	1.0	1.2	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.57	0.61	0.68	2.5
Trichloroethene	79-01-6	0.42	0.83	0.92	Not Detected
Vinyl Chloride	75-01-4	0.33	0.39	0.44	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	96
4-Bromofluorobenzene	460-00-4	70-130	109
Toluene-d8	2037-26-5	70-130	97

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAB-34380CAPITOL-03_061119	<b>Date/Time Analyzed:</b>	6/18/19 04:14 PM
<b>Lab ID:</b>	1906314R1-04A	<b>Dilution Factor:</b>	1.68
<b>Date/Time Collected:</b>	6/12/19 07:07 PM	<b>Instrument/Filename:</b>	msdv.i / v061813
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.32	0.60	0.67	Not Detected
1,4-Dioxane	123-91-1	0.35	0.54	0.60	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.40	0.60	0.67	Not Detected
Tetrachloroethene	127-18-4	0.57	1.0	1.1	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.56	0.60	0.67	2.1
Trichloroethene	79-01-6	0.41	0.81	0.90	Not Detected
Vinyl Chloride	75-01-4	0.32	0.39	0.43	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	91
4-Bromofluorobenzene	460-00-4	70-130	108
Toluene-d8	2037-26-5	70-130	99

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAG-34380CAPITOL-01_061119	<b>Date/Time Analyzed:</b>	6/18/19 04:53 PM
<b>Lab ID:</b>	1906314R1-05A	<b>Dilution Factor:</b>	1.87
<b>Date/Time Collected:</b>	6/12/19 07:09 PM	<b>Instrument/Filename:</b>	msdv.i / v061814
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.36	0.67	0.74	Not Detected
1,4-Dioxane	123-91-1	0.39	0.61	0.67	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.45	0.67	0.74	Not Detected
Tetrachloroethene	127-18-4	0.63	1.1	1.3	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.63	0.67	0.74	Not Detected
Trichloroethene	79-01-6	0.46	0.90	1.0	Not Detected
Vinyl Chloride	75-01-4	0.36	0.43	0.48	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	93
4-Bromofluorobenzene	460-00-4	70-130	100
Toluene-d8	2037-26-5	70-130	98

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	Lab Blank	<b>Date/Time Analyzed:</b>	6/18/19 11:15 AM
<b>Lab ID:</b>	1906314R1-06A	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061806a
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.19	0.36	0.40	Not Detected
1,4-Dioxane	123-91-1	0.21	0.32	0.36	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.24	0.36	0.40	Not Detected
Tetrachloroethene	127-18-4	0.34	0.61	0.68	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.34	0.36	0.40	Not Detected
Trichloroethene	79-01-6	0.25	0.48	0.54	Not Detected
Vinyl Chloride	75-01-4	0.19	0.23	0.26	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	94
4-Bromofluorobenzene	460-00-4	70-130	96
Toluene-d8	2037-26-5	70-130	102

MODIFIED EPA METHOD TO-15 GC/MS SIM  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	Lab Blank	<b>Date/Time Analyzed:</b>	6/18/19 11:15 AM
<b>Lab ID:</b>	1906314R1-06B	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061806simc
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Trichloroethene	79-01-6	0.073	0.032	0.11	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	98
4-Bromofluorobenzene	460-00-4	70-130	98
Toluene-d8	2037-26-5	70-130	103

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	CCV	<b>Date/Time Analyzed:</b>	6/18/19 08:39 AM
<b>Lab ID:</b>	1906314R1-07A	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061802
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
1,1-Dichloroethene	75-35-4	96
1,4-Dioxane	123-91-1	100
cis-1,2-Dichloroethene	156-59-2	95
Tetrachloroethene	127-18-4	95
trans-1,2-Dichloroethene	156-60-5	93
Trichloroethene	79-01-6	92
Vinyl Chloride	75-01-4	92

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	95
4-Bromofluorobenzene	460-00-4	70-130	103
Toluene-d8	2037-26-5	70-130	99



MODIFIED EPA METHOD TO-15 GC/MS SIM  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	CCV	<b>Date/Time Analyzed:</b>	6/18/19 08:39 AM
<b>Lab ID:</b>	1906314R1-07B	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061802sim
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
Trichloroethene	79-01-6	89

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	93
4-Bromofluorobenzene	460-00-4	70-130	103
Toluene-d8	2037-26-5	70-130	98

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	LCS	<b>Date/Time Analyzed:</b>	6/18/19 09:22 AM
<b>Lab ID:</b>	1906314R1-08A	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061803
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
1,1-Dichloroethene	75-35-4	95
1,4-Dioxane	123-91-1	104
cis-1,2-Dichloroethene	156-59-2	103
Tetrachloroethene	127-18-4	97
trans-1,2-Dichloroethene	156-60-5	79
Trichloroethene	79-01-6	92
Vinyl Chloride	75-01-4	93

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	94
4-Bromofluorobenzene	460-00-4	70-130	100
Toluene-d8	2037-26-5	70-130	100

\* % Recovery is calculated using unrounded analytical results.

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	LCSD	<b>Date/Time Analyzed:</b>	6/18/19 10:00 AM
<b>Lab ID:</b>	1906314R1-08AA	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061804
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
1,1-Dichloroethene	75-35-4	98
1,4-Dioxane	123-91-1	106
cis-1,2-Dichloroethene	156-59-2	108
Tetrachloroethene	127-18-4	95
trans-1,2-Dichloroethene	156-60-5	82
Trichloroethene	79-01-6	93
Vinyl Chloride	75-01-4	95

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	98
4-Bromofluorobenzene	460-00-4	70-130	102
Toluene-d8	2037-26-5	70-130	99

\* % Recovery is calculated using unrounded analytical results.

MODIFIED EPA METHOD TO-15 GC/MS SIM  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	LCS	<b>Date/Time Analyzed:</b>	6/18/19 09:22 AM
<b>Lab ID:</b>	1906314R1-08B	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061803sim
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
Trichloroethene	79-01-6	89

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	94
4-Bromofluorobenzene	460-00-4	70-130	100
Toluene-d8	2037-26-5	70-130	99

\* % Recovery is calculated using unrounded analytical results.

MODIFIED EPA METHOD TO-15 GC/MS SIM  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	LCSD	<b>Date/Time Analyzed:</b>	6/18/19 10:00 AM
<b>Lab ID:</b>	1906314R1-08BB	<b>Dilution Factor:</b>	1.00
<b>Date/Time Collected:</b>	NA - Not Applicable	<b>Instrument/Filename:</b>	msdv.i / v061804sim
<b>Media:</b>	NA - Not Applicable		

Compound	CAS#	%Recovery
Trichloroethene	79-01-6	89

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	94
4-Bromofluorobenzene	460-00-4	70-130	102
Toluene-d8	2037-26-5	70-130	98

\* % Recovery is calculated using unrounded analytical results.



# DATA VERIFICATION REPORT

REVISED REPORT: September 26, 2019

REVISION SUMMARY: Sample ID's updated per revised COC per client request.

Kris Hinskey  
Arcadis Inc  
10559 Citation Ave  
Suite 100  
Brighton, MI 48116

CADENA project ID: E203631

Project: Ford Livonia Transmission Project - OFF-SITE - Soil Gas and Groundwater

Project number: MI001454.0002/3/4.00002/2B/3B

Event Specific Scope of Work References: Sample COC

Laboratory: Eurofins-California

Laboratory submittal: 1906314

Sample date: 2019-06-12

Report received by CADENA: 2019-09-26

Initial Data Verification completed by CADENA: 2019-06-24

Number of Samples: 5

Sample Matrices: Air

Test Categories: TO-15 GCMS VOC

**Please see attached criteria report or sample result/qualified analytical result summary for qualifier flags assigned to sample data.**

There were no significant QC anomalies or exceptions to report.

Sample Surrogate Recovery, Blank/LCS Surrogate Recovery, LCS/LCD Recovery, Blank Contamination and Hold Time Exception were reviewed as part of our verification.

Data verification for the report specified above was completed using the Ford Motor Company Environmental Laboratory Technical Specification, the CADENA Standard Operating Procedure for the Verification of Environmental Analytical Data and the associated analytical methods as references for evaluating the batch QC, sample data and report content. The EPA National Functional Guidelines for validating organic and inorganic data were used as guidance when addressing out of control QC results and the associated data qualifiers.

Analytical results reported between RDL and MDL are flagged 'J' and considered estimated values.

The definitions of the qualifiers used for this data package are defined in the analytical report. CADENA valid qualifiers are defined in the table below. To view and download a PDF copy of the laboratory analytical report access the CADENA CLMS at <http://clms.cadenaco.com/index.cfm>.

Please contact me if you have any questions.

Sincerely,

Jim Tomalia, Project Scientist

## CADENA Valid Qualifiers

Valid Qualifiers	Description
<	Less than the reported concentration.
>	Greater than the reported concentration.
B	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was greater than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the reported concentration. For Inorganic methods the sample concentration was greater than the RDL and less than 10x the blank concentration and is considered non-detect at the reported concentration.
E	The analyte / Compound reported exceeds the calibration range and is considered estimated.
EMPC	Estimated Minimum Potential Contamination - Dioxin/Furan analyses only.
J	Indicates an estimated value. This flag is used either when estimating a concentration for a tentatively identified compound or when the data indicates the presence of an analyte / compound but the result is less than the sample Quantitation limit, but greater than zero. The flag is also used in data validation to indicate a reported value should be considered estimated due to associated quality assurance deficiencies.
J-	The result is an estimated quantity, but the result may be biased low.
JB	NON-DETECT AT THE CONCENTRATION REPORTED AND ESTIMATED
JH	The sample result is considered estimated and is potentially biased high.
JL	The sample result is considered estimated and is potentially biased low.
JUB	NON-DETECT AT THE REPORTING LIMIT AND ESTIMATED
NJ	Tentatively identified compound with approximated concentration.
R	Indicates the value is considered to be unusable. (Note: The analyte / compound may or may not be present.)
TNTC	Too Numerous to Count - Asbestos and Microbiological Results.
U	Indicates that the analyte / compound was analyzed for, but not detected.
UB	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was less than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the RDL. For Inorganic methods the sample concentration was less than the RDL and less than 10x the blank concentration and is considered non-detect at the RDL.
UJ	The analyte / compound was not detected above the reported sample Quantitation limit. However, the Quantitation limit is considered to be approximate due to associated quality assurance results and may or may not represent the actual limit of Quantitation to accurately and precisely report the analyte in the sample.

# Ford Motor Company – Livonia Transmission Project

## DATA REVIEW

### Livonia, Michigan

Volatile Organic Compounds (VOC) TO-15 Analysis

SDG #1906314R1

CADENA Verification Report: 2019-06-24

Analyses Performed By:  
Eurofins Air Toxics  
Folsom, California

Report #33404R rev1  
Review Level: Tier III  
Project: 30016346.00003 (MI001454.0004.00002)



## DATA REVIEW

### SUMMARY

This data quality assessment summarizes the review of Sample Delivery Group (SDG) # 1906314R1 for samples collected in association with the Ford – Livonia, Michigan site. The review was conducted as a Tier III validation in addition to a verification/Tier II validation review performed by CADENA Inc. and included review of level IV laboratory data package completeness. Only elements of a Tier III validation effort (Tier III includes a detailed review of laboratory raw data to check for errors in calculation, calibration review, internal standard review and compound identification) and omitted deviations from the CADENA verification/Tier II report are documented in this report. Only analytical data associated with constituents of concern were reviewed for this validation. Field documentation was not included in this review. Included with this assessment are the validation annotated sample result sheets, and chain of custody. Analyses were performed on the following samples:

SDG	Sample ID	Lab ID	Matrix	Sample Collection Date	Parent Sample	Analysis		
						TO-15 (Full Scan)	TO-15 (SIM)	MISC
1906314R1	AA-34380CAPITOL-01_061119	1906314R1-01A	Air	6/12/2019		X		
	IAF-34380CAPITOL-02_061119	1906314R1-02B	Air	6/12/2019		X	X	
	DUP-34380CAPITOL-01_061119	1906314R1-03A	Air	6/12/2019	IAB-34380CAPITOL-03_061119	X		
	IAB-34380CAPITOL-03_061119	1906314R1-04A	Air	6/12/2019		X		
	IAG-34380CAPITOL-01_061119	1906314R1-05A	Air	6/12/2019		X		

Note: This report was revised to incorporate the correct sample address designation of “CAPITOL” (Original report was CAPITAL) into the sample IDs.

## DATA REVIEW

### ANALYTICAL DATA PACKAGE DOCUMENTATION

The table below is the evaluation of the data package completeness.

Items Reviewed	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
1. Sample receipt condition		X		X	
2. Requested analyses and sample results		X		X	
3. Master tracking list		X		X	
4. Methods of analysis		X		X	
5. Reporting limits		X		X	
6. Sample collection date		X		X	
7. Laboratory sample received date		X		X	
8. Sample preservation verification (as applicable)		X		X	
9. Sample preparation/extraction/analysis dates		X		X	
10. Fully executed Chain-of-Custody (COC) form		X		X	
11. Narrative summary of Quality Assurance or sample problems provided		X		X	
12. Data Package Completeness and Compliance		X		X	

## DATA REVIEW

### ORGANIC ANALYSIS INTRODUCTION

Analyses were performed according to United States Environmental Protection Agency (USEPA) Method TO-15 (Full Scan) and TO-15-SIM. Data were reviewed in accordance with USEPA National Functional Guidelines of October 1999.

The data review process is an evaluation of data on a technical basis rather than a determination of contract compliance. As such, the standards against which the data are being weighed may differ from those specified in the analytical method. It is assumed that the data package represents the best efforts of the laboratory and had already been subjected to adequate and sufficient quality review prior to submission.

During the review process, laboratory qualified and unqualified data are verified against the supporting documentation. Based on this evaluation, qualifier codes may be added, deleted, or modified by the data reviewer. Results are qualified with the following codes in accordance with USEPA National Functional Guidelines:

- Concentration (C) Qualifiers
  - U The analyte was analyzed for but was not detected above the level of the reported sample quantitation limit.
  - B The compound has been found in the sample as well as its associated blank, its presence in the sample may be suspect.
- Quantitation (Q) Qualifiers
  - E The compound was quantitated above the calibration range.
  - D Concentration is based on a diluted sample analysis.
- Validation Qualifiers
  - J The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
  - UJ The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
  - J+ The result is an estimated quantity, but the result may be biased high.
  - J- The result is an estimated quantity, but the result may be biased low.
  - UB Analyte considered non-detect at the listed value due to associated blank contamination.
  - N The analysis indicates the presence of a compound for which there is presumptive evidence to make a tentative identification.
  - R The sample results are rejected.

Two facts should be noted by all data users. First, the "R" flag means that the associated value is unusable. In other words, due to significant quality control (QC) problems, the analysis is invalid and provides no information as to whether the compound is present or not. "R" values should not appear on data tables because they cannot be relied upon, even as a last resort. The second fact to keep in mind is that no compound concentration, even if it has passed all QC tests, is guaranteed to be accurate. Strict QC serves to increase confidence in data but any value potentially contains error.

## DATA REVIEW

### VOLATILE ORGANIC COMPOUND (VOC) ANALYSES

#### 1. Holding Times

The specified holding times for the following methods are presented in the following table.

Method	Matrix	Holding Time	Preservation	Return Canister Pressure
USEPA TO-15 and USEPA TO-15-SIM	Air	30 days from collection to analysis (Canister)	Ambient Temperature	< -2" Hg

All samples were analyzed within the specified holding time and canister return pressure / vacuum criteria.

#### 2. Mass Spectrometer Tuning

Mass spectrometer performance was acceptable and all analyses were performed within a 12-hour tune clock.

System performance and column resolution were acceptable.

#### 3. Calibration

Satisfactory instrument calibration is established to ensure that the instrument is capable of producing acceptable quantitative data. An initial calibration demonstrates that the instrument is capable of acceptable performance at the beginning of an experimental sequence. The continuing calibration verifies that the instrument daily performance is satisfactory.

##### 3.1 Initial Calibration

The method specifies percent relative standard deviation (%RSD) and relative response factor (RRF) limits for select compounds only. A technical review of the data applies limits to all compounds with no exceptions.

All target compounds associated with the initial calibration standards must exhibit a %RSD less than the control limit (30%) or a correlation coefficient greater than 0.99 and an RRF value greater than control limit (0.05).

All compounds associated with the initial calibrations were within the specified control limits.

##### 3.2 Continuing Calibration

All target compounds associated with the continuing calibration standard must exhibit a percent difference (%D) less than the control limit (30%) and RRF value greater than control limit (0.05).

All compounds associated with the continuing calibrations were within the specified control limits.

#### 4. Internal Standard Performance

Internal standard performance criteria ensure that the GC/MS sensitivity and response are stable during every sample analysis. The criteria requires the internal standard compounds associated with the VOC exhibit area counts that are not greater than 140% or less than 60% of the area counts of the associated continuing calibration standard.

All internal standard responses were within control limits.

## DATA REVIEW

### 5. Compound Identification

Compounds are identified on the GC/MS by using the analytes relative retention time and ion spectra. All identified compounds met the specified criteria.

### 6. Field Duplicate Sample Analysis

The field duplicate analysis is used to assess the precision of the field sampling procedures and analytical method. A control limit of 35% for air matrices is applied to the RPD between the parent sample and the field duplicate. In the instance when the parent and/or duplicate sample concentrations are not greater than five times the RL, a control limit of one times the RL is applied to the difference between the duplicate sample results.

Results (in  $\mu\text{g}/\text{m}^3$ ) for the field duplicate samples are summarized in the following table.

Sample ID / Duplicate ID	Compound	Sample Result	Duplicate Result	RPD
IAB-34380CAPITOL-03_061119/ DUP-34380CAPITOL-01_061119	trans-1,2-Dichloroethene	2.1	2.5	AC

AC Acceptable

The calculated RPDs between the parent sample and field duplicate were acceptable.

### 7. System Performance and Overall Assessment

Overall system performance was acceptable. Other than for those deviations specifically mentioned in this review, the overall data quality is within the guidelines specified in the method.

**DATA REVIEW**

**DATA VALIDATION CHECKLIST FOR VOCs**

VOCs: TO-15 ( Full Scan) and TO-15 SIM	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
<b>GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS)</b>					
<b>Tier II Validation</b>					
Canister return pressure (<-2"Hg)		X		X	
<b>Tier III Validation</b>					
System performance and column resolution		X		X	
Initial calibration %RSDs		X		X	
Continuing calibration RRFs		X		X	
Continuing calibration %Ds		X		X	
Instrument tune and performance check		X		X	
Ion abundance criteria for each instrument used		X		X	
Internal standard		X		X	
Field Duplicate Sample RPD		X		X	
Compound identification and quantitation					
A. Reconstructed ion chromatograms		X		X	
B. Quantitation Reports		X		X	
C. RT of sample compounds within the established RT windows		X		X	
D. Transcription/calculation errors present		X		X	
E. Reporting limits adjusted to reflect sample dilutions		X		X	

Notes:

%RSD Relative standard deviation

%R Percent recovery

RPD Relative percent difference

%D Percent difference

VALIDATION PERFORMED BY: Joseph C. Houser


SIGNATURE:



DATE: July 11, 2019

PEER REVIEW: Dennis Capria

DATE: July 11, 2019



**CHAIN OF CUSTODY  
CORRECTED SAMPLE ANALYSIS DATA  
SHEETS**





**NO CORRECTIONS/QUALIFIERS ADDED  
TO SAMPLE ANALYSIS DATA SHEETS**



MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	AA-34380CAPITOL-01_061119	<b>Date/Time Analyzed:</b>	6/18/19 02:20 PM
<b>Lab ID:</b>	1906314R1-01A	<b>Dilution Factor:</b>	1.87
<b>Date/Time Collected:</b>	6/12/19 07:12 PM	<b>Instrument/Filename:</b>	msdv.i / v061810
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier)		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.36	0.67	0.74	Not Detected
1,4-Dioxane	123-91-1	0.39	0.61	0.67	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.45	0.67	0.74	Not Detected
Tetrachloroethene	127-18-4	0.63	1.1	1.3	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.63	0.67	0.74	Not Detected
Trichloroethene	79-01-6	0.46	0.90	1.0	Not Detected
Vinyl Chloride	75-01-4	0.36	0.43	0.48	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	95
4-Bromofluorobenzene	460-00-4	70-130	98
Toluene-d8	2037-26-5	70-130	101

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAF-34380CAPITOL-02_061119	<b>Date/Time Analyzed:</b>	6/18/19 02:58 PM
<b>Lab ID:</b>	1906314R1-02A	<b>Dilution Factor:</b>	1.83
<b>Date/Time Collected:</b>	6/12/19 07:05 PM	<b>Instrument/Filename:</b>	msdv.i / v061811
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.35	0.65	0.72	Not Detected
1,4-Dioxane	123-91-1	0.38	0.59	0.66	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.44	0.65	0.72	Not Detected
Tetrachloroethene	127-18-4	0.62	1.1	1.2	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.62	0.65	0.72	Not Detected
Vinyl Chloride	75-01-4	0.35	0.42	0.47	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	94
4-Bromofluorobenzene	460-00-4	70-130	105
Toluene-d8	2037-26-5	70-130	101

MODIFIED EPA METHOD TO-15 GC/MS SIM  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAF-34380CAPITOL-02_061119	<b>Date/Time Analyzed:</b>	6/18/19 02:58 PM
<b>Lab ID:</b>	1906314R1-02B	<b>Dilution Factor:</b>	1.83
<b>Date/Time Collected:</b>	6/12/19 07:05 PM	<b>Instrument/Filename:</b>	msdv.i / v061811sim
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Trichloroethene	79-01-6	0.13	0.059	0.20	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	95
4-Bromofluorobenzene	460-00-4	70-130	103
Toluene-d8	2037-26-5	70-130	100

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	DUP-34380CAPITOL-01_061119	<b>Date/Time Analyzed:</b>	6/18/19 03:36 PM
<b>Lab ID:</b>	1906314R1-03A	<b>Dilution Factor:</b>	1.71
<b>Date/Time Collected:</b>	6/12/19 07:07 PM	<b>Instrument/Filename:</b>	msdv.i / v061812
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.33	0.61	0.68	Not Detected
1,4-Dioxane	123-91-1	0.36	0.55	0.62	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.41	0.61	0.68	Not Detected
Tetrachloroethene	127-18-4	0.58	1.0	1.2	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.57	0.61	0.68	2.5
Trichloroethene	79-01-6	0.42	0.83	0.92	Not Detected
Vinyl Chloride	75-01-4	0.33	0.39	0.44	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	96
4-Bromofluorobenzene	460-00-4	70-130	109
Toluene-d8	2037-26-5	70-130	97

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAB-34380CAPITOL-03_061119	<b>Date/Time Analyzed:</b>	6/18/19 04:14 PM
<b>Lab ID:</b>	1906314R1-04A	<b>Dilution Factor:</b>	1.68
<b>Date/Time Collected:</b>	6/12/19 07:07 PM	<b>Instrument/Filename:</b>	msdv.i / v061813
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.32	0.60	0.67	Not Detected
1,4-Dioxane	123-91-1	0.35	0.54	0.60	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.40	0.60	0.67	Not Detected
Tetrachloroethene	127-18-4	0.57	1.0	1.1	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.56	0.60	0.67	2.1
Trichloroethene	79-01-6	0.41	0.81	0.90	Not Detected
Vinyl Chloride	75-01-4	0.32	0.39	0.43	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	91
4-Bromofluorobenzene	460-00-4	70-130	108
Toluene-d8	2037-26-5	70-130	99

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN  
 Ford LTP Off-Site Sampling

<b>Client ID:</b>	IAG-34380CAPITOL-01_061119	<b>Date/Time Analyzed:</b>	6/18/19 04:53 PM
<b>Lab ID:</b>	1906314R1-05A	<b>Dilution Factor:</b>	1.87
<b>Date/Time Collected:</b>	6/12/19 07:09 PM	<b>Instrument/Filename:</b>	msdv.i / v061814
<b>Media:</b>	6 Liter Summa Canister (100% Cert Ambier		

Compound	CAS#	MDL (ug/m3)	LOD (ug/m3)	Rpt. Limit (ug/m3)	Amount (ug/m3)
1,1-Dichloroethene	75-35-4	0.36	0.67	0.74	Not Detected
1,4-Dioxane	123-91-1	0.39	0.61	0.67	Not Detected
cis-1,2-Dichloroethene	156-59-2	0.45	0.67	0.74	Not Detected
Tetrachloroethene	127-18-4	0.63	1.1	1.3	Not Detected
trans-1,2-Dichloroethene	156-60-5	0.63	0.67	0.74	Not Detected
Trichloroethene	79-01-6	0.46	0.90	1.0	Not Detected
Vinyl Chloride	75-01-4	0.36	0.43	0.48	Not Detected

D: Analyte not within the DoD scope of accreditation.

Surrogates	CAS#	Limits	%Recovery
1,2-Dichloroethane-d4	17060-07-0	70-130	93
4-Bromofluorobenzene	460-00-4	70-130	100
Toluene-d8	2037-26-5	70-130	98

### Analysis Request /Canister Chain of Custody

For Laboratory Use Only

180 Blue Ravine Rd. Suite B, Folsom, CA 95630  
 Phone (800) 985-5955; Fax (916) 351-8279

P.O. # \_\_\_\_\_

Workorder# \_\_\_\_\_

Page 1 of 1

Client: Arcadis		P.O. #		Special Instructions/Notes: Report ONLY: 1,1-DCE, cis-1,2-DCE, trans-1,2-DCE, 1,4-Dioxane, PCE, TCE and VC. Submit results through Cadena at jim.bentall@cadenalabs.com. Cadena #E203631 Level IV Reporting				Turnaround Time (Rush surcharges may apply)										
Project Name: Ford LTP GH Site Sampling		P.O. # M001454 0003 00003																
Project Manager: Kris Hmskey		P.O. #		5 Day Turnaround Time				Requested Analysis										
Sampler: Shantel Johnson		P.O. #																
Site Name: 34380 CAPITOL		P.O. #		Canister Vacuum/Pressure				Lab Use Only										
Lab ID		Sample Identification										Start Sampling Information		Stop Sampling Information		Initial (in Hg)		Final (in Hg)
	AA-34380CAPITOL_01_061119	6L2358	22154	06/11/2019	19:12	06/12/2019	19:09	-29	-7									X
	IAF-34380CAPITOL_02_061119	6L0740	30814	06/11/2019	19:05	06/12/2019	18:39	-29	-7									X
	DUP-34380CAPITOL_01_061119	6L0427	22605	06/11/2019	19:07	06/12/2019	16:02	-29	-6.5									X
	IAF-34380CAPITOL_03_061119	6L0480	21936	06/11/2019	19:07	06/12/2019	18:02	-29	-8									X
	IAF-34380CAPITOL_01_061119	6L2407	20408	06/11/2019	19:09	06/12/2019	18:08	-29	-7									X
Relinquished by (Signature/Affiliation)				Date	Time	Received by (Signature/Affiliation)				Date	Time							
Relinquished by (Signature/Affiliation)				Date	Time	Received by (Signature/Affiliation)				Date	Time							
Relinquished by (Signature/Affiliation)				Date	Time	Received by (Signature/Affiliation)				Date	Time							
Lab Use Only																		
Shipper Name		Custody Seals Intact?		Yes		No		None										
Sample Transportation Notice: Relinquishing signature on this document indicates that samples are shipped in compliance with all applicable local, State, Federal, and international laws, regulations, and ordinances of any kind. Relinquishing signature also indicates agreement to hold harmless, defend, and indemnify Eurofins Air Toxics against any claim, demand, or action, of any kind, related to the collection, handling, or shipping of samples. D.O.T. Hotline (800) 467-4922																		

Revised COC 9/24/19 -AS



# Analysis Request /Canister Chain of Custody

For Laboratory Use Only

## 1906314

PID: \_\_\_\_\_

Workerorder#: \_\_\_\_\_

Page 1 of 1

180 Blue Ravine Rd. Suite B, Folsom, CA 95630

Phone (800) 985-5955; Fax (916) 351-8279

Client:	Arcadis	PID:		<b>Special Instructions/Notes:</b> Report ONLY: 1,1-DCE, cis-1,2-DCE, trans-1,2-DCE, 1,4-Dioxane, PCE, TCE and VC. Submit results through Cadena at jim.tomalia@cadena.com. Cadena #E203631. Level IV Reporting				<b>Turnaround Time (Rush surcharges may apply)</b>												
Project Name:	Ford LTP Off-Site Sampling	P.O.#	MI001454.0003.00002					<b>5 Day Turnaround Time</b>												
Project Manager:	Kris Hinskey			<b>Canister Vacuum/Pressure</b>				<b>Requested Analyses</b>												
Sampler:	Shantel Johnson			<b>Start Sampling Information</b>				<b>Stop Sampling Information</b>												
Site Name:	34380 CAPITOL			<b>Date</b>		<b>Time</b>		<b>Date</b>		<b>Time</b>		<b>Initial (in Hg)</b>		<b>Final (in Hg)</b>		<b>Lab Use Only</b>		<b>TO-15 (See Special Instructions/Notes)</b>		
				<b>Receipt</b>		<b>Final (psig) Gas: N2 / He</b>														
<b>Lab ID</b>	<b>Sample Identification</b>	<b>Canister #</b>	<b>Flow Controller #</b>																	
01A	AA-34380CAPITAL-01_061119	6L2368	22154	06/11/2019	19:12	06/12/2019	18:09	-29	-7	8.5" Hg	SP3								X	
02A	IAF-34380CAPITAL-02_061119	6L0740	30814	06/11/2019	19:05	06/12/2019	18:39	-29	-7	8.0" Hg	SP3								X	
03A	DUP-34380CAPITAL-01_061119	6L0427	22605	06/11/2019	19:07	06/12/2019	18:02	-29	-6.5	6.5" Hg	SP3								X	
04A	IAB-34380CAPITAL-03_061119	6L0480	21936	06/11/2019	19:07	06/12/2019	18:02	-29	-6	6.0" Hg	SP3								X	
05A	IAG-34380CAPITAL-01_061119	6L2407	20499	06/11/2019	19:09	06/12/2019	18:08	-29	-7	8.5" Hg	SP3								X	
Relinquished by: (Signature/Affiliation) <i>[Signature]</i>				Date	6-13-19	Time	12:00	Received by: (Signature/Affiliation) <i>[Signature]</i>				Date	6/17/19	Time	09:01					
Relinquished by: (Signature/Affiliation)				Date		Time		Received by: (Signature/Affiliation)				Date		Time						
Relinquished by: (Signature/Affiliation)				Date		Time		Received by: (Signature/Affiliation)				Date		Time						
<b>Lab Use Only</b>																				
Shipper Name: <i>Ford SX</i>		Custody Seals Intact? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> None																		
<p><b>Sample Transportation Notice:</b> Relinquishing signature on this document indicates that samples are shipped in compliance with all applicable local, State, Federal, and international laws, regulations, and ordinances of any kind. Relinquishing signature also indicates agreement to hold harmless, defend, and indemnify Eurofins Air Toxics against any claim, demand, or action, of any kind, related to the collection, handling, of shipping of samples. D.O.T Hotline (800) 467-4922</p>																				

Pressurized by: H/W  
 Date: 06/18/19  
 Nitrogen / Helium