

ANALYTICAL REPORT

Eurofins TestAmerica, Edison
777 New Durham Road
Edison, NJ 08817
Tel: (732)549-3900

Laboratory Job ID: 460-197429-1
Client Project/Site: Ford LTP On-Site

For:
ARCADIS U.S., Inc.
28550 Cabot Drive
Suite 500
Novi, Michigan 48377

Attn: Kristoffer Hinskey



Authorized for release by:
12/8/2019 2:29:01 PM

Michael DelMonico, Project Manager I
(330)497-9396
michael.delmonico@testamericainc.com

LINKS

Review your project
results through
TotalAccess

Have a Question?



Visit us at:
www.testamericainc.com

The test results in this report meet all 2003 NELAC and 2009 TNI requirements for accredited parameters, exceptions are noted in this report. This report may not be reproduced except in full, and with written approval from the laboratory. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Detection Summary	6
Client Sample Results	7
Surrogate Summary	10
QC Sample Results	11
QC Association Summary	14
Lab Chronicle	15
Certification Summary	16
Method Summary	17
Sample Summary	18
Chain of Custody	19
Receipt Checklists	21

Definitions/Glossary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
▫	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Job ID: 460-197429-1

Laboratory: Eurofins TestAmerica, Edison

Narrative

CASE NARRATIVE

Client: ARCADIS U.S., Inc.

Project: Ford LTP On-Site

Report Number: 460-197429-1

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

Eurofins TestAmerica, Edison attests to the validity of the laboratory data generated by Eurofins TestAmerica facilities reported herein. All analyses performed by Eurofins TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the application methods. Eurofins TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

The test results in this report meet all NELAP requirements for parameters for which accreditation is required or available. Any exceptions to NELAP requirements are noted in this report. Pursuant to NELAP, this report may not be reproduced, except in full, without the written approval of the laboratory.

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of Eurofins TestAmerica and its client.

RECEIPT

The samples were received on 11/22/2019 10:15 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperatures of the 2 coolers at receipt time were 4.3° C and 4.7° C.

VOLATILE ORGANIC COMPOUNDS (GCMS)

Samples TRIP BLANK (460-197429-1), MW-5_112019 (460-197429-2), MW-2_112019 (460-197429-3), MW-19_112019 (460-197429-4) and MW-29_112019 (460-197429-5) were analyzed for volatile organic compounds (GCMS) in accordance with EPA SW-846 Method 8260C. The samples were analyzed on 12/01/2019.

The following sample was diluted to bring the concentration of target analytes within the calibration range: MW-2_112019 (460-197429-3). Elevated reporting limits (RLs) are provided.

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

VOLATILE ORGANIC COMPOUNDS (GC/MS)

Samples MW-5_112019 (460-197429-2), MW-2_112019 (460-197429-3), MW-19_112019 (460-197429-4) and MW-29_112019 (460-197429-5) were analyzed for Volatile organic compounds (GC/MS) in accordance with SW-846 Method 8260C SIM. The samples were analyzed on 11/27/2019.

The following sample was diluted to bring the concentration of target analyte within the calibration range: MW-19_112019 (460-197429-4).

Case Narrative

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Job ID: 460-197429-1 (Continued)

Laboratory: Eurofins TestAmerica, Edison (Continued)

Elevated reporting limit (RL) is provided.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

Detection Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 460-197429-1

No Detections.

Client Sample ID: MW-5_112019

Lab Sample ID: 460-197429-2

No Detections.

Client Sample ID: MW-2_112019

Lab Sample ID: 460-197429-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	4.9		2.0	0.33	ug/L	1		8260C SIM	Total/NA
1,1-Dichloroethene	2.8	J	10	2.6	ug/L	10		8260C	Total/NA
cis-1,2-Dichloroethene	3000		10	2.2	ug/L	10		8260C	Total/NA
trans-1,2-Dichloroethene	760		10	2.4	ug/L	10		8260C	Total/NA
Trichloroethene	3.4	J	10	3.1	ug/L	10		8260C	Total/NA
Vinyl chloride	160		10	1.7	ug/L	10		8260C	Total/NA

Client Sample ID: MW-19_112019

Lab Sample ID: 460-197429-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	290		50	8.3	ug/L	25		8260C SIM	Total/NA
cis-1,2-Dichloroethene	1.1		1.0	0.22	ug/L	1		8260C	Total/NA
Trichloroethene	0.70	J	1.0	0.31	ug/L	1		8260C	Total/NA
Vinyl chloride	1.4		1.0	0.17	ug/L	1		8260C	Total/NA

Client Sample ID: MW-29_112019

Lab Sample ID: 460-197429-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	0.60	J	2.0	0.33	ug/L	1		8260C SIM	Total/NA
Vinyl chloride	0.20	J	1.0	0.17	ug/L	1		8260C	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins TestAmerica, Edison

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 460-197429-1

Date Collected: 11/20/19 00:00

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.26	ug/L	-		12/01/19 17:03	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.22	ug/L	-		12/01/19 17:03	1
Tetrachloroethene	1.0	U	1.0	0.25	ug/L	-		12/01/19 17:03	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.24	ug/L	-		12/01/19 17:03	1
Trichloroethene	1.0	U	1.0	0.31	ug/L	-		12/01/19 17:03	1
Vinyl chloride	1.0	U	1.0	0.17	ug/L	-		12/01/19 17:03	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		74 - 132		12/01/19 17:03	1
Toluene-d8 (Surr)	92		80 - 120		12/01/19 17:03	1
Dibromofluoromethane (Surr)	93		72 - 131		12/01/19 17:03	1
4-Bromofluorobenzene	93		77 - 124		12/01/19 17:03	1

Client Sample ID: MW-5_112019

Lab Sample ID: 460-197429-2

Date Collected: 11/20/19 10:37

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.33	ug/L	-		11/27/19 07:08	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	94		72 - 133		11/27/19 07:08	1

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.26	ug/L	-		12/01/19 17:51	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.22	ug/L	-		12/01/19 17:51	1
Tetrachloroethene	1.0	U	1.0	0.25	ug/L	-		12/01/19 17:51	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.24	ug/L	-		12/01/19 17:51	1
Trichloroethene	1.0	U	1.0	0.31	ug/L	-		12/01/19 17:51	1
Vinyl chloride	1.0	U	1.0	0.17	ug/L	-		12/01/19 17:51	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		74 - 132		12/01/19 17:51	1
Toluene-d8 (Surr)	94		80 - 120		12/01/19 17:51	1
Dibromofluoromethane (Surr)	94		72 - 131		12/01/19 17:51	1
4-Bromofluorobenzene	92		77 - 124		12/01/19 17:51	1

Client Sample ID: MW-2_112019

Lab Sample ID: 460-197429-3

Date Collected: 11/20/19 11:37

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	4.9		2.0	0.33	ug/L	-		11/27/19 07:33	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	90		72 - 133		11/27/19 07:33	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Client Sample ID: MW-2_112019

Lab Sample ID: 460-197429-3

Date Collected: 11/20/19 11:37

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	2.8	J	10	2.6	ug/L			12/01/19 21:29	10
cis-1,2-Dichloroethene	3000		10	2.2	ug/L			12/01/19 21:29	10
Tetrachloroethene	10	U	10	2.5	ug/L			12/01/19 21:29	10
trans-1,2-Dichloroethene	760		10	2.4	ug/L			12/01/19 21:29	10
Trichloroethene	3.4	J	10	3.1	ug/L			12/01/19 21:29	10
Vinyl chloride	160		10	1.7	ug/L			12/01/19 21:29	10

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	99		74 - 132		12/01/19 21:29	10
Toluene-d8 (Surr)	107		80 - 120		12/01/19 21:29	10
Dibromofluoromethane (Surr)	103		72 - 131		12/01/19 21:29	10
4-Bromofluorobenzene	108		77 - 124		12/01/19 21:29	10

Client Sample ID: MW-19_112019

Lab Sample ID: 460-197429-4

Date Collected: 11/20/19 12:57

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	290		50	8.3	ug/L			11/27/19 16:47	25

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	98		72 - 133		11/27/19 16:47	25

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.26	ug/L			12/01/19 18:16	1
cis-1,2-Dichloroethene	1.1		1.0	0.22	ug/L			12/01/19 18:16	1
Tetrachloroethene	1.0	U	1.0	0.25	ug/L			12/01/19 18:16	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.24	ug/L			12/01/19 18:16	1
Trichloroethene	0.70	J	1.0	0.31	ug/L			12/01/19 18:16	1
Vinyl chloride	1.4		1.0	0.17	ug/L			12/01/19 18:16	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	105		74 - 132		12/01/19 18:16	1
Toluene-d8 (Surr)	107		80 - 120		12/01/19 18:16	1
Dibromofluoromethane (Surr)	107		72 - 131		12/01/19 18:16	1
4-Bromofluorobenzene	109		77 - 124		12/01/19 18:16	1

Client Sample ID: MW-29_112019

Lab Sample ID: 460-197429-5

Date Collected: 11/20/19 14:17

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	0.60	J	2.0	0.33	ug/L			11/27/19 14:41	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	97		72 - 133		11/27/19 14:41	1

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Client Sample ID: MW-29_112019

Lab Sample ID: 460-197429-5

Date Collected: 11/20/19 14:17

Matrix: Water

Date Received: 11/22/19 10:15

Method: 8260C - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.26	ug/L			12/01/19 18:40	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.22	ug/L			12/01/19 18:40	1
Tetrachloroethene	1.0	U	1.0	0.25	ug/L			12/01/19 18:40	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.24	ug/L			12/01/19 18:40	1
Trichloroethene	1.0	U	1.0	0.31	ug/L			12/01/19 18:40	1
Vinyl chloride	0.20	J	1.0	0.17	ug/L			12/01/19 18:40	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	92		74 - 132		12/01/19 18:40	1
Toluene-d8 (Surr)	94		80 - 120		12/01/19 18:40	1
Dibromofluoromethane (Surr)	97		72 - 131		12/01/19 18:40	1
4-Bromofluorobenzene	96		77 - 124		12/01/19 18:40	1

Surrogate Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Method: 8260C - Volatile Organic Compounds by GC/MS

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA (74-132)	TOL (80-120)	DBFM (72-131)	BFB (77-124)
460-197429-1	TRIP BLANK	91	92	93	93
460-197429-2	MW-5_112019	91	94	94	92
460-197429-3	MW-2_112019	99	107	103	108
460-197429-4	MW-19_112019	105	107	107	109
460-197429-5	MW-29_112019	92	94	97	96
LCS 460-659359/3	Lab Control Sample	97	102	99	97
LCSD 460-659359/4	Lab Control Sample Dup	99	105	106	106
MB 460-659359/8	Method Blank	88	92	90	92

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

TOL = Toluene-d8 (Surr)

DBFM = Dibromofluoromethane (Surr)

BFB = 4-Bromofluorobenzene

Method: 8260C SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB (72-133)
460-197429-2	MW-5_112019	94
460-197429-3	MW-2_112019	90
460-197429-4	MW-19_112019	98
460-197429-5	MW-29_112019	97
LCS 460-658586/4	Lab Control Sample	96
LCS 460-658641/4	Lab Control Sample	97
LCSD 460-658586/5	Lab Control Sample Dup	93
LCSD 460-658641/5	Lab Control Sample Dup	95
MB 460-658586/9	Method Blank	95
MB 460-658641/8	Method Blank	95

Surrogate Legend

BFB = 4-Bromofluorobenzene

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Method: 8260C - Volatile Organic Compounds by GC/MS

Lab Sample ID: MB 460-659359/8
Matrix: Water
Analysis Batch: 659359

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.26	ug/L			12/01/19 16:16	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.22	ug/L			12/01/19 16:16	1
Tetrachloroethene	1.0	U	1.0	0.25	ug/L			12/01/19 16:16	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.24	ug/L			12/01/19 16:16	1
Trichloroethene	1.0	U	1.0	0.31	ug/L			12/01/19 16:16	1
Vinyl chloride	1.0	U	1.0	0.17	ug/L			12/01/19 16:16	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	88		74 - 132		12/01/19 16:16	1
Toluene-d8 (Surr)	92		80 - 120		12/01/19 16:16	1
Dibromofluoromethane (Surr)	90		72 - 131		12/01/19 16:16	1
4-Bromofluorobenzene	92		77 - 124		12/01/19 16:16	1

Lab Sample ID: LCS 460-659359/3
Matrix: Water
Analysis Batch: 659359

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	20.0	22.3		ug/L		112	74 - 123
cis-1,2-Dichloroethene	20.0	21.2		ug/L		106	80 - 120
Tetrachloroethene	20.0	20.5		ug/L		103	78 - 122
trans-1,2-Dichloroethene	20.0	22.1		ug/L		111	79 - 120
Trichloroethene	20.0	20.2		ug/L		101	77 - 120
Vinyl chloride	20.0	22.3		ug/L		111	62 - 138

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	97		74 - 132
Toluene-d8 (Surr)	102		80 - 120
Dibromofluoromethane (Surr)	99		72 - 131
4-Bromofluorobenzene	97		77 - 124

Lab Sample ID: LCSD 460-659359/4
Matrix: Water
Analysis Batch: 659359

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,1-Dichloroethene	20.0	21.7		ug/L		109	74 - 123	3	30
cis-1,2-Dichloroethene	20.0	20.7		ug/L		103	80 - 120	3	30
Tetrachloroethene	20.0	20.9		ug/L		105	78 - 122	2	30
trans-1,2-Dichloroethene	20.0	21.8		ug/L		109	79 - 120	2	30
Trichloroethene	20.0	19.8		ug/L		99	77 - 120	2	30
Vinyl chloride	20.0	22.0		ug/L		110	62 - 138	1	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	99		74 - 132
Toluene-d8 (Surr)	105		80 - 120
Dibromofluoromethane (Surr)	106		72 - 131

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Method: 8260C - Volatile Organic Compounds by GC/MS (Continued)

Lab Sample ID: LCSD 460-659359/4
Matrix: Water
Analysis Batch: 659359

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Surrogate	LCS %Recovery	LCS Qualifier	Limits
4-Bromofluorobenzene	106		77 - 124

Method: 8260C SIM - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 460-658586/9
Matrix: Water
Analysis Batch: 658586

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.33	ug/L			11/27/19 02:56	1
Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac			
4-Bromofluorobenzene	95		72 - 133		11/27/19 02:56	1			

Lab Sample ID: LCS 460-658586/4
Matrix: Water
Analysis Batch: 658586

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	5.00	5.01		ug/L		100	66 - 135
Surrogate	LCS %Recovery	LCS Qualifier	Limits				
4-Bromofluorobenzene	96		72 - 133				

Lab Sample ID: LCSD 460-658586/5
Matrix: Water
Analysis Batch: 658586

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,4-Dioxane	5.00	4.44		ug/L		89	66 - 135	12	30
Surrogate	LCSD %Recovery	LCSD Qualifier	Limits						
4-Bromofluorobenzene	93		72 - 133						

Lab Sample ID: MB 460-658641/8
Matrix: Water
Analysis Batch: 658641

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.33	ug/L			11/27/19 13:26	1
Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac			
4-Bromofluorobenzene	95		72 - 133		11/27/19 13:26	1			

QC Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Method: 8260C SIM - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 460-658641/4
Matrix: Water
Analysis Batch: 658641

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	5.00	4.63		ug/L	-	93	66 - 135
LCS LCS							
Surrogate	%Recovery	Qualifier	Limits				
4-Bromofluorobenzene	97		72 - 133				

Lab Sample ID: LCSD 460-658641/5
Matrix: Water
Analysis Batch: 658641

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,4-Dioxane	5.00	4.06		ug/L	-	81	66 - 135	13	30
LCSD LCSD									
Surrogate	%Recovery	Qualifier	Limits						
4-Bromofluorobenzene	95		72 - 133						



QC Association Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

GC/MS VOA

Analysis Batch: 658586

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
460-197429-2	MW-5_112019	Total/NA	Water	8260C SIM	
460-197429-3	MW-2_112019	Total/NA	Water	8260C SIM	
MB 460-658586/9	Method Blank	Total/NA	Water	8260C SIM	
LCS 460-658586/4	Lab Control Sample	Total/NA	Water	8260C SIM	
LCSD 460-658586/5	Lab Control Sample Dup	Total/NA	Water	8260C SIM	

Analysis Batch: 658641

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
460-197429-4	MW-19_112019	Total/NA	Water	8260C SIM	
460-197429-5	MW-29_112019	Total/NA	Water	8260C SIM	
MB 460-658641/8	Method Blank	Total/NA	Water	8260C SIM	
LCS 460-658641/4	Lab Control Sample	Total/NA	Water	8260C SIM	
LCSD 460-658641/5	Lab Control Sample Dup	Total/NA	Water	8260C SIM	

Analysis Batch: 659359

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
460-197429-1	TRIP BLANK	Total/NA	Water	8260C	
460-197429-2	MW-5_112019	Total/NA	Water	8260C	
460-197429-3	MW-2_112019	Total/NA	Water	8260C	
460-197429-4	MW-19_112019	Total/NA	Water	8260C	
460-197429-5	MW-29_112019	Total/NA	Water	8260C	
MB 460-659359/8	Method Blank	Total/NA	Water	8260C	
LCS 460-659359/3	Lab Control Sample	Total/NA	Water	8260C	
LCSD 460-659359/4	Lab Control Sample Dup	Total/NA	Water	8260C	

Lab Chronicle

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 460-197429-1

Date Collected: 11/20/19 00:00

Matrix: Water

Date Received: 11/22/19 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	659359	12/01/19 17:03	KLB	TAL EDI

Client Sample ID: MW-5_112019

Lab Sample ID: 460-197429-2

Date Collected: 11/20/19 10:37

Matrix: Water

Date Received: 11/22/19 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	659359	12/01/19 17:51	KLB	TAL EDI
Total/NA	Analysis	8260C SIM		1	658586	11/27/19 07:08	KLB	TAL EDI

Client Sample ID: MW-2_112019

Lab Sample ID: 460-197429-3

Date Collected: 11/20/19 11:37

Matrix: Water

Date Received: 11/22/19 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		10	659359	12/01/19 21:29	KLB	TAL EDI
Total/NA	Analysis	8260C SIM		1	658586	11/27/19 07:33	KLB	TAL EDI

Client Sample ID: MW-19_112019

Lab Sample ID: 460-197429-4

Date Collected: 11/20/19 12:57

Matrix: Water

Date Received: 11/22/19 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	659359	12/01/19 18:16	KLB	TAL EDI
Total/NA	Analysis	8260C SIM		25	658641	11/27/19 16:47	KLB	TAL EDI

Client Sample ID: MW-29_112019

Lab Sample ID: 460-197429-5

Date Collected: 11/20/19 14:17

Matrix: Water

Date Received: 11/22/19 10:15

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260C		1	659359	12/01/19 18:40	KLB	TAL EDI
Total/NA	Analysis	8260C SIM		1	658641	11/27/19 14:41	KLB	TAL EDI

Laboratory References:

TAL EDI = Eurofins TestAmerica, Edison, 777 New Durham Road, Edison, NJ 08817, TEL (732)549-3900

Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Laboratory: Eurofins TestAmerica, Edison

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Connecticut	State	PH-0200	09-30-20
DE Haz. Subst. Cleanup Act (HSCA)	State	<cert No.>	12-31-21
Georgia	State	12028 (NJ)	06-30-20
Massachusetts	State	M-NJ312	06-30-20
Massachusetts	State Program	M-NJ312	06-30-20
New Jersey	NELAP	12028	06-30-20
New York	NELAP	11452	04-01-20
Pennsylvania	NELAP	68-00522	02-28-20
Rhode Island	State	LAO00132	12-30-19
USDA	US Federal Programs	P330-18-00135	05-03-21

Laboratory: Eurofins TestAmerica, Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	2927	02-23-20
Connecticut	State	PH-0590	12-31-19
Florida	NELAP	E87225	06-30-20
Georgia	State	4062	02-23-20
Illinois	NELAP	004498	07-31-20
Iowa	State	421	06-01-20
Kansas	NELAP	E-10336	04-30-20
Kentucky (UST)	State	112225	02-23-20
Kentucky (WW)	State	KY98016	12-31-19
Minnesota	NELAP	OH00048	12-31-19
Minnesota (Petrofund)	State Program	3506	07-31-21
New Jersey	NELAP	OH001	06-30-20
New York	NELAP	10975	03-31-20
Ohio VAP	State	CL0024	06-05-21
Oregon	NELAP	4062	02-23-20
Pennsylvania	NELAP	68-00340	08-31-20
Texas	NELAP	T104704517-18-10	08-31-20
USDA	US Federal Programs	P330-16-00404	12-28-19
Virginia	NELAP	010101	09-14-20
Washington	State	C971	01-12-20
West Virginia DEP	State	210	12-31-19

Method Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Method	Method Description	Protocol	Laboratory
8260C	Volatile Organic Compounds by GC/MS	SW846	TAL EDI
8260C SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL EDI
5030C	Purge and Trap	SW846	TAL EDI

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL EDI = Eurofins TestAmerica, Edison, 777 New Durham Road, Edison, NJ 08817, TEL (732)549-3900



Sample Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 460-197429-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Asset ID
460-197429-1	TRIP BLANK	Water	11/20/19 00:00	11/22/19 10:15	
460-197429-2	MW-5_112019	Water	11/20/19 10:37	11/22/19 10:15	
460-197429-3	MW-2_112019	Water	11/20/19 11:37	11/22/19 10:15	
460-197429-4	MW-19_112019	Water	11/20/19 12:57	11/22/19 10:15	
460-197429-5	MW-29_112019	Water	11/20/19 14:17	11/22/19 10:15	

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

TestAmerica Laboratory location: Brighton --- 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

197429

Client Contact
Company Name: Arcadis
Address: 28550 Cabot Drive, Suite 500
City/State/Zip: Novi, MI, 48377
Phone: 248-994-2240
Project Name: Ford LTP On-Site
Project Number: 30016346.0001B
PO # 30016346.0001B

Regulatory program: DW NPDES RCRA Other

Client Project Manager: Kris Hinskey
Telephone: 248-994-2240
Email: kristoffer.hinskey@arcadis.com

Site Contact: Rachel Bielak
Telephone: 248-946-6331
Email: rbielak@arcadis.com

Lab Contact: Mike DeMonico
Telephone: 330-497-9396

Sampler Name: Heather Woodsum
Method of Shipment/Carrier:
Shipping/Tracking No.:

Sample Identification	Sample Date	Sample Time	Analysis											Sample Specific Notes / Special Instructions:				
			Air	Aqueous	Sediment	Solid	Other:	H2SO4	HNO3	HCl	NaOH	ZnAc	NaOH		Upres	Other:		
TRIP BLANK																		1 Trip Blank
MW-5-112019	11/20/19	1037	X															3 VOA POC 8260B 3 VOA POC 8260B SIM
MW-2-112019		1137	X															
MW-19-112019		1257	X															
MW-29-112019		1417	X															



460-197429 Chain of Custody

Possible Hazard Identification
 Non-Hazard Flammable Irritant Poison B Unknown

Special Instructions/QC Requirements & Comments:
Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)
 Return to Client Disposal By Lab Archive For _____ Months

Received by: **Tom Traiker** Company: **Arcadis** Date/Time: **11/20/19 1445**

Received by: **Novi Cab Storage** Company: **Arcadis** Date/Time: **11/20/19 1556**

Received in Laboratory: **Jim Head** Company: **ETA** Date/Time: **11/21/19 1100**

44/47 40/43

11/20/19 1445

11/20/19 1735

11/21-19 1100

11/20/19 1016

11-21-19 e 1309 ETA

©2008 TestAmerica Laboratories, Inc. All rights reserved.
TestAmerica & Design are trademarks of TestAmerica Laboratories, Inc.



Eurofins TestAmerica Edison
Receipt Temperature and pH Log

Job Number: 197429

Number of Coolers: 2 IR Gun # 11

Cooler Temperatures

	RAW	CORRECTED	RAW	CORRECTED	RAW	CORRECTED
Cooler #1:	49	47	Cooler #4:	°C	Cooler #7:	°C
Cooler #2:	40	43	Cooler #5:	°C	Cooler #8:	°C
Cooler #3:	°C	°C	Cooler #6:	°C	Cooler #9:	°C

TALS Sample Number	Ammonia (pH<2)	Nitrate Nitrite (pH<2)	Metals* (pH<2)	Hardness (pH<2)	Pest (pH 5-9)	EPH or OAM (pH<2)	Phenols (pH<2)	Sulfide (pH>9)	TKN (pH<2)	TOC (pH<2)	Total Cyanide (pH>12)	Total Phos (pH<2)	Other	Other

If pH adjustments are required record the information below:

Sample No(s). adjusted: _____

Preservative Name/Conc.: _____ Volume of Preservative used (ml): _____

Lot # of Preservative(s): _____ Expiration Date: _____

The appropriate Project Manager and Department Manager should be notified about the samples which were pH adjusted.
* Samples for Metal analysis which are out of compliance must be acidified at least 24 hours prior to analysis.

Initials: KL Date: 11/22/19



Login Sample Receipt Checklist

Client: ARCADIS U.S., Inc.

Job Number: 460-197429-1

Login Number: 197429

List Number: 1

Creator: Rivera, Kenneth

List Source: Eurofins TestAmerica, Edison

Question	Answer	Comment
Radioactivity wasn't checked or is \leq background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	False	Refer to Job Narrative for details.
Containers requiring zero headspace have no headspace or bubble is <math><6\text{mm}</math> (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

