



Environment Testing  
TestAmerica

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14

## ANALYTICAL REPORT

Eurofins TestAmerica, Canton  
4101 Shuffel Street NW  
North Canton, OH 44720  
Tel: (330)497-9396

Laboratory Job ID: 240-119308-1

Client Project/Site: Ford LTP Livonia MI - E203631

For:

ARCADIS U.S., Inc.  
28550 Cabot Drive  
Suite 500  
Novi, Michigan 48377

Attn: Kristoffer Hinskey

*Mike DelMonico*

Authorized for release by:

10/7/2019 3:55:16 PM

Michael DelMonico, Project Manager I

(330)497-9396

[michael.delmonico@testamericainc.com](mailto:michael.delmonico@testamericainc.com)

### LINKS

Review your project  
results through

**TotalAccess**

Have a Question?

Ask  
The  
Expert

Visit us at:

[www.testamericainc.com](http://www.testamericainc.com)

*This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.*

*Results relate only to the items tested and the sample(s) as received by the laboratory.*

# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Method Summary . . . . .	6
Sample Summary . . . . .	7
Detection Summary . . . . .	8
Client Sample Results . . . . .	9
Surrogate Summary . . . . .	13
QC Sample Results . . . . .	14
QC Association Summary . . . . .	17
Lab Chronicle . . . . .	18
Certification Summary . . . . .	19
Chain of Custody . . . . .	20

# Definitions/Glossary

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## Qualifiers

### GC/MS VOA

Qualifier	Qualifier Description
U	Indicates the analyte was analyzed for but not detected.
X	Surrogate is outside control limits

## Glossary

### Abbreviation

**These commonly used abbreviations may or may not be present in this report.**

<input checked="" type="checkbox"/>	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

1

2

3

4

5

6

7

8

9

10

11

12

13

14

# Case Narrative

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Job ID: 240-119308-1**

**Laboratory: Eurofins TestAmerica, Canton**

Narrative

## CASE NARRATIVE

**Client: ARCADIS U.S., Inc.**

**Project: Ford LTP Livonia MI - E203631**

**Report Number: 240-119308-1**

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

Eurofins TestAmerica, Canton attests to the validity of the laboratory data generated by Eurofins TestAmerica facilities reported herein. All analyses performed by Eurofins TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the application methods. Eurofins TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

The test results in this report meet all NELAP requirements for parameters for which accreditation is required or available. Any exceptions to NELAP requirements are noted in this report. Pursuant to NELAP, this report may not be reproduced, except in full, without the written approval of the laboratory.

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of Eurofins TestAmerica and its client.

### RECEIPT

The samples were received on 9/21/2019 9:50 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 2.3° C.

### VOLATILE ORGANIC COMPOUNDS (GCMS)

Samples MW-86\_091919 (240-119308-1), MW-86S\_091919 (240-119308-2), MW-97S\_091919 (240-119308-3) and TRIP BLANK (240-119308-4) were analyzed for volatile organic compounds (GCMS) in accordance with EPA SW-846 Method 8260B. The samples were analyzed on 10/01/2019.

1,2-Dichloroethane-d4 (Surr) failed the surrogate recovery criteria high for MW-86S\_091919 (240-119308-2) and LCS 240-403240/4. Refer to the QC report for details.

Surrogate recovery for the LCS and the following samples were outside the upper control limit: MW-86\_091919 (240-119308-1), MW-86S\_091919 (240-119308-2), MW-97S\_091919 (240-119308-3), TRIP BLANK (240-119308-4) and (LCS 240-403240/4). This sample did not contain any target analytes; therefore, re-extraction and/or re-analysis was not performed.

Surrogate recovery for the following sample was outside the upper control limit: MW-86S\_091919 (240-119308-2). This sample did not contain any target analytes; therefore, re-extraction and/or re-analysis was not performed.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

## Case Narrative

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

### Job ID: 240-119308-1 (Continued)

#### Laboratory: Eurofins TestAmerica, Canton (Continued)

##### **VOLATILE ORGANIC COMPOUNDS (GCMS SIM)**

Samples MW-86\_091919 (240-119308-1), MW-86S\_091919 (240-119308-2) and MW-97S\_091919 (240-119308-3) were analyzed for volatile organic compounds (GCMS SIM) in accordance with EPA SW-846 Method 8260B SIM. The samples were analyzed on 09/26/2019.

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

## Method Summary

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
8260B SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
5030B	Purge and Trap	SW846	TAL CAN

### Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

### Laboratory References:

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396

## Sample Summary

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Asset ID
240-119308-1	MW-86_091919	Water	09/19/19 11:30	09/21/19 09:50	
240-119308-2	MW-86S_091919	Water	09/19/19 13:55	09/21/19 09:50	
240-119308-3	MW-97S_091919	Water	09/19/19 15:54	09/21/19 09:50	
240-119308-4	TRIP BLANK	Water	09/19/19 00:00	09/21/19 09:50	

## Detection Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Client Sample ID: MW-86\_091919**

**Lab Sample ID: 240-119308-1**

No Detections.

**Client Sample ID: MW-86S\_091919**

**Lab Sample ID: 240-119308-2**

No Detections.

**Client Sample ID: MW-97S\_091919**

**Lab Sample ID: 240-119308-3**

No Detections.

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-119308-4**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins TestAmerica, Canton

# Client Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Client Sample ID: MW-86\_091919**

**Lab Sample ID: 240-119308-1**

Matrix: Water

Date Collected: 09/19/19 11:30

Date Received: 09/21/19 09:50

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			09/26/19 19:54	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	102		63 - 125					09/26/19 19:54	1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 04:09	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			10/01/19 04:09	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			10/01/19 04:09	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 04:09	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			10/01/19 04:09	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			10/01/19 04:09	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	118		70 - 121					10/01/19 04:09	1
4-Bromofluorobenzene (Surr)	99		59 - 120					10/01/19 04:09	1
Toluene-d8 (Surr)	100		70 - 123					10/01/19 04:09	1
Dibromofluoromethane (Surr)	88		75 - 128					10/01/19 04:09	1

# Client Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Client Sample ID: MW-86S\_091919**

**Lab Sample ID: 240-119308-2**

Matrix: Water

Date Collected: 09/19/19 13:55

Date Received: 09/21/19 09:50

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			09/26/19 20:19	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	101		63 - 125					09/26/19 20:19	1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 04:31	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			10/01/19 04:31	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			10/01/19 04:31	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 04:31	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			10/01/19 04:31	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			10/01/19 04:31	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	125	X	70 - 121					10/01/19 04:31	1
4-Bromofluorobenzene (Surr)	104		59 - 120					10/01/19 04:31	1
Toluene-d8 (Surr)	108		70 - 123					10/01/19 04:31	1
Dibromofluoromethane (Surr)	93		75 - 128					10/01/19 04:31	1

# Client Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Client Sample ID: MW-97S\_091919**

**Lab Sample ID: 240-119308-3**

**Matrix: Water**

Date Collected: 09/19/19 15:54

Date Received: 09/21/19 09:50

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			09/26/19 20:44	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	101		63 - 125					09/26/19 20:44	1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 04:53	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			10/01/19 04:53	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			10/01/19 04:53	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 04:53	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			10/01/19 04:53	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			10/01/19 04:53	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	118		70 - 121					10/01/19 04:53	1
4-Bromofluorobenzene (Surr)	95		59 - 120					10/01/19 04:53	1
Toluene-d8 (Surr)	100		70 - 123					10/01/19 04:53	1
Dibromofluoromethane (Surr)	90		75 - 128					10/01/19 04:53	1

# Client Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Client Sample ID: TRIP BLANK**

Date Collected: 09/19/19 00:00

Date Received: 09/21/19 09:50

**Lab Sample ID: 240-119308-4**

Matrix: Water

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 05:15	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			10/01/19 05:15	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			10/01/19 05:15	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			10/01/19 05:15	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			10/01/19 05:15	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			10/01/19 05:15	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)	114		70 - 121					10/01/19 05:15	1
4-Bromofluorobenzene (Surr)	97		59 - 120					10/01/19 05:15	1
Toluene-d8 (Surr)	98		70 - 123					10/01/19 05:15	1
Dibromofluoromethane (Surr)	86		75 - 128					10/01/19 05:15	1

# Surrogate Summary

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCA (70-121)	BFB (59-120)	TOL (70-123)	DBFM (75-128)
240-119304-B-1 MSD	Matrix Spike Duplicate	113	97	97	93
240-119304-D-1 MS	Matrix Spike	109	98	98	82
240-119308-1	MW-86_091919	118	99	100	88
240-119308-2	MW-86S_091919	125 X	104	108	93
240-119308-3	MW-97S_091919	118	95	100	90
240-119308-4	TRIP BLANK	114	97	98	86
LCS 240-403240/4	Lab Control Sample	123 X	107	104	96
MB 240-403240/6	Method Blank	117	98	101	91

### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

BFB = 4-Bromofluorobenzene (Surr)

TOL = Toluene-d8 (Surr)

DBFM = Dibromofluoromethane (Surr)

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCA (63-125)			
240-119306-A-4 MS	Matrix Spike	101			
240-119306-A-4 MSD	Matrix Spike Duplicate	101			
240-119308-1	MW-86_091919	102			
240-119308-2	MW-86S_091919	101			
240-119308-3	MW-97S_091919	101			
LCS 240-402639/4	Lab Control Sample	99			
MB 240-402639/5	Method Blank	98			

### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

# QC Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 240-403240/6**

**Matrix: Water**

**Analysis Batch: 403240**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			09/30/19 21:51	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			09/30/19 21:51	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			09/30/19 21:51	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			09/30/19 21:51	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			09/30/19 21:51	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			09/30/19 21:51	1

**MB MB**

Surrogate	MB	MB	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
1,2-Dichloroethane-d4 (Surr)	117		70 - 121				09/30/19 21:51	1
4-Bromofluorobenzene (Surr)	98		59 - 120				09/30/19 21:51	1
Toluene-d8 (Surr)	101		70 - 123				09/30/19 21:51	1
Dibromofluoromethane (Surr)	91		75 - 128				09/30/19 21:51	1

**Lab Sample ID: LCS 240-403240/4**

**Matrix: Water**

**Analysis Batch: 403240**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spikes	LCS	LCS	Result	Qualifier	Unit	D	%Rec	%Rec.
	Added	Result	Qualifier						
1,1-Dichloroethene	10.0	9.16		ug/L			92	65 - 139	
cis-1,2-Dichloroethene	10.0	10.3		ug/L			103	76 - 128	
Tetrachloroethene	10.0	9.44		ug/L			94	74 - 130	
trans-1,2-Dichloroethene	10.0	9.90		ug/L			99	78 - 133	
Trichloroethene	10.0	8.92		ug/L			89	76 - 125	
Vinyl chloride	10.0	8.80		ug/L			88	58 - 143	

**LCS LCS**

Surrogate	LCSS	LCSS	%Recovery	Qualifier	Limits
	Result	Qualifier			
1,2-Dichloroethane-d4 (Surr)	123	X	70 - 121		
4-Bromofluorobenzene (Surr)	107		59 - 120		
Toluene-d8 (Surr)	104		70 - 123		
Dibromofluoromethane (Surr)	96		75 - 128		

**Lab Sample ID: 240-119304-B-1 MSD**

**Matrix: Water**

**Analysis Batch: 403240**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec.
	Result	Qualifier	Added	Result	Qualifier				
1,1-Dichloroethene	1.0	U	10.0	7.79		ug/L	78	53 - 140	8
cis-1,2-Dichloroethene	1.0	U	10.0	9.05		ug/L	90	64 - 130	3
Tetrachloroethene	1.0	U	10.0	7.07		ug/L	71	51 - 136	2
trans-1,2-Dichloroethene	1.0	U	10.0	8.46		ug/L	85	68 - 133	4
Trichloroethene	1.0	U	10.0	7.18		ug/L	72	55 - 131	0
Vinyl chloride	1.0	U	10.0	7.11		ug/L	71	43 - 154	2

**MSD MSD**

Surrogate	MSD	MSD	%Recovery	Qualifier	Limits
	Result	Qualifier			
1,2-Dichloroethane-d4 (Surr)	113	X	70 - 121		
4-Bromofluorobenzene (Surr)	97		59 - 120		
Toluene-d8 (Surr)	97		70 - 123		

Eurofins TestAmerica, Canton

# QC Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID:** 240-119304-B-1 MSD

**Matrix:** Water

**Analysis Batch:** 403240

**Client Sample ID:** Matrix Spike Duplicate  
**Prep Type:** Total/NA

Surrogate	MSD	MSD	%Recovery	Qualifier	Limits
Dibromofluoromethane (Surr)			93		75 - 128

**Lab Sample ID:** 240-119304-D-1 MS

**Matrix:** Water

**Analysis Batch:** 403240

**Client Sample ID:** Matrix Spike  
**Prep Type:** Total/NA

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	Limits
	Result	Qualifier	Added	Result	Qualifier				
1,1-Dichloroethene	1.0	U	10.0	7.21		ug/L	72	53 - 140	
cis-1,2-Dichloroethene	1.0	U	10.0	8.79		ug/L	88	64 - 130	
Tetrachloroethene	1.0	U	10.0	7.24		ug/L	72	51 - 136	
trans-1,2-Dichloroethene	1.0	U	10.0	8.15		ug/L	81	68 - 133	
Trichloroethene	1.0	U	10.0	7.21		ug/L	72	55 - 131	
Vinyl chloride	1.0	U	10.0	7.00		ug/L	70	43 - 154	

Surrogate	MS	MS	%Recovery	Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	109				70 - 121
4-Bromofluorobenzene (Surr)	98				59 - 120
Toluene-d8 (Surr)	98				70 - 123
Dibromofluoromethane (Surr)	82				75 - 128

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

**Lab Sample ID:** MB 240-402639/5

**Matrix:** Water

**Analysis Batch:** 402639

**Client Sample ID:** Method Blank  
**Prep Type:** Total/NA

Analyte	MB	MB	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane			2.0	U		0.86	ug/L			09/26/19 13:18	1
<b>Surrogate</b>	<b>MB</b>	<b>MB</b>							<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,2-Dichloroethane-d4 (Surr)			98		63 - 125					09/26/19 13:18	1

**Lab Sample ID:** LCS 240-402639/4

**Matrix:** Water

**Analysis Batch:** 402639

**Client Sample ID:** Lab Control Sample  
**Prep Type:** Total/NA

Analyte	Spike	LCS	LCS	Result	Qualifier	Unit	D	%Rec	Limits
	Added								
1,4-Dioxane				10.0		10.8		108	59 - 131
<b>Surrogate</b>	<b>LCS</b>	<b>LCS</b>							
1,2-Dichloroethane-d4 (Surr)			99		63 - 125				

**Lab Sample ID:** 240-119306-A-4 MS

**Matrix:** Water

**Analysis Batch:** 402639

**Client Sample ID:** Matrix Spike  
**Prep Type:** Total/NA

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	Limits
	Result	Qualifier	Added	Result	Qualifier				
1,4-Dioxane	2.0	U	10.0	10.5		ug/L	105	52 - 129	

Eurofins TestAmerica, Canton

# QC Sample Results

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS) (Continued)

Surrogate	MS %Recovery	MS Qualifier	Limits
1,2-Dichloroethane-d4 (Surrogate)	101		63 - 125

Lab Sample ID: 240-119306-A-4 MSD

Matrix: Water

Analysis Batch: 402639

Client Sample ID: Matrix Spike Duplicate  
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec. Limits	RPD	RPD Limit
1,4-Dioxane	2.0	U	10.0	11.3		ug/L	113	52 - 129	8	13

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surrogate)	101		63 - 125

# QC Association Summary

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## GC/MS VOA

### Analysis Batch: 402639

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-119308-1	MW-86_091919	Total/NA	Water	8260B SIM	
240-119308-2	MW-86S_091919	Total/NA	Water	8260B SIM	
240-119308-3	MW-97S_091919	Total/NA	Water	8260B SIM	
MB 240-402639/5	Method Blank	Total/NA	Water	8260B SIM	
LCS 240-402639/4	Lab Control Sample	Total/NA	Water	8260B SIM	
240-119306-A-4 MS	Matrix Spike	Total/NA	Water	8260B SIM	
240-119306-A-4 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B SIM	

### Analysis Batch: 403240

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-119308-1	MW-86_091919	Total/NA	Water	8260B	
240-119308-2	MW-86S_091919	Total/NA	Water	8260B	
240-119308-3	MW-97S_091919	Total/NA	Water	8260B	
240-119308-4	TRIP BLANK	Total/NA	Water	8260B	
MB 240-403240/6	Method Blank	Total/NA	Water	8260B	
LCS 240-403240/4	Lab Control Sample	Total/NA	Water	8260B	
240-119304-B-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B	
240-119304-D-1 MS	Matrix Spike	Total/NA	Water	8260B	

# Lab Chronicle

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

**Client Sample ID: MW-86\_091919**  
Date Collected: 09/19/19 11:30  
Date Received: 09/21/19 09:50

**Lab Sample ID: 240-119308-1**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	403240	10/01/19 04:09	LEE	TAL CAN
Total/NA	Analysis	8260B SIM		1	402639	09/26/19 19:54	SAM	TAL CAN

**Client Sample ID: MW-86S\_091919**  
Date Collected: 09/19/19 13:55  
Date Received: 09/21/19 09:50

**Lab Sample ID: 240-119308-2**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	403240	10/01/19 04:31	LEE	TAL CAN
Total/NA	Analysis	8260B SIM		1	402639	09/26/19 20:19	SAM	TAL CAN

**Client Sample ID: MW-97S\_091919**  
Date Collected: 09/19/19 15:54  
Date Received: 09/21/19 09:50

**Lab Sample ID: 240-119308-3**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	403240	10/01/19 04:53	LEE	TAL CAN
Total/NA	Analysis	8260B SIM		1	402639	09/26/19 20:44	SAM	TAL CAN

**Client Sample ID: TRIP BLANK**  
Date Collected: 09/19/19 00:00  
Date Received: 09/21/19 09:50

**Lab Sample ID: 240-119308-4**  
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	403240	10/01/19 05:15	LEE	TAL CAN

## Laboratory References:

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396

Eurofins TestAmerica, Canton

# Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.

Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-119308-1

## Laboratory: Eurofins TestAmerica, Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	2927	02-23-20
Connecticut	State	PH-0590	12-31-19
Florida	NELAP	E87225	06-30-20
Georgia	State	4062	02-23-20
Illinois	NELAP	004498	07-31-20
Iowa	State	421	06-01-20
Kansas	NELAP	E-10336	04-30-20
Kentucky (UST)	State	112225	02-23-20
Kentucky (WW)	State	KY98016	12-31-19
Minnesota	NELAP	OH00048	12-31-19
Minnesota (Petrofund)	State Program	3506	07-31-21
New Jersey	NELAP	OH001	06-30-20
New York	NELAP	10975	03-31-20
Ohio VAP	State	CL0024	06-05-21
Oregon	NELAP	4062	02-23-20
Pennsylvania	NELAP	68-00340	08-31-20
Texas	NELAP	T104704517-18-10	08-31-20
USDA	US Federal Programs	P330-16-00404	12-28-19
Virginia	NELAP	010101	09-14-20
Washington	State	C971	01-12-20
West Virginia DEP	State	210	12-31-19



**Chain of Custody Record**

**TestAmerica Laboratory location:** Brighton --- 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

MICHIGAN  
190

**TestAmerica**  
THE LEADER IN ENVIRONMENTAL TESTING

Possible Hazard Identification	<input checked="" type="checkbox"/> Non-Hazard	<input type="checkbox"/> Flammable
<input type="checkbox"/> Special Instructions/QC Requirements & Comments:		

**Sample Disposal** (A fee may be assessed if samples are retained longer than 1 month)  
 Return to Client     Disposal By Lab     Archive For \_\_\_\_\_ Months

Submit all results through Cadena at [jim.tomalia@cadena.com](mailto:jim.tomalia@cadena.com). Cadena #E203631

Submit all results through Cadena at [jim.tomalia@](mailto:jim.tomalia@)

Date/Time: 9/19/19	Received by: <u>Mel O.</u>	Company: <u>Archadis</u>	Date/Time: 9/19/19
Date/Time: 9/19/19	Received by: <u>Tom</u>	Company: <u>Archadis</u>	Date/Time: 9/19/19
Date/Time: 9/20/19	Received in Laboratory by: <u>Tom</u>	Company: <u>ESFA</u>	Date/Time: 9-20-19

©2003, Intertech Data Works, Inc. All rights reserved.  
Intertek & Design™ are trademarks of Intertech Data Works, Inc.

## Eurofins TestAmerica Canton Sample Receipt Form/Narrative

## Canton Facility

Login # : 119308

Client <u>Arcadis</u>	Site Name _____	Cooler unpacked by: <u>A</u>
Cooler Received on <u>9-21-19</u>	Opened on <u>9-23-19</u>	
FedEx: 1 <sup>st</sup> Grd <u>Exp</u> UPS FAS Clipper	Client Drop Off	TestAmerica Courier

<b>Receipt After-hours: Drop-off Date/Time</b>		<b>Storage Location</b>
--	--	-------------------------

TestAmerica Cooler # <u>TA</u>	Foam Box	Client Cooler	Box	Other _____
Packing material used:	Bubble Wrap	Foam	Plastic Bag	None Other _____
COOLANT:	Wet Ice	Blue Ice	Dry Ice	Water None

1. Cooler temperature upon receipt  
 IR GUN# IR-10 (CF +0.7 °C) Observed Cooler Temp. 1.6 °C Corrected Cooler Temp. -3 °C  
 IR GUN #IR-11 (CF +0.9°C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C

2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 1 Yes No  
 -Were the seals on the outside of the cooler(s) signed & dated? Yes No NA  
 -Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No  
 -Were tamper/custody seals intact and uncompromised? Yes No NA
3. Shippers' packing slip attached to the cooler(s)? Yes No  
 4. Did custody papers accompany the sample(s)? Yes No  
 5. Were the custody papers relinquished & signed in the appropriate place? Yes No  
 6. Was/were the person(s) who collected the samples clearly identified on the COC? Yes No  
 7. Did all bottles arrive in good condition (Unbroken)? Yes No  
 8. Could all bottle labels be reconciled with the COC? Yes No  
 9. Were correct bottle(s) used for the test(s) indicated? Yes No  
 10. Sufficient quantity received to perform indicated analyses? Yes No  
 11. Are these work share samples? Yes No

If yes, Questions 12-16 have been checked at the originating laboratory.

12. Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC991818  
 13. Were VOAs on the COC? Yes No  
 14. Were air bubbles >6 mm in any VOA vials? Larger than this. Yes No NA  
 15. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # \_\_\_\_\_ Yes No  
 16. Was a LL Hg or Me Hg trip blank present? Yes No

Tests that are not checked for pH by Receiving:  
 VOAs  
 Oil and Grease  
 TOC

Contacted PM \_\_\_\_\_ Date \_\_\_\_\_ by \_\_\_\_\_ via Verbal Voice Mail Other \_\_\_\_\_

Concerning \_\_\_\_\_

## 17. CHAIN OF CUSTODY &amp; SAMPLE DISCREPANCIES

Samples processed by: BC

## 18. SAMPLE CONDITION

Sample(s) \_\_\_\_\_ were received after the recommended holding time had expired.

Sample(s) \_\_\_\_\_ were received in a broken container.

Sample(s) \_\_\_\_\_ were received with bubble >6 mm in diameter. (Notify PM)

## 19. SAMPLE PRESERVATION

Sample(s) \_\_\_\_\_ were further preserved in the laboratory.  
 Time preserved: \_\_\_\_\_ Preservative(s) added/Lot number(s): \_\_\_\_\_

VOA Sample Preservation - Date/Time VOAs Frozen: \_\_\_\_\_