

## ANALYTICAL REPORT

Eurofins TestAmerica, Canton  
4101 Shuffel Street NW  
North Canton, OH 44720  
Tel: (330)497-9396

Laboratory Job ID: 240-113320-1

Client Project/Site: Ford LTP Livonia MI - E203631

**For:**

ARCADIS U.S., Inc.  
28550 Cabot Drive  
Suite 500  
Novi, Michigan 48377

Attn: Kristoffer Hinskey



Authorized for release by:  
6/11/2019 4:33:25 PM

Michael DelMonico, Project Manager I  
(330)497-9396  
[michael.delmonico@testamericainc.com](mailto:michael.delmonico@testamericainc.com)

### LINKS

Review your project  
results through  
**TotalAccess**

Have a Question?



Visit us at:  
[www.testamericainc.com](http://www.testamericainc.com)

*This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.*

*Results relate only to the items tested and the sample(s) as received by the laboratory.*



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	2
Definitions/Glossary . . . . .	3
Case Narrative . . . . .	4
Method Summary . . . . .	5
Sample Summary . . . . .	6
Detection Summary . . . . .	7
Client Sample Results . . . . .	8
Surrogate Summary . . . . .	10
QC Sample Results . . . . .	11
QC Association Summary . . . . .	15
Lab Chronicle . . . . .	16
Certification Summary . . . . .	17
Chain of Custody . . . . .	18

# Definitions/Glossary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Qualifiers

### GC/MS VOA

Qualifier	Qualifier Description
F2	MS/MSD RPD exceeds control limits
U	Indicates the analyte was analyzed for but not detected.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
▫	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

# Case Narrative

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

**Job ID: 240-113320-1**

**Laboratory: Eurofins TestAmerica, Canton**

**Narrative**

## CASE NARRATIVE

**Client: ARCADIS U.S., Inc.**

**Project: Ford LTP Livonia MI - E203631**

**Report Number: 240-113320-1**

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

Eurofins TestAmerica, Canton attests to the validity of the laboratory data generated by Eurofins TestAmerica facilities reported herein. All analyses performed by Eurofins TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the application methods. Eurofins TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

The test results in this report meet all NELAP requirements for parameters for which accreditation is required or available. Any exceptions to NELAP requirements are noted in this report. Pursuant to NELAP, this report may not be reproduced, except in full, without the written approval of the laboratory.

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of Eurofins TestAmerica and its client.

### **RECEIPT**

The samples were received on 5/25/2019 10:00 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperatures of the 2 coolers at receipt time were 3.0° C and 4.0° C.

### **VOLATILE ORGANIC COMPOUNDS (GCMS)**

Samples MW-142S\_052319 (240-113320-1) and TRIP BLANK (240-113320-2) were analyzed for volatile organic compounds (GCMS) in accordance with EPA SW-846 Method 8260B. The samples were analyzed on 06/04/2019 and 06/06/2019.

The continuing calibration verification (CCV) associated with batch 384267 recovered above the upper control limit for Vinyl Chloride. The samples associated with this CCV were non-detects for the affected analyte; therefore, the data have been reported. The following samples are impacted: MW-142S\_052319 (240-113320-1).

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

### **VOLATILE ORGANIC COMPOUNDS (GCMS SIM)**

Sample MW-142S\_052319 (240-113320-1) was analyzed for volatile organic compounds (GCMS SIM) in accordance with EPA SW-846 Method 8260B SIM. The samples were analyzed on 05/31/2019.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

# Method Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
8260B SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
5030B	Purge and Trap	SW846	TAL CAN

**Protocol References:**

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

**Laboratory References:**

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396



# Sample Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Asset ID
240-113320-1	MW-142S_052319	Water	05/23/19 09:38	05/25/19 10:00	
240-113320-2	TRIP BLANK	Water	05/23/19 00:00	05/25/19 10:00	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

# Detection Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

**Client Sample ID: MW-142S\_052319**

**Lab Sample ID: 240-113320-1**

No Detections.

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-113320-2**

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

This Detection Summary does not include radiochemical test results.

Eurofins TestAmerica, Canton

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

**Client Sample ID: MW-142S\_052319**

**Lab Sample ID: 240-113320-1**

Date Collected: 05/23/19 09:38

Matrix: Water

Date Received: 05/25/19 10:00

**Method: 8260B SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L	-		05/31/19 17:02	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	109		63 - 125		05/31/19 17:02	1

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L	-		06/04/19 03:44	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L	-		06/04/19 03:44	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L	-		06/04/19 03:44	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L	-		06/04/19 03:44	1
Trichloroethene	1.0	U	1.0	0.10	ug/L	-		06/04/19 03:44	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L	-		06/04/19 03:44	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	97		70 - 121		06/04/19 03:44	1
4-Bromofluorobenzene (Surr)	89		59 - 120		06/04/19 03:44	1
Toluene-d8 (Surr)	101		70 - 123		06/04/19 03:44	1
Dibromofluoromethane (Surr)	103		75 - 128		06/04/19 03:44	1



# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-113320-2**

**Date Collected: 05/23/19 00:00**

**Matrix: Water**

**Date Received: 05/25/19 10:00**

**Method: 8260B - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/06/19 14:25	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/06/19 14:25	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/06/19 14:25	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/06/19 14:25	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/06/19 14:25	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/06/19 14:25	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	118		70 - 121		06/06/19 14:25	1
4-Bromofluorobenzene (Surr)	99		59 - 120		06/06/19 14:25	1
Toluene-d8 (Surr)	102		70 - 123		06/06/19 14:25	1
Dibromofluoromethane (Surr)	104		75 - 128		06/06/19 14:25	1

# Surrogate Summary

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCA (70-121)	BFB (59-120)	TOL (70-123)	DBFM (75-128)
240-113320-1	MW-142S_052319	97	89	101	103
240-113320-2	TRIP BLANK	118	99	102	104
240-113326-E-1 MSD	Matrix Spike Duplicate	88	94	95	89
240-113326-F-1 MS	Matrix Spike	84	96	96	89
240-113327-B-24 MS	Matrix Spike	110	104	105	101
240-113327-B-24 MSD	Matrix Spike Duplicate	110	104	104	99
LCS 240-384267/4	Lab Control Sample	91	107	106	101
LCS 240-384832/5	Lab Control Sample	108	103	102	100
MB 240-384267/6	Method Blank	95	91	100	105
MB 240-384832/8	Method Blank	113	101	99	102

#### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)  
 BFB = 4-Bromofluorobenzene (Surr)  
 TOL = Toluene-d8 (Surr)  
 DBFM = Dibromofluoromethane (Surr)

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA
		(63-125)
240-113320-1	MW-142S_052319	109
240-113406-C-1 MS	Matrix Spike	110
240-113406-C-1 MSD	Matrix Spike Duplicate	110
LCS 240-383941/4	Lab Control Sample	105
MB 240-383941/5	Method Blank	109

#### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

# QC Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Method: 8260B - Volatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 240-384267/6**  
**Matrix: Water**  
**Analysis Batch: 384267**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/03/19 22:11	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/03/19 22:11	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/03/19 22:11	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/03/19 22:11	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/03/19 22:11	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/03/19 22:11	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	95		70 - 121		06/03/19 22:11	1
4-Bromofluorobenzene (Surr)	91		59 - 120		06/03/19 22:11	1
Toluene-d8 (Surr)	100		70 - 123		06/03/19 22:11	1
Dibromofluoromethane (Surr)	105		75 - 128		06/03/19 22:11	1

**Lab Sample ID: LCS 240-384267/4**  
**Matrix: Water**  
**Analysis Batch: 384267**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	10.0	11.7		ug/L		117	65 - 139
cis-1,2-Dichloroethene	10.0	11.3		ug/L		113	76 - 128
Tetrachloroethene	10.0	9.27		ug/L		93	74 - 130
trans-1,2-Dichloroethene	10.0	11.2		ug/L		112	78 - 133
Trichloroethene	10.0	9.49		ug/L		95	76 - 125
Vinyl chloride	10.0	12.7		ug/L		127	58 - 143

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	91		70 - 121
4-Bromofluorobenzene (Surr)	107		59 - 120
Toluene-d8 (Surr)	106		70 - 123
Dibromofluoromethane (Surr)	101		75 - 128

**Lab Sample ID: 240-113326-E-1 MSD**  
**Matrix: Water**  
**Analysis Batch: 384267**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,1-Dichloroethene	1.0	U	10.0	9.78		ug/L		98	53 - 140	11	35
cis-1,2-Dichloroethene	1.0	U	10.0	9.81		ug/L		98	64 - 130	11	21
Tetrachloroethene	1.0	U	10.0	7.38		ug/L		74	51 - 136	1	23
trans-1,2-Dichloroethene	1.0	U	10.0	9.41		ug/L		94	68 - 133	10	24
Trichloroethene	1.0	U	10.0	7.74		ug/L		77	55 - 131	4	23
Vinyl chloride	1.0	U	10.0	10.7		ug/L		107	43 - 154	1	29

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	88		70 - 121
4-Bromofluorobenzene (Surr)	94		59 - 120
Toluene-d8 (Surr)	95		70 - 123

Eurofins TestAmerica, Canton

# QC Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 240-113326-E-1 MSD**  
**Matrix: Water**  
**Analysis Batch: 384267**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

<i>Surrogate</i>	<i>%Recovery</i>	<i>MSD Qualifier</i>	<i>MSD Limits</i>
<i>Dibromofluoromethane (Surr)</i>	89		75 - 128

**Lab Sample ID: 240-113326-F-1 MS**  
**Matrix: Water**  
**Analysis Batch: 384267**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

<b>Analyte</b>	<b>Sample Result</b>	<b>Sample Qualifier</b>	<b>Spike Added</b>	<b>MS Result</b>	<b>MS Qualifier</b>	<b>Unit</b>	<b>D</b>	<b>%Rec</b>	<b>%Rec. Limits</b>
1,1-Dichloroethene	1.0	U	10.0	8.75		ug/L		88	53 - 140
cis-1,2-Dichloroethene	1.0	U	10.0	8.82		ug/L		88	64 - 130
Tetrachloroethene	1.0	U	10.0	7.31		ug/L		73	51 - 136
trans-1,2-Dichloroethene	1.0	U	10.0	8.53		ug/L		85	68 - 133
Trichloroethene	1.0	U	10.0	7.40		ug/L		74	55 - 131
Vinyl chloride	1.0	U	10.0	10.5		ug/L		105	43 - 154

<i>Surrogate</i>	<i>%Recovery</i>	<i>MS Qualifier</i>	<i>MS Limits</i>
<i>1,2-Dichloroethane-d4 (Surr)</i>	84		70 - 121
<i>4-Bromofluorobenzene (Surr)</i>	96		59 - 120
<i>Toluene-d8 (Surr)</i>	96		70 - 123
<i>Dibromofluoromethane (Surr)</i>	89		75 - 128

**Lab Sample ID: MB 240-384832/8**  
**Matrix: Water**  
**Analysis Batch: 384832**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

<b>Analyte</b>	<b>MB Result</b>	<b>MB Qualifier</b>	<b>RL</b>	<b>MDL</b>	<b>Unit</b>	<b>D</b>	<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/06/19 12:21	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/06/19 12:21	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/06/19 12:21	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/06/19 12:21	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/06/19 12:21	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/06/19 12:21	1

<i>Surrogate</i>	<i>%Recovery</i>	<i>MB Qualifier</i>	<i>MB Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>1,2-Dichloroethane-d4 (Surr)</i>	113		70 - 121		06/06/19 12:21	1
<i>4-Bromofluorobenzene (Surr)</i>	101		59 - 120		06/06/19 12:21	1
<i>Toluene-d8 (Surr)</i>	99		70 - 123		06/06/19 12:21	1
<i>Dibromofluoromethane (Surr)</i>	102		75 - 128		06/06/19 12:21	1

**Lab Sample ID: LCS 240-384832/5**  
**Matrix: Water**  
**Analysis Batch: 384832**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

<b>Analyte</b>	<b>Spike Added</b>	<b>LCS Result</b>	<b>LCS Qualifier</b>	<b>Unit</b>	<b>D</b>	<b>%Rec</b>	<b>%Rec. Limits</b>
1,1-Dichloroethene	20.0	18.1		ug/L		91	65 - 139
cis-1,2-Dichloroethene	20.0	17.5		ug/L		88	76 - 128
Tetrachloroethene	20.0	19.5		ug/L		97	74 - 130
trans-1,2-Dichloroethene	20.0	18.4		ug/L		92	78 - 133
Trichloroethene	20.0	18.4		ug/L		92	76 - 125

Eurofins TestAmerica, Canton

# QC Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 240-384832/5**

**Matrix: Water**

**Analysis Batch: 384832**

**Client Sample ID: Lab Control Sample**

**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
Vinyl chloride	20.0	17.2		ug/L		86	58 - 143

  

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	108		70 - 121
4-Bromofluorobenzene (Surr)	103		59 - 120
Toluene-d8 (Surr)	102		70 - 123
Dibromofluoromethane (Surr)	100		75 - 128

**Lab Sample ID: 240-113327-B-24 MS**

**Matrix: Water**

**Analysis Batch: 384832**

**Client Sample ID: Matrix Spike**

**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	130	U	2500	2150		ug/L		86	53 - 140
cis-1,2-Dichloroethene	4100		2500	5930		ug/L		75	64 - 130
Tetrachloroethene	130	U	2500	2110		ug/L		85	51 - 136
trans-1,2-Dichloroethene	28	J	2500	2250		ug/L		89	68 - 133
Trichloroethene	950		2500	2940		ug/L		80	55 - 131
Vinyl chloride	250		2500	2980		ug/L		109	43 - 154

  

Surrogate	MS %Recovery	MS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 121
4-Bromofluorobenzene (Surr)	104		59 - 120
Toluene-d8 (Surr)	105		70 - 123
Dibromofluoromethane (Surr)	101		75 - 128

**Lab Sample ID: 240-113327-B-24 MSD**

**Matrix: Water**

**Analysis Batch: 384832**

**Client Sample ID: Matrix Spike Duplicate**

**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,1-Dichloroethene	130	U	2500	2210		ug/L		88	53 - 140	3	35
cis-1,2-Dichloroethene	4100		2500	5980		ug/L		77	64 - 130	1	21
Tetrachloroethene	130	U	2500	2130		ug/L		85	51 - 136	1	23
trans-1,2-Dichloroethene	28	J	2500	2330		ug/L		92	68 - 133	3	24
Trichloroethene	950		2500	3040		ug/L		84	55 - 131	3	23
Vinyl chloride	250		2500	2860		ug/L		104	43 - 154	4	29

  

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	110		70 - 121
4-Bromofluorobenzene (Surr)	104		59 - 120
Toluene-d8 (Surr)	104		70 - 123
Dibromofluoromethane (Surr)	99		75 - 128

# QC Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 240-383941/5**  
**Matrix: Water**  
**Analysis Batch: 383941**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/31/19 13:44	1
Surrogate	MB %Recovery	MB Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	109		63 - 125					05/31/19 13:44	1

**Lab Sample ID: LCS 240-383941/4**  
**Matrix: Water**  
**Analysis Batch: 383941**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	10.0	12.3		ug/L		123	59 - 131
Surrogate	LCS %Recovery	LCS Qualifier	Limits				
1,2-Dichloroethane-d4 (Surr)	105		63 - 125				

**Lab Sample ID: 240-113406-C-1 MS**  
**Matrix: Water**  
**Analysis Batch: 383941**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	2.0	U F2	10.0	8.52		ug/L		85	52 - 129
Surrogate	MS %Recovery	MS Qualifier	Limits						
1,2-Dichloroethane-d4 (Surr)	110		63 - 125						

**Lab Sample ID: 240-113406-C-1 MSD**  
**Matrix: Water**  
**Analysis Batch: 383941**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,4-Dioxane	2.0	U F2	10.0	12.2	F2	ug/L		122	52 - 129	36	13
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
1,2-Dichloroethane-d4 (Surr)	110		63 - 125								

# QC Association Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## GC/MS VOA

### Analysis Batch: 383941

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-113320-1	MW-142S_052319	Total/NA	Water	8260B SIM	
MB 240-383941/5	Method Blank	Total/NA	Water	8260B SIM	
LCS 240-383941/4	Lab Control Sample	Total/NA	Water	8260B SIM	
240-113406-C-1 MS	Matrix Spike	Total/NA	Water	8260B SIM	
240-113406-C-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B SIM	

### Analysis Batch: 384267

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-113320-1	MW-142S_052319	Total/NA	Water	8260B	
MB 240-384267/6	Method Blank	Total/NA	Water	8260B	
LCS 240-384267/4	Lab Control Sample	Total/NA	Water	8260B	
240-113326-E-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B	
240-113326-F-1 MS	Matrix Spike	Total/NA	Water	8260B	

### Analysis Batch: 384832

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-113320-2	TRIP BLANK	Total/NA	Water	8260B	
MB 240-384832/8	Method Blank	Total/NA	Water	8260B	
LCS 240-384832/5	Lab Control Sample	Total/NA	Water	8260B	
240-113327-B-24 MS	Matrix Spike	Total/NA	Water	8260B	
240-113327-B-24 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B	

# Lab Chronicle

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

**Client Sample ID: MW-142S\_052319**

**Lab Sample ID: 240-113320-1**

**Date Collected: 05/23/19 09:38**

**Matrix: Water**

**Date Received: 05/25/19 10:00**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	384267	06/04/19 03:44	LEE	TAL CAN
Total/NA	Analysis	8260B SIM		1	383941	05/31/19 17:02	SAM	TAL CAN

**Client Sample ID: TRIP BLANK**

**Lab Sample ID: 240-113320-2**

**Date Collected: 05/23/19 00:00**

**Matrix: Water**

**Date Received: 05/25/19 10:00**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	384832	06/06/19 14:25	HMB	TAL CAN

**Laboratory References:**

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396



# Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP Livonia MI - E203631

Job ID: 240-113320-1

## Laboratory: Eurofins TestAmerica, Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	EPA Region	Identification Number	Expiration Date
California	State Program	9	2927	02-23-20
Connecticut	State Program	1	PH-0590	12-31-19
Florida	NELAP	4	E87225	06-30-19 *
Illinois	NELAP	5	200004	07-31-19 *
Iowa	State Program	7	421	06-01-21
Kansas	NELAP	7	E-10336	04-30-20
Kentucky (UST)	State Program	4	58	02-23-20
Kentucky (WW)	State Program	4	98016	12-31-19
Minnesota	NELAP	5	039-999-348	12-31-19 *
Minnesota (Petrofund)	State Program	1	3506	07-31-19 *
Nevada	State Program	9	OH00048	07-31-19
New Jersey	NELAP	2	OH001	06-30-19 *
New York	NELAP	2	10975	03-31-20
Ohio VAP	State Program	5	CL0024	06-05-21
Oregon	NELAP	10	4062	02-23-20
Pennsylvania	NELAP	3	68-00340	08-31-19 *
Texas	NELAP	6	T104704517-18-10	08-31-19 *
USDA	Federal		P330-16-00404	12-28-19
Virginia	NELAP	3	460175	09-14-19 *
Washington	State Program	10	C971	01-12-20 *
West Virginia DEP	State Program	3	210	12-31-19

\* Accreditation/Certification renewal pending - accreditation/certification considered valid.

**euofins TestAmerica, Canton**  
4101 Shuffel Street NW  
North Canton, OH 44720  
Phone (330) 497-9396 Fax (330) 497-0772

**MICHIGAN Chain of Custody Record**  
190

4.2 | C4,0  
3.2 | C3,0

<b>Client Information</b> Client Contact: Caitlin O'Neill Company: ARCADIS U.S. Inc Address: 28550 Cabot Drive Suite 500 City: Novi State, Zip: MI, 48377 Phone: _____ Email: Caitlin.O'Neill@arcadis.com Project Name: Ford LTP Livonia MI - E203631 Site: _____		Lab P/N: DelMonico, Michael E-Mail: michael.delmonico@testamericainc.com Phone: 248-662-7233 Sampler: S. Johnson Due Date Requested: _____ TAT Requested (days): 10 PO #: M1004948-000200002 WO #: M1004948-000200002 Cadena #: E203631 Project #: 24015353 SSOW#: _____		Carrier Tracking No(s): _____ COC No: 240-60548-25803.8 Page: _____ Page 8 of 13 Job #: 1051	
<b>Analysis Requested</b> Perform MS/MSD (Yes or No) <input checked="" type="checkbox"/> A Field Filtered Sample (Yes or No) <input checked="" type="checkbox"/> A 8260B - VOCs (Short List) <input checked="" type="checkbox"/> A 8260B, 8260B SIM <input checked="" type="checkbox"/> A 8260B - VOCs (Short List) <input checked="" type="checkbox"/> A		Preservation Codes: M - Hexane N - None O - AgNaO2 P - Na2O4S Q - Na2SO3 R - NaOH S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4.5 X - EDTA Y - EDA Z - other (specify) _____ Other: _____		Total Number of Containers: 6 Special Instructions/Note: _____	
<b>Sample Identification</b> MW-192S TRIP BLANK		Sample Date: 5-23-19 Sample Time: 9:58 Sample Type (C=comp, G=grab): 6 Preservation Code: Water	Matrix (Water, Solid, Over-sat, BT=Teabag, AF=AF)	Barcode: 240-113320 Chain of Custody	
Possible Hazard Identification <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological Deliverable Requested: I, II, III, IV Other (specify) _____					
Empty Kit Relinquished by: _____ Date: _____ Relinquished by: _____ Date/Time: 5-23-19 / 18:36 Company: ARCADIS Relinquished by: _____ Date/Time: 5-24-19 / 19:00 Company: ARCADIS Relinquished by: _____ Date/Time: 5/24/19 1315 Company: ETA Custody Seal No.: _____ Cooler Temperature(s) °C and Other Remarks: _____					
Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months Special Instructions/QC Requirements: _____					

TestAmerica Canton Sample Receipt Form/Narrative

Login # : 113320

Canton Facility

Client Arcadis Site Name \_\_\_\_\_  
 Cooler Received on 5-25-19 Opened on 5-25-19 1000  
 FedEx: 1<sup>st</sup> Grd  Exp UPS FAS Clipper Client Drop Off TestAmerica Courier Other \_\_\_\_\_

Cooler unpacked by:

Ryan Cribley

Receipt After-hours: Drop-off Date/Time \_\_\_\_\_ Storage Location \_\_\_\_\_

TestAmerica Cooler # 714 Foam Box Client Cooler Box Other \_\_\_\_\_  
 Packing material used:  Bubble Wrap  Foam  Plastic Bag None Other \_\_\_\_\_  
 COOLANT:  Wet Ice Blue Ice Dry Ice Water None

1. Cooler temperature upon receipt  See Multiple Cooler Form  
 IR GUN# IR-8 (CF -0.2 °C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C  
 IR GUN #36 (CF +0.7°C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C
2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 2 Yes No  
 -Were the seals on the outside of the cooler(s) signed & dated? total Yes No NA  
 -Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes  No  
 -Were tamper/custody seals intact and uncompromised?  Yes No NA
3. Shippers' packing slip attached to the cooler(s)?  Yes No
4. Did custody papers accompany the sample(s)?  Yes No
5. Were the custody papers relinquished & signed in the appropriate place?  Yes No
6. Was/were the person(s) who collected the samples clearly identified on the COC?  Yes No
7. Did all bottles arrive in good condition (Unbroken)?  Yes No
8. Could all bottle labels be reconciled with the COC?  Yes No
9. Were correct bottle(s) used for the test(s) indicated?  Yes No
10. Sufficient quantity received to perform indicated analyses?  Yes No
11. Are these work share samples? Yes  No
12. Were all preserved sample(s) at the correct pH upon receipt? Yes No  NA pH Strip Lot# HC984738
13. Were VOAs on the COC?  Yes No
14. Were air bubbles >6 mm in any VOA vials?  Yes  No  NA Larger than this.
15. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # \_\_\_\_\_  Yes No
16. Was a LL Hg or Me Hg trip blank present? \_\_\_\_\_  Yes  No

Tests that are not checked for pH by Receiving:  
 VOAs  
 Oil and Grease  
 TOC

Contacted PM \_\_\_\_\_ Date \_\_\_\_\_ by \_\_\_\_\_ via Verbal Voice Mail Other \_\_\_\_\_

Concerning \_\_\_\_\_

17. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES

Samples processed by:

Ryan

18. SAMPLE CONDITION

Sample(s) \_\_\_\_\_ were received after the recommended holding time had expired.  
 Sample(s) \_\_\_\_\_ were received in a broken container.  
 Sample(s) \_\_\_\_\_ were received with bubble >6 mm in diameter. (Notify PM)

19. SAMPLE PRESERVATION

Sample(s) \_\_\_\_\_ were further preserved in the laboratory.  
 Time preserved: \_\_\_\_\_ Preservative(s) added/Lot number(s): \_\_\_\_\_

VOA Sample Preservation - Date/Time VOAs Frozen: \_\_\_\_\_

