

ANALYTICAL REPORT

Eurofins TestAmerica, Canton
4101 Shuffel Street NW
North Canton, OH 44720
Tel: (330)497-9396

Laboratory Job ID: 240-130976-1
Client Project/Site: Ford LTP On-Site

For:
ARCADIS U.S., Inc.
28550 Cabot Drive
Suite 500
Novi, Michigan 48377

Attn: Kristoffer Hinskey



Authorized for release by:
6/12/2020 9:28:26 AM

Michael DelMonico, Project Manager I
(330)497-9396
michael.delmonico@testamericainc.com

LINKS

Review your project
results through
TotalAccess

Have a Question?



Visit us at:

www.eurofinsus.com/Env

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Method Summary	6
Sample Summary	7
Detection Summary	8
Client Sample Results	9
Surrogate Summary	15
QC Sample Results	16
QC Association Summary	19
Lab Chronicle	20
Certification Summary	21
Chain of Custody	22

Definitions/Glossary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
U	Indicates the analyte was analyzed for but not detected.
X	Surrogate recovery exceeds control limits

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

Case Narrative

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Job ID: 240-130976-1

Laboratory: Eurofins TestAmerica, Canton

Narrative

CASE NARRATIVE

Client: ARCADIS U.S., Inc.

Project: Ford LTP On-Site

Report Number: 240-130976-1

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

Eurofins TestAmerica, Canton attests to the validity of the laboratory data generated by Eurofins TestAmerica facilities reported herein. All analyses performed by Eurofins TestAmerica facilities were done using established laboratory SOPs that incorporate QA/QC procedures described in the application methods. Eurofins TestAmerica's operations groups have reviewed the data for compliance with the laboratory QA/QC plan, and data have been found to be compliant with laboratory protocols unless otherwise noted below.

The test results in this report meet all NELAP requirements for parameters for which accreditation is required or available. Any exceptions to NELAP requirements are noted in this report. Pursuant to NELAP, this report may not be reproduced, except in full, without the written approval of the laboratory.

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

All solid sample results are reported on an "as received" basis unless otherwise indicated by the presence of a % solids value in the method header.

This laboratory report is confidential and is intended for the sole use of Eurofins TestAmerica and its client.

RECEIPT

The samples were received on 5/29/2020 10:30 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperature of the cooler at receipt was 3.8° C.

VOLATILE ORGANIC COMPOUNDS (GCMS)

Samples TRIP BLANK (240-130976-1), MW-3_052720 (240-130976-2), MW-5_052720 (240-130976-3), MW-49_052720 (240-130976-4), MW-44_052720 (240-130976-5) and DUP-03 (240-130976-6) were analyzed for volatile organic compounds (GCMS) in accordance with EPA SW-846 Method 8260B. The samples were analyzed on 06/05/2020.

Samples MW-49_052720 (240-130976-4)[1000X], MW-44_052720 (240-130976-5)[10X] and DUP-03 (240-130976-6)[714.28X] required dilution prior to analysis. The reporting limits have been adjusted accordingly.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

VOLATILE ORGANIC COMPOUNDS (GCMS SIM)

Samples MW-3_052720 (240-130976-2), MW-5_052720 (240-130976-3), MW-49_052720 (240-130976-4), MW-44_052720 (240-130976-5) and DUP-03 (240-130976-6) were analyzed for volatile organic compounds (GCMS SIM) in accordance with EPA SW-846 Method 8260B SIM. The samples were analyzed on 06/09/2020.

1,2-Dichloroethane-d4 (Surr) failed the surrogate recovery criteria high for MW-44_052720MSD (240-130976-5MSD). Refer to the QC

Case Narrative

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Job ID: 240-130976-1 (Continued)

Laboratory: Eurofins TestAmerica, Canton (Continued)

report for details.

Surrogate recovery for the following sample was outside control limits: (240-130976-C-5 MSD).

No additional analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Method Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
8260B SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
5030B	Purge and Trap	SW846	TAL CAN

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396



Sample Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Asset ID
240-130976-1	TRIP BLANK	Water	05/27/20 00:00	05/29/20 10:30	
240-130976-2	MW-3_052720	Water	05/27/20 10:42	05/29/20 10:30	
240-130976-3	MW-5_052720	Water	05/27/20 12:07	05/29/20 10:30	
240-130976-4	MW-49_052720	Water	05/27/20 13:47	05/29/20 10:30	
240-130976-5	MW-44_052720	Water	05/27/20 15:12	05/29/20 10:30	
240-130976-6	DUP-03	Water	05/27/20 00:00	05/29/20 10:30	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Detection Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 240-130976-1

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
cis-1,2-Dichloroethene	0.20	J	1.0	0.16	ug/L	1		8260B	Total/NA
Vinyl chloride	0.45	J	1.0	0.20	ug/L	1		8260B	Total/NA

Client Sample ID: MW-3_052720

Lab Sample ID: 240-130976-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	1.4	J	2.0	0.86	ug/L	1		8260B SIM	Total/NA

Client Sample ID: MW-5_052720

Lab Sample ID: 240-130976-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
cis-1,2-Dichloroethene	0.18	J	1.0	0.16	ug/L	1		8260B	Total/NA

Client Sample ID: MW-49_052720

Lab Sample ID: 240-130976-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	5.8		2.0	0.86	ug/L	1		8260B SIM	Total/NA
cis-1,2-Dichloroethene	26000		1000	160	ug/L	1000		8260B	Total/NA
Vinyl chloride	15000		1000	200	ug/L	1000		8260B	Total/NA

Client Sample ID: MW-44_052720

Lab Sample ID: 240-130976-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	11		2.0	0.86	ug/L	1		8260B SIM	Total/NA
cis-1,2-Dichloroethene	2.2	J	10	1.6	ug/L	10		8260B	Total/NA
Vinyl chloride	290		10	2.0	ug/L	10		8260B	Total/NA

Client Sample ID: DUP-03

Lab Sample ID: 240-130976-6

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	9.8		2.0	0.86	ug/L	1		8260B SIM	Total/NA
cis-1,2-Dichloroethene	27000		710	110	ug/L	714.28		8260B	Total/NA
trans-1,2-Dichloroethene	140	J	710	140	ug/L	714.28		8260B	Total/NA
Vinyl chloride	15000		710	140	ug/L	714.28		8260B	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins TestAmerica, Canton

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 240-130976-1

Date Collected: 05/27/20 00:00

Matrix: Water

Date Received: 05/29/20 10:30

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/05/20 20:50	1
cis-1,2-Dichloroethene	0.20	J	1.0	0.16	ug/L			06/05/20 20:50	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/05/20 20:50	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/05/20 20:50	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/05/20 20:50	1
Vinyl chloride	0.45	J	1.0	0.20	ug/L			06/05/20 20:50	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	95		75 - 130		06/05/20 20:50	1
4-Bromofluorobenzene (Surr)	102		47 - 134		06/05/20 20:50	1
Toluene-d8 (Surr)	94		69 - 122		06/05/20 20:50	1
Dibromofluoromethane (Surr)	94		78 - 129		06/05/20 20:50	1

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: MW-3_052720

Lab Sample ID: 240-130976-2

Date Collected: 05/27/20 10:42

Matrix: Water

Date Received: 05/29/20 10:30

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	1.4	J	2.0	0.86	ug/L			06/09/20 11:01	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	99		70 - 133		06/09/20 11:01	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/05/20 21:15	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/05/20 21:15	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/05/20 21:15	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/05/20 21:15	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/05/20 21:15	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/05/20 21:15	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		75 - 130		06/05/20 21:15	1
4-Bromofluorobenzene (Surr)	103		47 - 134		06/05/20 21:15	1
Toluene-d8 (Surr)	95		69 - 122		06/05/20 21:15	1
Dibromofluoromethane (Surr)	98		78 - 129		06/05/20 21:15	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: MW-5_052720

Lab Sample ID: 240-130976-3

Date Collected: 05/27/20 12:07

Matrix: Water

Date Received: 05/29/20 10:30

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L	-		06/09/20 11:27	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	99		70 - 133		06/09/20 11:27	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L	-		06/05/20 21:46	1
cis-1,2-Dichloroethene	0.18	J	1.0	0.16	ug/L	-		06/05/20 21:46	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L	-		06/05/20 21:46	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L	-		06/05/20 21:46	1
Trichloroethene	1.0	U	1.0	0.10	ug/L	-		06/05/20 21:46	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L	-		06/05/20 21:46	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	104		75 - 130		06/05/20 21:46	1
4-Bromofluorobenzene (Surr)	101		47 - 134		06/05/20 21:46	1
Toluene-d8 (Surr)	97		69 - 122		06/05/20 21:46	1
Dibromofluoromethane (Surr)	95		78 - 129		06/05/20 21:46	1

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: MW-49_052720

Lab Sample ID: 240-130976-4

Date Collected: 05/27/20 13:47

Matrix: Water

Date Received: 05/29/20 10:30

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	5.8		2.0	0.86	ug/L			06/09/20 11:53	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		70 - 133					06/09/20 11:53	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1000	U	1000	190	ug/L			06/05/20 22:19	1000
cis-1,2-Dichloroethene	26000		1000	160	ug/L			06/05/20 22:19	1000
Tetrachloroethene	1000	U	1000	150	ug/L			06/05/20 22:19	1000
trans-1,2-Dichloroethene	1000	U	1000	190	ug/L			06/05/20 22:19	1000
Trichloroethene	1000	U	1000	100	ug/L			06/05/20 22:19	1000
Vinyl chloride	15000		1000	200	ug/L			06/05/20 22:19	1000
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	97		75 - 130					06/05/20 22:19	1000
4-Bromofluorobenzene (Surr)	97		47 - 134					06/05/20 22:19	1000
Toluene-d8 (Surr)	91		69 - 122					06/05/20 22:19	1000
Dibromofluoromethane (Surr)	97		78 - 129					06/05/20 22:19	1000

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: MW-44_052720

Lab Sample ID: 240-130976-5

Date Collected: 05/27/20 15:12

Matrix: Water

Date Received: 05/29/20 10:30

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	11		2.0	0.86	ug/L			06/09/20 12:18	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	98		70 - 133		06/09/20 12:18	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	10	U	10	1.9	ug/L			06/05/20 22:44	10
cis-1,2-Dichloroethene	2.2	J	10	1.6	ug/L			06/05/20 22:44	10
Tetrachloroethene	10	U	10	1.5	ug/L			06/05/20 22:44	10
trans-1,2-Dichloroethene	10	U	10	1.9	ug/L			06/05/20 22:44	10
Trichloroethene	10	U	10	1.0	ug/L			06/05/20 22:44	10
Vinyl chloride	290		10	2.0	ug/L			06/05/20 22:44	10

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	99		75 - 130		06/05/20 22:44	10
4-Bromofluorobenzene (Surr)	101		47 - 134		06/05/20 22:44	10
Toluene-d8 (Surr)	93		69 - 122		06/05/20 22:44	10
Dibromofluoromethane (Surr)	98		78 - 129		06/05/20 22:44	10

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: DUP-03
Date Collected: 05/27/20 00:00
Date Received: 05/29/20 10:30

Lab Sample ID: 240-130976-6
Matrix: Water

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	9.8		2.0	0.86	ug/L			06/09/20 13:36	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	90		70 - 133					06/09/20 13:36	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	710	U	710	140	ug/L			06/05/20 23:09	714.28
cis-1,2-Dichloroethene	27000		710	110	ug/L			06/05/20 23:09	714.28
Tetrachloroethene	710	U	710	110	ug/L			06/05/20 23:09	714.28
trans-1,2-Dichloroethene	140	J	710	140	ug/L			06/05/20 23:09	714.28
Trichloroethene	710	U	710	71	ug/L			06/05/20 23:09	714.28
Vinyl chloride	15000		710	140	ug/L			06/05/20 23:09	714.28
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	102		75 - 130					06/05/20 23:09	714.28
4-Bromofluorobenzene (Surr)	104		47 - 134					06/05/20 23:09	714.28
Toluene-d8 (Surr)	94		69 - 122					06/05/20 23:09	714.28
Dibromofluoromethane (Surr)	99		78 - 129					06/05/20 23:09	714.28

Surrogate Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA	BFB	TOL	DBFM
		(75-130)	(47-134)	(69-122)	(78-129)
240-130976-1	TRIP BLANK	95	102	94	94
240-130976-2	MW-3_052720	101	103	95	98
240-130976-3	MW-5_052720	104	101	97	95
240-130976-4	MW-49_052720	97	97	91	97
240-130976-5	MW-44_052720	99	101	93	98
240-130976-6	DUP-03	102	104	94	99
240-131287-A-1 MSD	Matrix Spike Duplicate	101	110	94	100
240-131287-C-1 MS	Matrix Spike	96	111	89	96
LCS 240-437098/4	Lab Control Sample	97	107	92	96
MB 240-437098/7	Method Blank	104	102	90	96

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

BFB = 4-Bromofluorobenzene (Surr)

TOL = Toluene-d8 (Surr)

DBFM = Dibromofluoromethane (Surr)

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA
		(70-133)
240-130976-2	MW-3_052720	99
240-130976-3	MW-5_052720	99
240-130976-4	MW-49_052720	91
240-130976-5	MW-44_052720	98
240-130976-5 MS	MW-44_052720	101
240-130976-5 MSD	MW-44_052720	189 X
240-130976-6	DUP-03	90
LCS 240-437419/4	Lab Control Sample	94
MB 240-437419/5	Method Blank	95

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 240-437098/7
Matrix: Water
Analysis Batch: 437098

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/05/20 15:25	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.16	ug/L			06/05/20 15:25	1
Tetrachloroethene	1.0	U	1.0	0.15	ug/L			06/05/20 15:25	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.19	ug/L			06/05/20 15:25	1
Trichloroethene	1.0	U	1.0	0.10	ug/L			06/05/20 15:25	1
Vinyl chloride	1.0	U	1.0	0.20	ug/L			06/05/20 15:25	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	104		75 - 130		06/05/20 15:25	1
4-Bromofluorobenzene (Surr)	102		47 - 134		06/05/20 15:25	1
Toluene-d8 (Surr)	90		69 - 122		06/05/20 15:25	1
Dibromofluoromethane (Surr)	96		78 - 129		06/05/20 15:25	1

Lab Sample ID: LCS 240-437098/4
Matrix: Water
Analysis Batch: 437098

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	10.0	10.4		ug/L		104	73 - 129
cis-1,2-Dichloroethene	10.0	10.4		ug/L		104	75 - 124
Tetrachloroethene	10.0	11.2		ug/L		112	70 - 125
trans-1,2-Dichloroethene	10.0	10.3		ug/L		103	74 - 130
Trichloroethene	10.0	9.47		ug/L		95	71 - 121
Vinyl chloride	10.0	13.0		ug/L		130	61 - 134

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	97		75 - 130
4-Bromofluorobenzene (Surr)	107		47 - 134
Toluene-d8 (Surr)	92		69 - 122
Dibromofluoromethane (Surr)	96		78 - 129

Lab Sample ID: 240-131287-A-1 MSD
Matrix: Water
Analysis Batch: 437098

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,1-Dichloroethene	1.0	U	10.0	10.8		ug/L		108	64 - 132	1	35
cis-1,2-Dichloroethene	1.0	U	10.0	10.1		ug/L		101	68 - 121	9	35
Tetrachloroethene	1.0	U	10.0	10.5		ug/L		105	52 - 129	4	35
trans-1,2-Dichloroethene	1.0	U	10.0	10.2		ug/L		102	69 - 126	3	35
Trichloroethene	1.0	U	10.0	9.11		ug/L		91	56 - 124	6	35
Vinyl chloride	1.0	U F1	10.0	14.9	F1	ug/L		149	49 - 136	1	35

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	101		75 - 130
4-Bromofluorobenzene (Surr)	110		47 - 134
Toluene-d8 (Surr)	94		69 - 122

Eurofins TestAmerica, Canton

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 240-131287-A-1 MSD
Matrix: Water
Analysis Batch: 437098

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

Surrogate	MSD %Recovery	MSD Qualifier	Limits
Dibromofluoromethane (Surr)	100		78 - 129

Lab Sample ID: 240-131287-C-1 MS
Matrix: Water
Analysis Batch: 437098

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
1,1-Dichloroethene	1.0	U	10.0	10.8		ug/L		108	64 - 132
cis-1,2-Dichloroethene	1.0	U	10.0	11.0		ug/L		110	68 - 121
Tetrachloroethene	1.0	U	10.0	10.1		ug/L		101	52 - 129
trans-1,2-Dichloroethene	1.0	U	10.0	10.4		ug/L		104	69 - 126
Trichloroethene	1.0	U	10.0	9.69		ug/L		97	56 - 124
Vinyl chloride	1.0	U F1	10.0	14.7	F1	ug/L		147	49 - 136

Surrogate	MS %Recovery	MS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	96		75 - 130
4-Bromofluorobenzene (Surr)	111		47 - 134
Toluene-d8 (Surr)	89		69 - 122
Dibromofluoromethane (Surr)	96		78 - 129

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 240-437419/5
Matrix: Water
Analysis Batch: 437419

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			06/09/20 06:44	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	95		70 - 133		06/09/20 06:44	1

Lab Sample ID: LCS 240-437419/4
Matrix: Water
Analysis Batch: 437419

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	10.0	8.71		ug/L		87	80 - 135

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	94		70 - 133

Lab Sample ID: 240-130976-5 MS
Matrix: Water
Analysis Batch: 437419

Client Sample ID: MW-44_052720
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	11		10.0	20.0		ug/L		85	46 - 170

Eurofins TestAmerica, Canton

QC Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Method: 8260B SIM - Volatile Organic Compounds (GC/MS) (Continued)

<i>Surrogate</i>	<i>MS</i> <i>%Recovery</i>	<i>MS</i> <i>Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	101		70 - 133

Lab Sample ID: 240-130976-5 MSD
 Matrix: Water
 Analysis Batch: 437419

Client Sample ID: MW-44_052720
 Prep Type: Total/NA

<i>Analyte</i>	<i>Sample</i> <i>Result</i>	<i>Sample</i> <i>Qualifier</i>	<i>Spike</i> <i>Added</i>	<i>MSD</i> <i>Result</i>	<i>MSD</i> <i>Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec.</i> <i>Limits</i>	<i>RPD</i>	<i>RPD</i> <i>Limit</i>
1,4-Dioxane	11		10.0	19.3		ug/L		78	46 - 170	4	26

<i>Surrogate</i>	<i>MSD</i> <i>%Recovery</i>	<i>MSD</i> <i>Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	189	X	70 - 133



QC Association Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

GC/MS VOA

Analysis Batch: 437098

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-130976-1	TRIP BLANK	Total/NA	Water	8260B	
240-130976-2	MW-3_052720	Total/NA	Water	8260B	
240-130976-3	MW-5_052720	Total/NA	Water	8260B	
240-130976-4	MW-49_052720	Total/NA	Water	8260B	
240-130976-5	MW-44_052720	Total/NA	Water	8260B	
240-130976-6	DUP-03	Total/NA	Water	8260B	
MB 240-437098/7	Method Blank	Total/NA	Water	8260B	
LCS 240-437098/4	Lab Control Sample	Total/NA	Water	8260B	
240-131287-A-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B	
240-131287-C-1 MS	Matrix Spike	Total/NA	Water	8260B	

Analysis Batch: 437419

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-130976-2	MW-3_052720	Total/NA	Water	8260B SIM	
240-130976-3	MW-5_052720	Total/NA	Water	8260B SIM	
240-130976-4	MW-49_052720	Total/NA	Water	8260B SIM	
240-130976-5	MW-44_052720	Total/NA	Water	8260B SIM	
240-130976-6	DUP-03	Total/NA	Water	8260B SIM	
MB 240-437419/5	Method Blank	Total/NA	Water	8260B SIM	
LCS 240-437419/4	Lab Control Sample	Total/NA	Water	8260B SIM	
240-130976-5 MS	MW-44_052720	Total/NA	Water	8260B SIM	
240-130976-5 MSD	MW-44_052720	Total/NA	Water	8260B SIM	

Lab Chronicle

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Client Sample ID: TRIP BLANK

Lab Sample ID: 240-130976-1

Date Collected: 05/27/20 00:00

Matrix: Water

Date Received: 05/29/20 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	437098	06/05/20 20:50	LRW	TAL CAN

Client Sample ID: MW-3_052720

Lab Sample ID: 240-130976-2

Date Collected: 05/27/20 10:42

Matrix: Water

Date Received: 05/29/20 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	437098	06/05/20 21:15	LRW	TAL CAN
Total/NA	Analysis	8260B SIM		1	437419	06/09/20 11:01	SAM	TAL CAN

Client Sample ID: MW-5_052720

Lab Sample ID: 240-130976-3

Date Collected: 05/27/20 12:07

Matrix: Water

Date Received: 05/29/20 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	437098	06/05/20 21:46	LRW	TAL CAN
Total/NA	Analysis	8260B SIM		1	437419	06/09/20 11:27	SAM	TAL CAN

Client Sample ID: MW-49_052720

Lab Sample ID: 240-130976-4

Date Collected: 05/27/20 13:47

Matrix: Water

Date Received: 05/29/20 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1000	437098	06/05/20 22:19	LRW	TAL CAN
Total/NA	Analysis	8260B SIM		1	437419	06/09/20 11:53	SAM	TAL CAN

Client Sample ID: MW-44_052720

Lab Sample ID: 240-130976-5

Date Collected: 05/27/20 15:12

Matrix: Water

Date Received: 05/29/20 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		10	437098	06/05/20 22:44	LRW	TAL CAN
Total/NA	Analysis	8260B SIM		1	437419	06/09/20 12:18	SAM	TAL CAN

Client Sample ID: DUP-03

Lab Sample ID: 240-130976-6

Date Collected: 05/27/20 00:00

Matrix: Water

Date Received: 05/29/20 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		714.28	437098	06/05/20 23:09	LRW	TAL CAN
Total/NA	Analysis	8260B SIM		1	437419	06/09/20 13:36	SAM	TAL CAN

Laboratory References:

TAL CAN = Eurofins TestAmerica, Canton, 4101 Shuffel Street NW, North Canton, OH 44720, TEL (330)497-9396

Eurofins TestAmerica, Canton

Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP On-Site

Job ID: 240-130976-1

Laboratory: Eurofins TestAmerica, Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	2927	02-23-21
Connecticut	State	PH-0590	12-31-21
Florida	NELAP	E87225	06-30-20
Georgia	State	4062	02-23-21
Illinois	NELAP	004498	07-31-20
Iowa	State	421	06-01-21
Kansas	NELAP	E-10336	04-30-21
Kentucky (UST)	State	112225	02-23-21
Kentucky (WW)	State	KY98016	12-31-20
Minnesota	NELAP	OH00048	12-31-20
Minnesota (Petrofund)	State	3506	08-01-21
New Jersey	NELAP	OH001	06-30-20
New York	NELAP	10975	03-31-21
Ohio VAP	State	CL0024	06-05-21
Oregon	NELAP	4062	02-24-21
Pennsylvania	NELAP	68-00340	08-31-20
Texas	NELAP	T104704517-18-10	08-31-20
USDA	US Federal Programs	P330-18-00281	09-17-21
Virginia	NELAP	010101	09-14-20
Washington	State	C971	01-12-21
West Virginia DEP	State	210	12-31-20



TestAmerica Laboratory location: Brighton --- 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

Regulatory program: DW NPDES RCRA Other

Client Contact: Arcadis
Address: 28550 Cabot Drive, Suite 500
City/State/Zip: Novi, MI, 48377
Phone: 248-994-2240

Client Project Manager: Kris Hinsky
Telephone: 248-994-2240
Email: kristoffer.hinsky@arcadis.com

Site Contact: Julia McClafferty
Telephone: 734-644-5131

Lab Contact: Mike DelMontico
Telephone: 330-497-9396

Company Name: Arcadis
Address: 28550 Cabot Drive, Suite 500
City/State/Zip: Novi, MI, 48377
Phone: 248-994-2240

Project Name: Ford LTP On-Site
Project Number: 30050315.401.03
PO # 30050315.401.03

Sample Identification	Sample Date	Sample Time	Matrix			Containers & Preservatives					Filtered Sample (Y/N)	Analyses						Sample Specific Notes / Special Instructions:		
			Air	Aqueous	Solid	Other:	H2SO4	HNO3	HCl	NaOH		ZnAc	NaOH	Other:	1,1-DCE 8260B	GIS-1,2-DCE 8260B	Trans-1,2-DCE 8260B		PCE 8260B	TCE 8260B
TRIP BLANK	5/27/20	1042	6									NG	X	X	X	X	X	X	X	TRIP BLANK
MW-3-052720	5/27/20	1207	6									NG	X	X	X	X	X	X	X	3 VOA for 8260B 3 VOA for 8260B
MW-5-052720	5/27/20	1347	6									NG	X	X	X	X	X	X	X	
MW-44-052720	5/27/20	1512	6									NG	X	X	X	X	X	X	X	
DUP-03	5/27/20		6									NG	X	X	X	X	X	X	X	



Possible Hazard Identification
 Non-Hazard Irritant Inflammable Corrosive Toxic Other

Special Instructions/QC Requirements & Comments:

Submit all results through Cadena at jtomalia@cadenaco.com, Cadena #E203728
Level IV Reporting requested.

Relinquished by:	Company:	Date/Time:	Received by:	Company:	Date/Time:
<i>Heather Woodrum</i>	Arcadis	5/27/20 1830	<i>Novi Cold Storage</i>	Arcadis	5/27/20 1830
<i>John Banta</i>	Arcadis	5/28/20 0856	<i>ETA</i>	ETA	5/28/20 0857
<i>ETA</i>	ETA	5/27/20 0857	<i>ETA</i>	ETA	5/27/20 1050


**Eurofins TestAmerica Canton Sample Receipt Form/Narrative
Canton Facility**

Login # : 130976

Client Arcadis Site Name _____ Cooler unpacked by: [Signature]
 Cooler Received on 5-29-20 Opened on 5-29-20
 FedEx: 1st Grd Exp UPS FAS Clipper Client Drop Off TestAmerica Courier Other _____

Receipt After-hours: Drop-off Date/Time _____ Storage Location _____

TestAmerica Cooler # NA Foam Box _____ Client Cooler _____ Box _____ Other _____
 Packing material used: Bubble Wrap Foam Plastic Bag None _____ Other _____
 COOLANT: Wet Ice Blue Ice Dry Ice Water None _____

1. Cooler temperature upon receipt See Multiple Cooler Form
 IR GUN# IR-10 (CF +0.7°C) Observed Cooler Temp. 3.1 °C Corrected Cooler Temp. 3.8 °C
 IR GUN #IR-11 (CF +0.9°C) Observed Cooler Temp. _____ °C Corrected Cooler Temp. _____ °C
2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 1 Yes No
 -Were the seals on the outside of the cooler(s) signed & dated? Yes No NA
 -Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No
 -Were tamper/custody seals intact and uncompromised? Yes No NA
3. Shippers' packing slip attached to the cooler(s)? Yes No
4. Did custody papers accompany the sample(s)? Yes No
5. Were the custody papers relinquished & signed in the appropriate place? Yes No
6. Was/were the person(s) who collected the samples clearly identified on the COC? Yes No
7. Did all bottles arrive in good condition (Unbroken)? Yes No
8. Could all bottle labels be reconciled with the COC? Yes No
9. Were correct bottle(s) used for the test(s) indicated? Yes No
10. Sufficient quantity received to perform indicated analyses? Yes No
11. Are these work share samples? Yes No
12. Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC902937
13. Were VOAs on the COC? Yes No
14. Were air bubbles >6 mm in any VOA vials?  Larger than this. NA Yes No NA
15. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # NA Yes No
16. Was a LL Hg or Me Hg trip blank present? Yes No

Tests that are not checked for pH by Receiving:
 VOAs
 Oil and Grease
 TOC

Contacted PM _____ Date _____ by _____ via Verbal Voice Mail Other _____
 Concerning _____

17. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES Samples processed by: MS

18. SAMPLE CONDITION
 Sample(s) _____ were received after the recommended holding time had expired.
 Sample(s) _____ were received in a broken container.
 Sample(s) _____ were received with bubble >6 mm in diameter. (Notify PM)

19. SAMPLE PRESERVATION
 Sample(s) _____ were further preserved in the laboratory.
 Time preserved: _____ Preservative(s) added/Lot number(s): _____
 VOA Sample Preservation - Date/Time VOAs Frozen: _____