

ANALYTICAL REPORT

Eurofins Canton
180 S. Van Buren Avenue
Barberton, OH 44203
Tel: (330)497-9396

Laboratory Job ID: 240-162541-1
Client Project/Site: Ford LTP - Off-Site

For:
ARCADIS U.S., Inc.
28550 Cabot Drive
Suite 500
Novi, Michigan 48377

Attn: Kristoffer Hinskey



Authorized for release by:
2/23/2022 9:44:32 AM

Michael DelMonico, Project Manager I
(330)497-9396
Michael.DelMonico@Eurofinset.com

LINKS

Review your project
results through
TotalAccess

Have a Question?



Visit us at:

www.eurofinsus.com/Env

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Method Summary	5
Sample Summary	6
Detection Summary	7
Client Sample Results	8
Surrogate Summary	12
QC Sample Results	13
QC Association Summary	16
Lab Chronicle	17
Certification Summary	18
Chain of Custody	19

Definitions/Glossary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Job ID: 240-162541-1

Laboratory: Eurofins Canton

Narrative

**Job Narrative
240-162541-1**

Comments

No additional comments.

Receipt

The samples were received on 2/9/2022 12:55 PM. Unless otherwise noted below, the samples arrived in good condition, and where required, properly preserved and on ice. The temperature of the cooler at receipt was 1.3° C.

GC/MS VOA

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

VOA Prep

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Method Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Method	Method Description	Protocol	Laboratory
8260B	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
8260B SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
5030B	Purge and Trap	SW846	TAL CAN

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL CAN = Eurofins Canton, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Sample Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
240-162541-1	TRIP BLANK_01	Water	02/07/22 00:00	02/09/22 12:55
240-162541-2	MW-73D_020722	Water	02/07/22 10:25	02/09/22 12:55
240-162541-3	MW-73SR_020722	Water	02/07/22 11:46	02/09/22 12:55
240-162541-4	MW-74_020722	Water	02/07/22 13:17	02/09/22 12:55

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Detection Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Client Sample ID: TRIP BLANK_01

Lab Sample ID: 240-162541-1

No Detections.

Client Sample ID: MW-73D_020722

Lab Sample ID: 240-162541-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
cis-1,2-Dichloroethene	2.1		1.0	0.46	ug/L	1		8260B	Total/NA

Client Sample ID: MW-73SR_020722

Lab Sample ID: 240-162541-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	1.2	J	2.0	0.86	ug/L	1		8260B SIM	Total/NA

Client Sample ID: MW-74_020722

Lab Sample ID: 240-162541-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	0.90	J	2.0	0.86	ug/L	1		8260B SIM	Total/NA
Vinyl chloride	2.5		1.0	0.45	ug/L	1		8260B	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Canton

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Client Sample ID: TRIP BLANK_01

Lab Sample ID: 240-162541-1

Date Collected: 02/07/22 00:00

Matrix: Water

Date Received: 02/09/22 12:55

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			02/10/22 12:23	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			02/10/22 12:23	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 12:23	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			02/10/22 12:23	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 12:23	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			02/10/22 12:23	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	94		62 - 137		02/10/22 12:23	1
4-Bromofluorobenzene (Surr)	95		56 - 136		02/10/22 12:23	1
Toluene-d8 (Surr)	99		78 - 122		02/10/22 12:23	1
Dibromofluoromethane (Surr)	105		73 - 120		02/10/22 12:23	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Client Sample ID: MW-73D_020722

Lab Sample ID: 240-162541-2

Date Collected: 02/07/22 10:25

Matrix: Water

Date Received: 02/09/22 12:55

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			02/10/22 17:57	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	75		66 - 120		02/10/22 17:57	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			02/10/22 13:11	1
cis-1,2-Dichloroethene	2.1		1.0	0.46	ug/L			02/10/22 13:11	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 13:11	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			02/10/22 13:11	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 13:11	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			02/10/22 13:11	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	90		62 - 137		02/10/22 13:11	1
4-Bromofluorobenzene (Surr)	96		56 - 136		02/10/22 13:11	1
Toluene-d8 (Surr)	98		78 - 122		02/10/22 13:11	1
Dibromofluoromethane (Surr)	102		73 - 120		02/10/22 13:11	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Client Sample ID: MW-73SR_020722

Lab Sample ID: 240-162541-3

Date Collected: 02/07/22 11:46

Matrix: Water

Date Received: 02/09/22 12:55

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	1.2	J	2.0	0.86	ug/L			02/10/22 19:13	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	77		66 - 120		02/10/22 19:13	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			02/10/22 13:35	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			02/10/22 13:35	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 13:35	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			02/10/22 13:35	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 13:35	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			02/10/22 13:35	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	92		62 - 137		02/10/22 13:35	1
4-Bromofluorobenzene (Surr)	95		56 - 136		02/10/22 13:35	1
Toluene-d8 (Surr)	95		78 - 122		02/10/22 13:35	1
Dibromofluoromethane (Surr)	101		73 - 120		02/10/22 13:35	1

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Client Sample ID: MW-74_020722

Lab Sample ID: 240-162541-4

Date Collected: 02/07/22 13:17

Matrix: Water

Date Received: 02/09/22 12:55

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	0.90	J	2.0	0.86	ug/L			02/10/22 19:38	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	76		66 - 120					02/10/22 19:38	1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			02/10/22 13:58	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			02/10/22 13:58	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 13:58	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			02/10/22 13:58	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 13:58	1
Vinyl chloride	2.5		1.0	0.45	ug/L			02/10/22 13:58	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	92		62 - 137					02/10/22 13:58	1
4-Bromofluorobenzene (Surr)	92		56 - 136					02/10/22 13:58	1
Toluene-d8 (Surr)	98		78 - 122					02/10/22 13:58	1
Dibromofluoromethane (Surr)	104		73 - 120					02/10/22 13:58	1

Surrogate Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCA (62-137)	BFB (56-136)	TOL (78-122)	DBFM (73-120)
240-162541-1	TRIP BLANK_01	94	95	99	105
240-162541-2	MW-73D_020722	90	96	98	102
240-162541-3	MW-73SR_020722	92	95	95	101
240-162541-4	MW-74_020722	92	92	98	104
240-162542-B-3 MS	Matrix Spike	98	105	107	110
240-162542-B-3 MSD	Matrix Spike Duplicate	100	105	109	111
LCS 240-517755/5	Lab Control Sample	88	99	103	101
MB 240-517755/7	Method Blank	91	98	101	103

Surrogate Legend
DCA = 1,2-Dichloroethane-d4 (Surr)
BFB = 4-Bromofluorobenzene (Surr)
TOL = Toluene-d8 (Surr)
DBFM = Dibromofluoromethane (Surr)

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		DCA (66-120)
240-162541-2	MW-73D_020722	75
240-162541-2 MS	MW-73D_020722	76
240-162541-2 MSD	MW-73D_020722	80
240-162541-3	MW-73SR_020722	77
240-162541-4	MW-74_020722	76
LCS 240-517810/4	Lab Control Sample	80
MB 240-517810/5	Method Blank	80

Surrogate Legend
DCA = 1,2-Dichloroethane-d4 (Surr)

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		DCA (10-150)
MRL 240-517810/6	Lab Control Sample	77

Surrogate Legend
DCA = 1,2-Dichloroethane-d4 (Surr)

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Method: 8260B - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 240-517755/7
Matrix: Water
Analysis Batch: 517755

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			02/10/22 11:59	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			02/10/22 11:59	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 11:59	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			02/10/22 11:59	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			02/10/22 11:59	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			02/10/22 11:59	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
1,2-Dichloroethane-d4 (Surr)	91		62 - 137		02/10/22 11:59	1
4-Bromofluorobenzene (Surr)	98		56 - 136		02/10/22 11:59	1
Toluene-d8 (Surr)	101		78 - 122		02/10/22 11:59	1
Dibromofluoromethane (Surr)	103		73 - 120		02/10/22 11:59	1

Lab Sample ID: LCS 240-517755/5
Matrix: Water
Analysis Batch: 517755

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec. Limits
		Result	Qualifier				
1,1-Dichloroethene	25.0	27.9		ug/L		111	63 - 134
cis-1,2-Dichloroethene	25.0	25.4		ug/L		101	77 - 123
Tetrachloroethene	25.0	26.8		ug/L		107	76 - 123
trans-1,2-Dichloroethene	25.0	26.5		ug/L		106	75 - 124
Trichloroethene	25.0	25.2		ug/L		101	70 - 122
Vinyl chloride	25.0	22.9		ug/L		92	60 - 144

Surrogate	LCS	LCS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	88		62 - 137
4-Bromofluorobenzene (Surr)	99		56 - 136
Toluene-d8 (Surr)	103		78 - 122
Dibromofluoromethane (Surr)	101		73 - 120

Lab Sample ID: 240-162542-B-3 MS
Matrix: Water
Analysis Batch: 517755

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample	Sample	Spike Added	MS	MS	Unit	D	%Rec	%Rec. Limits
	Result	Qualifier		Result	Qualifier				
1,1-Dichloroethene	33	U	833	898		ug/L		108	56 - 135
cis-1,2-Dichloroethene	1100		833	1910		ug/L		103	66 - 128
Tetrachloroethene	33	U	833	842		ug/L		101	62 - 131
trans-1,2-Dichloroethene	33	U	833	844		ug/L		101	56 - 136
Trichloroethene	33	U	833	833		ug/L		100	61 - 124
Vinyl chloride	870		833	1570		ug/L		83	43 - 157

Surrogate	MS	MS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	98		62 - 137
4-Bromofluorobenzene (Surr)	105		56 - 136
Toluene-d8 (Surr)	107		78 - 122

Eurofins Canton

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Method: 8260B - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 240-162542-B-3 MS
Matrix: Water
Analysis Batch: 517755

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Surrogate	MS %Recovery	MS Qualifier	Limits
Dibromofluoromethane (Surr)	110		73 - 120

Lab Sample ID: 240-162542-B-3 MSD
Matrix: Water
Analysis Batch: 517755

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	RPD Limit
1,1-Dichloroethene	33	U	833	940		ug/L		113	56 - 135	5	26
cis-1,2-Dichloroethene	1100		833	1900		ug/L		102	66 - 128	0	14
Tetrachloroethene	33	U	833	847		ug/L		102	62 - 131	1	20
trans-1,2-Dichloroethene	33	U	833	866		ug/L		104	56 - 136	3	15
Trichloroethene	33	U	833	860		ug/L		103	61 - 124	3	15
Vinyl chloride	870		833	1600		ug/L		87	43 - 157	2	24

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	100		62 - 137
4-Bromofluorobenzene (Surr)	105		56 - 136
Toluene-d8 (Surr)	109		78 - 122
Dibromofluoromethane (Surr)	111		73 - 120

Method: 8260B SIM - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 240-517810/5
Matrix: Water
Analysis Batch: 517810

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			02/10/22 17:07	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	80		66 - 120		02/10/22 17:07	1

Lab Sample ID: LCS 240-517810/4
Matrix: Water
Analysis Batch: 517810

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	10.0	9.38		ug/L		94	80 - 122

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	80		66 - 120

Lab Sample ID: MRL 240-517810/6
Matrix: Water
Analysis Batch: 517810

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec. Limits
1,4-Dioxane	0.00100	0.0020	U	ng/uL		83	10 - 150

Eurofins Canton

QC Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Method: 8260B SIM - Volatile Organic Compounds (GC/MS) (Continued)

<i>Surrogate</i>	<i>MRL %Recovery</i>	<i>MRL Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	77		10 - 150

Lab Sample ID: 240-162541-2 MS
Matrix: Water
Analysis Batch: 517810

Client Sample ID: MW-73D_020722
Prep Type: Total/NA

<i>Analyte</i>	<i>Sample Result</i>	<i>Sample Qualifier</i>	<i>Spike Added</i>	<i>MS Result</i>	<i>MS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec. Limits</i>
1,4-Dioxane	2.0	U	10.0	10.1		ug/L		101	51 - 153

<i>Surrogate</i>	<i>MS %Recovery</i>	<i>MS Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	76		66 - 120

Lab Sample ID: 240-162541-2 MSD
Matrix: Water
Analysis Batch: 517810

Client Sample ID: MW-73D_020722
Prep Type: Total/NA

<i>Analyte</i>	<i>Sample Result</i>	<i>Sample Qualifier</i>	<i>Spike Added</i>	<i>MSD Result</i>	<i>MSD Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec. Limits</i>	<i>RPD</i>	<i>RPD Limit</i>
1,4-Dioxane	2.0	U	10.0	9.91		ug/L		99	51 - 153	2	16

<i>Surrogate</i>	<i>MSD %Recovery</i>	<i>MSD Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	80		66 - 120

QC Association Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

GC/MS VOA

Analysis Batch: 517755

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-162541-1	TRIP BLANK_01	Total/NA	Water	8260B	
240-162541-2	MW-73D_020722	Total/NA	Water	8260B	
240-162541-3	MW-73SR_020722	Total/NA	Water	8260B	
240-162541-4	MW-74_020722	Total/NA	Water	8260B	
MB 240-517755/7	Method Blank	Total/NA	Water	8260B	
LCS 240-517755/5	Lab Control Sample	Total/NA	Water	8260B	
240-162542-B-3 MS	Matrix Spike	Total/NA	Water	8260B	
240-162542-B-3 MSD	Matrix Spike Duplicate	Total/NA	Water	8260B	

Analysis Batch: 517810

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-162541-2	MW-73D_020722	Total/NA	Water	8260B SIM	
240-162541-3	MW-73SR_020722	Total/NA	Water	8260B SIM	
240-162541-4	MW-74_020722	Total/NA	Water	8260B SIM	
MB 240-517810/5	Method Blank	Total/NA	Water	8260B SIM	
LCS 240-517810/4	Lab Control Sample	Total/NA	Water	8260B SIM	
MRL 240-517810/6	Lab Control Sample	Total/NA	Water	8260B SIM	
240-162541-2 MS	MW-73D_020722	Total/NA	Water	8260B SIM	
240-162541-2 MSD	MW-73D_020722	Total/NA	Water	8260B SIM	

Lab Chronicle

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Client Sample ID: TRIP BLANK_01

Lab Sample ID: 240-162541-1

Date Collected: 02/07/22 00:00

Matrix: Water

Date Received: 02/09/22 12:55

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	517755	02/10/22 12:23	SAM	TAL CAN

Client Sample ID: MW-73D_020722

Lab Sample ID: 240-162541-2

Date Collected: 02/07/22 10:25

Matrix: Water

Date Received: 02/09/22 12:55

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	517755	02/10/22 13:11	SAM	TAL CAN
Total/NA	Analysis	8260B SIM		1	517810	02/10/22 17:57	CS	TAL CAN

Client Sample ID: MW-73SR_020722

Lab Sample ID: 240-162541-3

Date Collected: 02/07/22 11:46

Matrix: Water

Date Received: 02/09/22 12:55

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	517755	02/10/22 13:35	SAM	TAL CAN
Total/NA	Analysis	8260B SIM		1	517810	02/10/22 19:13	CS	TAL CAN

Client Sample ID: MW-74_020722

Lab Sample ID: 240-162541-4

Date Collected: 02/07/22 13:17

Matrix: Water

Date Received: 02/09/22 12:55

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260B		1	517755	02/10/22 13:58	SAM	TAL CAN
Total/NA	Analysis	8260B SIM		1	517810	02/10/22 19:38	CS	TAL CAN

Laboratory References:

TAL CAN = Eurofins Canton, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - Off-Site

Job ID: 240-162541-1

Laboratory: Eurofins Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	2927	02-23-22
Connecticut	State	PH-0590	12-31-21 *
Florida	NELAP	E87225	06-30-22
Georgia	State	4062	02-23-22
Illinois	NELAP	200004	07-31-22
Iowa	State	421	06-01-23
Kansas	NELAP	E-10336	04-30-22
Kentucky (UST)	State	112225	02-23-22
Kentucky (WW)	State	KY98016	12-31-22
Minnesota	NELAP	039-999-348	12-31-22
Minnesota (Petrofund)	State	3506	08-01-23
New Jersey	NELAP	OH001	11-06-22
New York	NELAP	10975	03-31-22
Ohio	State	8303	02-23-23
Ohio VAP	State	CL0024	12-21-23
Oregon	NELAP	4062	02-23-22
Pennsylvania	NELAP	68-00340	08-31-22
Texas	NELAP	T104704517-21-14	08-31-22
Virginia	NELAP	11570	09-14-22
Washington	State	C971	01-12-23
West Virginia DEP	State	210	12-31-22

* Accreditation/Certification renewal pending - accreditation/certification considered valid.



TestAmerica Laboratory location: Brighton — 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

Client Contact Company Name: Arcadis Address: 28550 Cabot Drive, Suite 500 City/State/Zip: Novi, MI, 48377 Phone: 248-994-2240 Project Name: Ford LTP Off-Site Project Number: 30080642.402.04 PO # 30080642.402.04		Regulatory program: <input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other	
Client Project Manager: Kris Hinskey Telephone: 248-994-2240 Email: kristoffer.hinskey@arcadis.com		Lab Contact: Mike DelMonico Telephone: 330-497-5396	
Sampler Name: <i>Christy Ken Gumbel</i>		Analyses Walk-in client Lab sampling Job/SDG No.:	
Method of Shipment/Carrier: Shipping/Tracking No.:		For lab use only 1 of 1 COCs	
Analysis Turnaround Time TAT, if different from below 10 day <input type="checkbox"/> 3 weeks <input checked="" type="checkbox"/> 2 weeks <input type="checkbox"/> 1 week <input type="checkbox"/> 2 days <input type="checkbox"/> 1 day		Filtered Sample (Y/N) Composite C / Grab = G 1,1-DCE 8260B cis-1,2-DCE 8260B Trans-1,2-DCE 8260B PCE 8260B TCE 8260B Vinyl Chloride 8260B 1,4-Dioxane 8260B SIM	
Containers & Preservatives H2SO4 HNO3 HCl NaOH ZnAc Unpres Other:		Sample Specific Notes / Special Instructions: 1 Trip Blank 3 VOAs for 8260B 3 VOAs for 8260B SIM	
Matrix Air Aqueous Sediment Solid Other:		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return to Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months	
Possible Hazard Identification <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant		Sample Date Sample Time 02/07/22 ✓ 2/7/22 1025 2/7/22 1146 2/7/22 1317	
Special Instructions/OC Requirements & Comments: Sample Address: <i>Bellevue Ct. Row</i> Submit all results through Cadena at jtomalia@cadenaco.com . Cadena #E203631 Level IV Reporting requested.		Date/Time: 2/7/22 Date/Time: 2/8/22 Date/Time: 2-9-22	
Relinquished by: <i>Christy Ken Gumbel</i> Relinquished by: <i>NOVI COLD STORAGE</i> Relinquished by: <i>Daniel</i>		Company: Arcadis Company: Arcadis Company: Eurofins	
Received by: <i>Novi Cold Storage</i> Received by: <i>Daniel</i> Received in Laboratory by: <i>Danny Boyer</i>		Date/Time: 2/7/22 1420 Date/Time: 2-8-22 1000 Date/Time: 2-9-22 1255	



©2008 TestAmerica Laboratories, Inc. All rights reserved. TestAmerica & Design are trademarks of TestAmerica Laboratories, Inc.

Eurofins TestAmerica Canton Sample Receipt Form/Narrative Login # : 162541
Canton Facility

Client Arcadis Site Name _____ Cooler unpacked by: _____

Cooler Received on 2-9-22 Opened on 2-9-22

FedEx: 1st Grd Exp UPS FAS Clipper Client Drop Off TestAmerica Courier Other

Receipt After-hours: Drop-off Date/Time _____ Storage Location _____

TestAmerica Cooler # 1A Foam Box Client Cooler Box Other _____

Packing material used: Bubble Wrap Foam Plastic Bag None Other _____

COOLANT: Wet Ice Blue Ice Dry Ice Water None

1. Cooler temperature upon receipt See Multiple Cooler Form

IR GUN# IR-14 (CF +0.1 °C) Observed Cooler Temp. 1.2 °C Corrected Cooler Temp. 1.3 °C

IR GUN #IR-15 (CF +0.2°C) Observed Cooler Temp. _____ °C Corrected Cooler Temp. _____ °C

2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 1 Yes No

-Were the seals on the outside of the cooler(s) signed & dated? Yes No NA

-Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No NA

-Were tamper/custody seals intact and uncompromised? Yes No NA

3. Shippers' packing slip attached to the cooler(s)? Yes No

4. Did custody papers accompany the sample(s)? Yes No

5. Were the custody papers relinquished & signed in the appropriate place? Yes No

6. Was/were the person(s) who collected the samples clearly identified on the COC? Yes No

7. Did all bottles arrive in good condition (Unbroken)? Yes No

8. Could all bottle labels (ID/Date/Time) be reconciled with the COC? Yes No

9. For each sample, does the COC specify preservatives (Y/N), # of containers (Y/N), and sample type of grab/comp (Y/N)?

10. Were correct bottle(s) used for the test(s) indicated? Yes No

11. Sufficient quantity received to perform indicated analyses? Yes No

12. Are these work share samples and all listed on the COC? Yes No

If yes, Questions 13-17 have been checked at the originating laboratory.

13. Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC157842

14. Were VOAs on the COC? Yes No

15. Were air bubbles >6 mm in any VOA vials? Larger than this. Yes No NA

16. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # 60358 Yes No

17. Was a LL Hg or Me Hg trip blank present? Yes No

Tests that are not checked for pH by Receiving:
VOAs
Oil and Grease
TOC

Contacted PM _____ Date _____ by _____ via Verbal Voice Mail Other

Concerning _____

18. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES additional next page Samples processed by: _____

The COC is not marked for SIM. The samples were logged for SIM per client. The TD is not logged for SIM due to insufficient volume. JME 2/9/22

19. SAMPLE CONDITION

Sample(s) _____ were received after the recommended holding time had expired.

Sample(s) _____ were received in a broken container.

Sample(s) _____ were received with bubble >6 mm in diameter. (Notify PM)

20. SAMPLE PRESERVATION

Sample(s) _____ were further preserved in the laboratory.

Time preserved: _____ Preservative(s) added/Lot number(s): _____

VOA Sample Preservation - Date/Time VOAs Frozen: _____