


ANALYTICAL REPORT

Eurofins Canton
180 S. Van Buren Avenue
Barberton, OH 44203
Tel: (330)497-9396

Laboratory Job ID: 240-166736-1
Client Project/Site: Ford LTP - On Site

For:
ARCADIS U.S., Inc.
28550 Cabot Drive
Suite 500
Novi, Michigan 48377

Attn: Kristoffer Hinskey



Authorized for release by:

5/27/2022 7:13:54 PM

Patrick O'Meara, Manager of Project Management
(330)966-5725

Patrick.O'Meara@et.eurofinsus.com

Designee for

Michael DelMonico, Project Manager I
(330)497-9396

Michael.DelMonico@et.eurofinsus.com

LINKS

Review your project
results through



Have a Question?



Visit us at:

www.eurofinsus.com/Env

This report has been electronically signed and authorized by the signatory. Electronic signature is intended to be the legally binding equivalent of a traditionally handwritten signature.

Results relate only to the items tested and the sample(s) as received by the laboratory.



Table of Contents

Cover Page	1
Table of Contents	2
Definitions/Glossary	3
Case Narrative	4
Method Summary	5
Sample Summary	6
Detection Summary	7
Client Sample Results	8
Surrogate Summary	13
QC Sample Results	14
QC Association Summary	17
Lab Chronicle	18
Certification Summary	19
Chain of Custody	20

Definitions/Glossary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Job ID: 240-166736-1

Laboratory: Eurofins Canton

Narrative

**Job Narrative
240-166736-1**

Comments

No additional comments.

Receipt

The samples were received on 5/17/2022 @ 9:30 AM. Unless otherwise noted below, the samples arrived in good condition, and where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 0.6° C and 2.1° C.

GC/MS VOA

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

VOA Prep

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Method Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Method	Method Description	Protocol	Laboratory
8260D	Volatile Organic Compounds by GC/MS	SW846	TAL CAN
8260D SIM	Volatile Organic Compounds (GC/MS)	SW846	TAL CAN
5030C	Purge and Trap	SW846	TAL CAN

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

TAL CAN = Eurofins Canton, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Sample Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
240-166736-1	TRIP BLANK_107	Water	05/13/22 00:00	05/17/22 09:30
240-166736-2	MW-67_051322	Water	05/13/22 10:20	05/17/22 09:30
240-166736-3	MW-54S_051322	Water	05/13/22 11:30	05/17/22 09:30
240-166736-4	MW-54_051322	Water	05/13/22 12:45	05/17/22 09:30
240-166736-5	MW-57_051322	Water	05/13/22 14:35	05/17/22 09:30

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Detection Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: TRIP BLANK_107

Lab Sample ID: 240-166736-1

No Detections.

Client Sample ID: MW-67_051322

Lab Sample ID: 240-166736-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
cis-1,2-Dichloroethene	3.3		1.0	0.46	ug/L	1		8260D	Total/NA
trans-1,2-Dichloroethene	0.59	J	1.0	0.51	ug/L	1		8260D	Total/NA
Trichloroethene	52		1.0	0.44	ug/L	1		8260D	Total/NA

Client Sample ID: MW-54S_051322

Lab Sample ID: 240-166736-3

No Detections.

Client Sample ID: MW-54_051322

Lab Sample ID: 240-166736-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	1.3	J	2.0	0.86	ug/L	1		8260D SIM	Total/NA
Vinyl chloride	0.89	J	1.0	0.45	ug/L	1		8260D	Total/NA

Client Sample ID: MW-57_051322

Lab Sample ID: 240-166736-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	0.98	J	2.0	0.86	ug/L	1		8260D SIM	Total/NA
Vinyl chloride	0.77	J	1.0	0.45	ug/L	1		8260D	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Canton

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: TRIP BLANK_107

Lab Sample ID: 240-166736-1

Date Collected: 05/13/22 00:00

Matrix: Water

Date Received: 05/17/22 09:30

Method: 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/25/22 17:22	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/25/22 17:22	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 17:22	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/25/22 17:22	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 17:22	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/25/22 17:22	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	105		62 - 137		05/25/22 17:22	1
4-Bromofluorobenzene (Surr)	97		56 - 136		05/25/22 17:22	1
Toluene-d8 (Surr)	101		78 - 122		05/25/22 17:22	1
Dibromofluoromethane (Surr)	96		73 - 120		05/25/22 17:22	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: MW-67_051322

Lab Sample ID: 240-166736-2

Date Collected: 05/13/22 10:20

Matrix: Water

Date Received: 05/17/22 09:30

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/24/22 02:24	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	105		66 - 120		05/24/22 02:24	1

Method: 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/25/22 17:44	1
cis-1,2-Dichloroethene	3.3		1.0	0.46	ug/L			05/25/22 17:44	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 17:44	1
trans-1,2-Dichloroethene	0.59	J	1.0	0.51	ug/L			05/25/22 17:44	1
Trichloroethene	52		1.0	0.44	ug/L			05/25/22 17:44	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/25/22 17:44	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	105		62 - 137		05/25/22 17:44	1
4-Bromofluorobenzene (Surr)	97		56 - 136		05/25/22 17:44	1
Toluene-d8 (Surr)	103		78 - 122		05/25/22 17:44	1
Dibromofluoromethane (Surr)	96		73 - 120		05/25/22 17:44	1

Client Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: MW-54S_051322

Lab Sample ID: 240-166736-3

Date Collected: 05/13/22 11:30

Matrix: Water

Date Received: 05/17/22 09:30

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/24/22 02:49	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		66 - 120					05/24/22 02:49	1

Method: 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/25/22 18:29	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/25/22 18:29	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 18:29	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/25/22 18:29	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 18:29	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/25/22 18:29	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	104		62 - 137					05/25/22 18:29	1
4-Bromofluorobenzene (Surr)	97		56 - 136					05/25/22 18:29	1
Toluene-d8 (Surr)	102		78 - 122					05/25/22 18:29	1
Dibromofluoromethane (Surr)	96		73 - 120					05/25/22 18:29	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: MW-54_051322

Lab Sample ID: 240-166736-4

Date Collected: 05/13/22 12:45

Matrix: Water

Date Received: 05/17/22 09:30

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	1.3	J	2.0	0.86	ug/L			05/24/22 03:13	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		66 - 120		05/24/22 03:13	1

Method: 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/25/22 18:52	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/25/22 18:52	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 18:52	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/25/22 18:52	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 18:52	1
Vinyl chloride	0.89	J	1.0	0.45	ug/L			05/25/22 18:52	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	107		62 - 137		05/25/22 18:52	1
4-Bromofluorobenzene (Surr)	97		56 - 136		05/25/22 18:52	1
Toluene-d8 (Surr)	101		78 - 122		05/25/22 18:52	1
Dibromofluoromethane (Surr)	97		73 - 120		05/25/22 18:52	1

Client Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: MW-57_051322

Lab Sample ID: 240-166736-5

Date Collected: 05/13/22 14:35

Matrix: Water

Date Received: 05/17/22 09:30

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	0.98	J	2.0	0.86	ug/L			05/24/22 03:38	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	104		66 - 120					05/24/22 03:38	1

Method: 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/25/22 19:14	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/25/22 19:14	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 19:14	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/25/22 19:14	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 19:14	1
Vinyl chloride	0.77	J	1.0	0.45	ug/L			05/25/22 19:14	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	105		62 - 137					05/25/22 19:14	1
4-Bromofluorobenzene (Surr)	95		56 - 136					05/25/22 19:14	1
Toluene-d8 (Surr)	101		78 - 122					05/25/22 19:14	1
Dibromofluoromethane (Surr)	95		73 - 120					05/25/22 19:14	1

Surrogate Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Method: 8260D - Volatile Organic Compounds by GC/MS

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		DCA (62-137)	BFB (56-136)	TOL (78-122)	DBFM (73-120)
240-166734-B-2 MS	Matrix Spike	99	109	107	94
240-166734-B-2 MSD	Matrix Spike Duplicate	100	107	105	94
240-166736-1	TRIP BLANK_107	105	97	101	96
240-166736-2	MW-67_051322	105	97	103	96
240-166736-3	MW-54S_051322	104	97	102	96
240-166736-4	MW-54_051322	107	97	101	97
240-166736-5	MW-57_051322	105	95	101	95
LCS 240-527876/5	Lab Control Sample	97	106	104	93
MB 240-527876/8	Method Blank	104	99	103	95

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)
BFB = 4-Bromofluorobenzene (Surr)
TOL = Toluene-d8 (Surr)
DBFM = Dibromofluoromethane (Surr)

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		DCA (66-120)
240-166722-B-2 MS	Matrix Spike	102
240-166722-B-2 MSD	Matrix Spike Duplicate	99
240-166736-2	MW-67_051322	105
240-166736-3	MW-54S_051322	101
240-166736-4	MW-54_051322	101
240-166736-5	MW-57_051322	104
LCS 240-527589/3	Lab Control Sample	101
MB 240-527589/6	Method Blank	102

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Method: 8260D - Volatile Organic Compounds by GC/MS

Lab Sample ID: MB 240-527876/8
Matrix: Water
Analysis Batch: 527876

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/25/22 11:45	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/25/22 11:45	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 11:45	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/25/22 11:45	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/25/22 11:45	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/25/22 11:45	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
1,2-Dichloroethane-d4 (Surr)	104		62 - 137		05/25/22 11:45	1
4-Bromofluorobenzene (Surr)	99		56 - 136		05/25/22 11:45	1
Toluene-d8 (Surr)	103		78 - 122		05/25/22 11:45	1
Dibromofluoromethane (Surr)	95		73 - 120		05/25/22 11:45	1

Lab Sample ID: LCS 240-527876/5
Matrix: Water
Analysis Batch: 527876

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
1,1-Dichloroethene	20.0	20.5		ug/L		102	63 - 134
cis-1,2-Dichloroethene	20.0	19.0		ug/L		95	77 - 123
Tetrachloroethene	20.0	21.2		ug/L		106	76 - 123
trans-1,2-Dichloroethene	20.0	20.2		ug/L		101	75 - 124
Trichloroethene	20.0	19.1		ug/L		96	70 - 122
Vinyl chloride	20.0	17.1		ug/L		85	60 - 144

Surrogate	LCS	LCS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	97		62 - 137
4-Bromofluorobenzene (Surr)	106		56 - 136
Toluene-d8 (Surr)	104		78 - 122
Dibromofluoromethane (Surr)	93		73 - 120

Lab Sample ID: 240-166734-B-2 MS
Matrix: Water
Analysis Batch: 527876

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample	Sample	Spike Added	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier		Result	Qualifier				
1,1-Dichloroethene	20	U	400	408		ug/L		102	56 - 135
cis-1,2-Dichloroethene	53		400	430		ug/L		94	66 - 128
Tetrachloroethene	20	U	400	413		ug/L		103	62 - 131
trans-1,2-Dichloroethene	120		400	515		ug/L		98	56 - 136
Trichloroethene	880		400	1190		ug/L		79	61 - 124
Vinyl chloride	20	U	400	332		ug/L		83	43 - 157

Surrogate	MS	MS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	99		62 - 137
4-Bromofluorobenzene (Surr)	109		56 - 136
Toluene-d8 (Surr)	107		78 - 122

QC Sample Results

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Method: 8260D - Volatile Organic Compounds by GC/MS (Continued)

Lab Sample ID: 240-166734-B-2 MS
Matrix: Water
Analysis Batch: 527876

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Surrogate	MS %Recovery	MS Qualifier	Limits
Dibromofluoromethane (Surr)	94		73 - 120

Lab Sample ID: 240-166734-B-2 MSD
Matrix: Water
Analysis Batch: 527876

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1,1-Dichloroethene	20	U	400	405		ug/L		101	56 - 135	1	26
cis-1,2-Dichloroethene	53		400	428		ug/L		94	66 - 128	1	14
Tetrachloroethene	20	U	400	400		ug/L		100	62 - 131	3	20
trans-1,2-Dichloroethene	120		400	511		ug/L		97	56 - 136	1	15
Trichloroethene	880		400	1190		ug/L		78	61 - 124	0	15
Vinyl chloride	20	U	400	336		ug/L		84	43 - 157	1	24

Surrogate	MSD %Recovery	MSD Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	100		62 - 137
4-Bromofluorobenzene (Surr)	107		56 - 136
Toluene-d8 (Surr)	105		78 - 122
Dibromofluoromethane (Surr)	94		73 - 120

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 240-527589/6
Matrix: Water
Analysis Batch: 527589

Client Sample ID: Method Blank
Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/23/22 21:00	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	102		66 - 120		05/23/22 21:00	1

Lab Sample ID: LCS 240-527589/3
Matrix: Water
Analysis Batch: 527589

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1,4-Dioxane	10.0	9.54		ug/L		95	80 - 122

Surrogate	LCS %Recovery	LCS Qualifier	Limits
1,2-Dichloroethane-d4 (Surr)	101		66 - 120

Lab Sample ID: 240-166722-B-2 MS
Matrix: Water
Analysis Batch: 527589

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1,4-Dioxane	2.0	U	10.0	9.87		ug/L		99	51 - 153

Eurofins Canton

QC Sample Results

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Method: 8260D SIM - Volatile Organic Compounds (GC/MS) (Continued)

<i>Surrogate</i>	<i>MS</i> <i>%Recovery</i>	<i>MS</i> <i>Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	102		66 - 120

Lab Sample ID: 240-166722-B-2 MSD
Matrix: Water
Analysis Batch: 527589

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

<i>Analyte</i>	<i>Sample</i> <i>Result</i>	<i>Sample</i> <i>Qualifier</i>	<i>Spike</i> <i>Added</i>	<i>MSD</i> <i>Result</i>	<i>MSD</i> <i>Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i> <i>Limits</i>	<i>RPD</i>	<i>RPD</i> <i>Limit</i>
1,4-Dioxane	2.0	U	10.0	9.81		ug/L		98	51 - 153	1	16

<i>Surrogate</i>	<i>MSD</i> <i>%Recovery</i>	<i>MSD</i> <i>Qualifier</i>	<i>Limits</i>
1,2-Dichloroethane-d4 (Surr)	99		66 - 120

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

QC Association Summary

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

GC/MS VOA

Analysis Batch: 527589

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-166736-2	MW-67_051322	Total/NA	Water	8260D SIM	
240-166736-3	MW-54S_051322	Total/NA	Water	8260D SIM	
240-166736-4	MW-54_051322	Total/NA	Water	8260D SIM	
240-166736-5	MW-57_051322	Total/NA	Water	8260D SIM	
MB 240-527589/6	Method Blank	Total/NA	Water	8260D SIM	
LCS 240-527589/3	Lab Control Sample	Total/NA	Water	8260D SIM	
240-166722-B-2 MS	Matrix Spike	Total/NA	Water	8260D SIM	
240-166722-B-2 MSD	Matrix Spike Duplicate	Total/NA	Water	8260D SIM	

Analysis Batch: 527876

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-166736-1	TRIP BLANK_107	Total/NA	Water	8260D	
240-166736-2	MW-67_051322	Total/NA	Water	8260D	
240-166736-3	MW-54S_051322	Total/NA	Water	8260D	
240-166736-4	MW-54_051322	Total/NA	Water	8260D	
240-166736-5	MW-57_051322	Total/NA	Water	8260D	
MB 240-527876/8	Method Blank	Total/NA	Water	8260D	
LCS 240-527876/5	Lab Control Sample	Total/NA	Water	8260D	
240-166734-B-2 MS	Matrix Spike	Total/NA	Water	8260D	
240-166734-B-2 MSD	Matrix Spike Duplicate	Total/NA	Water	8260D	

Lab Chronicle

Client: ARCADIS U.S., Inc.
Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Client Sample ID: TRIP BLANK_107

Lab Sample ID: 240-166736-1

Date Collected: 05/13/22 00:00

Matrix: Water

Date Received: 05/17/22 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260D		1	527876	05/25/22 17:22	TJL1	TAL CAN

Client Sample ID: MW-67_051322

Lab Sample ID: 240-166736-2

Date Collected: 05/13/22 10:20

Matrix: Water

Date Received: 05/17/22 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260D		1	527876	05/25/22 17:44	TJL1	TAL CAN
Total/NA	Analysis	8260D SIM		1	527589	05/24/22 02:24	CS	TAL CAN

Client Sample ID: MW-54S_051322

Lab Sample ID: 240-166736-3

Date Collected: 05/13/22 11:30

Matrix: Water

Date Received: 05/17/22 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260D		1	527876	05/25/22 18:29	TJL1	TAL CAN
Total/NA	Analysis	8260D SIM		1	527589	05/24/22 02:49	CS	TAL CAN

Client Sample ID: MW-54_051322

Lab Sample ID: 240-166736-4

Date Collected: 05/13/22 12:45

Matrix: Water

Date Received: 05/17/22 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260D		1	527876	05/25/22 18:52	TJL1	TAL CAN
Total/NA	Analysis	8260D SIM		1	527589	05/24/22 03:13	CS	TAL CAN

Client Sample ID: MW-57_051322

Lab Sample ID: 240-166736-5

Date Collected: 05/13/22 14:35

Matrix: Water

Date Received: 05/17/22 09:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	8260D		1	527876	05/25/22 19:14	TJL1	TAL CAN
Total/NA	Analysis	8260D SIM		1	527589	05/24/22 03:38	CS	TAL CAN

Laboratory References:

TAL CAN = Eurofins Canton, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.
 Project/Site: Ford LTP - On Site

Job ID: 240-166736-1

Laboratory: Eurofins Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	2927	02-27-23
Connecticut	State	PH-0590	12-31-23
Florida	NELAP	E87225	06-30-22
Georgia	State	4062	02-23-22 *
Illinois	NELAP	200004	07-31-22
Iowa	State	421	06-01-23
Kentucky (UST)	State	112225	02-27-23
Kentucky (WW)	State	KY98016	12-31-22
Minnesota	NELAP	039-999-348	12-31-22
Minnesota (Petrofund)	State	3506	08-01-23
New Jersey	NELAP	OH001	06-30-22
New York	NELAP	10975	04-01-23
Ohio	State	8303	02-23-23
Ohio VAP	State	CL0024	02-27-23
Oregon	NELAP	4062	02-27-23
Pennsylvania	NELAP	68-00340	08-31-22
Texas	NELAP	T104704517-22-16	08-31-22
Virginia	NELAP	11570	09-14-22
Washington	State	C971	01-12-23
West Virginia DEP	State	210	12-31-22

* Accreditation/Certification renewal pending - accreditation/certification considered valid.



Client Contact Company Name: Arcadis Address: 28550 Cabot Drive, Suite 500 City/State/Zip: Novi, MI, 48377 Phone: 248-994-2240 Project Name: Ford LTP On-Site Project Number: 30080642-401.03 PO # 30080642-401.03		Regulatory program: <input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other	
Client Project Manager: Kris Hinskey Telephone: 269-832-7478 Email: Kristoffer.Hinskey@arcadis.com Sampler Name: <u>Sommer Guy</u> Method of Shipment/Carrier: Shipping/Tracking No:		Site Contact: Christina Weaver Telephone: 248-994-2329 Analysis Turnaround Time: TAT if different from below: <input type="checkbox"/> 3 weeks <input checked="" type="checkbox"/> 2 weeks <input type="checkbox"/> 1 week <input type="checkbox"/> 2 days <input type="checkbox"/> 1 day	
Client Project Manager: Mike DeMonico Telephone: 330-966-9783		Lab Contact: Mike DeMonico Telephone: 330-966-9783	
TestAmerica Laboratories, Inc. COC No: 1 of 1 COCs		For lab use only Walk-in client Lab sampling Job/SDG No:	
Sample Identification		Sample Specific Notes / Special Instructions:	
TRIP BLANK_107	5/13/22	1	1 Trip Blank
MW-67-051322	5/13/22 1020	6	3 VOAs for 8260D 3 VOAs for 8260D SIM
MW-54S-051322	5/13/22 1130	6	"
MW-54-051322	5/13/22 1245	6	"
MW-57-051322	5/13/22 1435	6	"
S6 2/13			
Possible Hazard Identification <input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown		Sample Disposal: <input type="checkbox"/> A fee may be assessed <input type="checkbox"/> Return to Client <input checked="" type="checkbox"/> Disposal by:	
Special Instructions/QC Requirements & Comments: Submit all results through Cadena at jtomalia@cadenaco.com, Cadena #E203728 Level IV Reporting requested.			
Relinquished by: <u>Sommer Guy</u> Relinquished by: <u>[Signature]</u> Relinquished by: <u>[Signature]</u>		Received by: <u>Novi Cold Storage</u> Received by: <u>[Signature]</u> Received in Laboratory by: <u>[Signature]</u>	
Company: Arcadis Date/Time: 5/13/22 15:30		Company: Arcadis Date/Time: 5/13/22 15:30	
Company: Arcadis Date/Time: 5/16/22 1200		Company: Arcadis Date/Time: 5/16/22 1200	
Company: FETA Date/Time: 5/16/22 1250		Company: EETNL Date/Time: 5-17-22 0930	

Eurofins TestAmerica Canton Sample Receipt Form/Narrative
 Canton Facility _____ Login # : 166736

Client Accadis Site Name Ford LTP Cooler unpacked by: JMP

Cooler Received on 5-17-22 Opened on 5-17-22

FedEx: 1st Grd. (Exp) UPS FAS Clipper Client Drop Off TestAmerica Courier Other _____

Receipt After-hours: Drop-off Date/Time _____ Storage Location _____

TestAmerica Cooler # TA Foam Box Client Cooler Box Other _____
 Packing material used: Bubble Wrap Foam Plastic Bag None Other _____
 COOLANT: Wet Ice Blue Ice Dry Ice Water None

1. Cooler temperature upon receipt See Multiple Cooler Form
 IR GUN# IR-13 (CF 0.0 °C) Observed Cooler Temp. _____ °C Corrected Cooler Temp. _____ °C
 IR GUN #IR-15 (CF -0.7°C) Observed Cooler Temp. _____ °C Corrected Cooler Temp. _____ °C
2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 1 ea Yes No
 -Were the seals on the outside of the cooler(s) signed & dated? Yes No NA
 -Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No NA
 -Were tamper/custody seals intact and uncompromised? Yes No NA
3. Shippers' packing slip attached to the cooler(s)? Yes No
4. Did custody papers accompany the sample(s)? Yes No
5. Were the custody papers relinquished & signed in the appropriate place? Yes No
6. Was/were the person(s) who collected the samples clearly identified on the COC? Yes No
7. Did all bottles arrive in good condition (Unbroken)? Yes No
8. Could all bottle labels (ID/Date/Time) be reconciled with the COC? Yes No
9. For each sample, does the COC specify preservatives (Y/N), # of containers (Y/N), and sample type of grab/comp (Y/N)?
10. Were correct bottle(s) used for the test(s) indicated? Yes No
11. Sufficient quantity received to perform indicated analyses? Yes No
12. Are these work share samples and all listed on the COC? Yes No
 If yes, Questions 13-17 have been checked at the originating laboratory.
13. Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC157842
14. Were VOAs on the COC? Yes No
15. Were air bubbles >6 mm in any VOA vials? Yes No NA Larger than this.
16. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # covered Yes No
17. Was a LL Hg or Me Hg trip blank present? Yes No

Tests that are not checked for pH by Receiving:
 VOAs
 Oil and Grease
 TOC

Contacted PM _____ Date _____ by _____ via Verbal Voice Mail Other _____
 Concerning _____

18. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES additional next page Samples processed by: _____

19. SAMPLE CONDITION
 Sample(s) _____ were received after the recommended holding time had expired.
 Sample(s) _____ were received in a broken container.
 Sample(s) 3x40-MW-67, 1x40-MW-545 were received with bubble >6 mm in diameter. (Notify PM)

20. SAMPLE PRESERVATION
 Sample(s) _____ were further preserved in the laboratory.
 Time preserved: _____ Preservative(s) added/Lot number(s): _____
 VOA Sample Preservation - Date/Time VOAs Frozen: _____

Login #: 1166736

Eurofins - Canton Sample Receipt Multiple Cooler Form									
Cooler Description (Circle)				IR Gun # (Circle)	Observed Temp °C	Corrected Temp °C	Coolant (Circle)		
TA	Client	Box	Other	IR-13 IR-15	2.1	2.1	Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15	0.6	0.6	Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	
TA	Client	Box	Other	IR-13 IR-15			Wet Ice	Blue Ice	Dry Ice
							Water	None	

See Temperature Excursion Form