

 **ANALYTICAL REPORT****PREPARED FOR**

Attn: Kristoffer Hinskey  
ARCADIS U.S., Inc.  
28550 Cabot Drive  
Suite 500  
Novi, Michigan 48377

Generated 12/2/2022 8:28:04 AM

**JOB DESCRIPTION**

Ford LTP - Off Site

**JOB NUMBER**

240-176530-1

# Eurofins Canton

## Job Notes

The test results in this report meet all NELAP requirements for parameters for which accreditation is required or available. Any exceptions to the NELAP requirements are noted in this report. Pursuant to NELAP, this report may not be reproduced, except in full, without the written approval of the laboratory. This report is confidential and is intended for the sole use of Eurofins Environment Testing North Central, LLC and its client. All questions regarding this report should be directed to the Eurofins Environment Testing North Central, LLC Project Manager who has signed this report.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Environment Testing North Central, LLC Project Manager.

## Authorization



Generated  
12/2/2022 8:28:04 AM

Authorized for release by  
Opal Johnson, Project Manager II  
[Opal.Johnson@et.eurofinsus.com](mailto:Opal.Johnson@et.eurofinsus.com)  
Designee for  
Michael DeMonico, Project Manager I  
[Michael.DeMonico@et.eurofinsus.com](mailto:Michael.DeMonico@et.eurofinsus.com)  
(330)497-9396



# Table of Contents

|                                  |    |
|----------------------------------|----|
| Cover Page . . . . .             | 1  |
| Table of Contents . . . . .      | 3  |
| Definitions/Glossary . . . . .   | 4  |
| Case Narrative . . . . .         | 5  |
| Method Summary . . . . .         | 6  |
| Sample Summary . . . . .         | 7  |
| Detection Summary . . . . .      | 8  |
| Client Sample Results . . . . .  | 9  |
| Surrogate Summary . . . . .      | 14 |
| QC Sample Results . . . . .      | 15 |
| QC Association Summary . . . . . | 19 |
| Lab Chronicle . . . . .          | 20 |
| Certification Summary . . . . .  | 21 |
| Chain of Custody . . . . .       | 22 |

# Definitions/Glossary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Qualifiers

### GC/MS VOA

| Qualifier | Qualifier Description                                    |
|-----------|--|
| F1        | MS and/or MSD recovery exceeds control limits.           |
| U         | Indicates the analyte was analyzed for but not detected. |

## Glossary

| Abbreviation   | These commonly used abbreviations may or may not be present in this report.                                 |
|----------------|---|
| α              | Listed under the "D" column to designate that the result is reported on a dry weight basis                  |
| %R             | Percent Recovery  |
| CFL            | Contains Free Liquid  |
| CFU            | Colony Forming Unit   |
| CNF            | Contains No Free Liquid   |
| DER            | Duplicate Error Ratio (normalized absolute difference)  |
| Dil Fac        | Dilution Factor   |
| DL             | Detection Limit (DoD/DOE)   |
| DL, RA, RE, IN | Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample |
| DLC            | Decision Level Concentration (Radiochemistry)   |
| EDL            | Estimated Detection Limit (Dioxin)  |
| LOD            | Limit of Detection (DoD/DOE)  |
| LOQ            | Limit of Quantitation (DoD/DOE)   |
| MCL            | EPA recommended "Maximum Contaminant Level"   |
| MDA            | Minimum Detectable Activity (Radiochemistry)  |
| MDC            | Minimum Detectable Concentration (Radiochemistry)   |
| MDL            | Method Detection Limit  |
| ML             | Minimum Level (Dioxin)  |
| MPN            | Most Probable Number  |
| MQL            | Method Quantitation Limit   |
| NC             | Not Calculated  |
| ND             | Not Detected at the reporting limit (or MDL or EDL if shown)  |
| NEG            | Negative / Absent   |
| POS            | Positive / Present  |
| PQL            | Practical Quantitation Limit  |
| PRES           | Presumptive   |
| QC             | Quality Control   |
| RER            | Relative Error Ratio (Radiochemistry)   |
| RL             | Reporting Limit or Requested Limit (Radiochemistry)   |
| RPD            | Relative Percent Difference, a measure of the relative difference between two points                        |
| TEF            | Toxicity Equivalent Factor (Dioxin)   |
| TEQ            | Toxicity Equivalent Quotient (Dioxin)   |
| TNTC           | Too Numerous To Count   |

# Case Narrative

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

---

**Job ID: 240-176530-1**

---

**Laboratory: Eurofins Canton**

---

**Narrative**

**Job Narrative  
240-176530-1**

**Comments**

No additional comments.

**Receipt**

The samples were received on 11/16/2022 8:00 AM. Unless otherwise noted below, the samples arrived in good condition, and where required, properly preserved and on ice. The temperature of the cooler at receipt was 3.9° C.

**GC/MS VOA**

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

**VOA Prep**

No analytical or quality issues were noted, other than those described in the Definitions/Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

# Method Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

| Method    | Method Description                  | Protocol | Laboratory |
|-----------|-------------------------------------|----------|------------|
| 8260D     | Volatile Organic Compounds by GC/MS | SW846    | EET CAN    |
| 8260D SIM | Volatile Organic Compounds (GC/MS)  | SW846    | EET CAN    |
| 5030C     | Purge and Trap                      | SW846    | EET CAN    |

**Protocol References:**

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

**Laboratory References:**

EET CAN = Eurofins Canton, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396



# Sample Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

| Lab Sample ID | Client Sample ID | Matrix | Collected      | Received       |
|---------------|------------------|--------|----------------|----------------|
| 240-176530-1  | TRIP BLANK_10    | Water  | 11/14/22 00:00 | 11/16/22 08:00 |
| 240-176530-2  | MW-228S_111422   | Water  | 11/14/22 10:10 | 11/16/22 08:00 |
| 240-176530-3  | MW-229S_111422   | Water  | 11/14/22 11:20 | 11/16/22 08:00 |
| 240-176530-4  | MW-229_111422    | Water  | 11/14/22 12:15 | 11/16/22 08:00 |
| 240-176530-5  | MW-229D_111422   | Water  | 11/14/22 13:15 | 11/16/22 08:00 |

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

# Detection Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: TRIP BLANK\_10** **Lab Sample ID: 240-176530-1**

No Detections.

**Client Sample ID: MW-228S\_111422** **Lab Sample ID: 240-176530-2**

No Detections.

**Client Sample ID: MW-229S\_111422** **Lab Sample ID: 240-176530-3**

No Detections.

**Client Sample ID: MW-229\_111422** **Lab Sample ID: 240-176530-4**

No Detections.

**Client Sample ID: MW-229D\_111422** **Lab Sample ID: 240-176530-5**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Canton





# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: TRIP BLANK\_10**

**Lab Sample ID: 240-176530-1**

**Date Collected: 11/14/22 00:00**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 14:40 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 14:40 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 14:40 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 14:40 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 14:40 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 14:40 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 106       |           | 62 - 137 |          | 11/23/22 14:40 | 1       |
| 4-Bromofluorobenzene (Surr)  | 88        |           | 56 - 136 |          | 11/23/22 14:40 | 1       |
| Toluene-d8 (Surr)            | 106       |           | 78 - 122 |          | 11/23/22 14:40 | 1       |
| Dibromofluoromethane (Surr)  | 99        |           | 73 - 120 |          | 11/23/22 14:40 | 1       |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-228S\_111422**

**Lab Sample ID: 240-176530-2**

**Date Collected: 11/14/22 10:10**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte     | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/23/22 03:11 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 79        |           | 66 - 120 |          | 11/23/22 03:11 | 1       |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 15:04 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 15:04 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:04 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 15:04 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:04 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 15:04 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 106       |           | 62 - 137 |          | 11/23/22 15:04 | 1       |
| 4-Bromofluorobenzene (Surr)  | 91        |           | 56 - 136 |          | 11/23/22 15:04 | 1       |
| Toluene-d8 (Surr)            | 109       |           | 78 - 122 |          | 11/23/22 15:04 | 1       |
| Dibromofluoromethane (Surr)  | 97        |           | 73 - 120 |          | 11/23/22 15:04 | 1       |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-229S\_111422**

**Lab Sample ID: 240-176530-3**

**Date Collected: 11/14/22 11:20**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte                      | Result           | Qualifier        | RL            | MDL  | Unit | D | Prepared        | Analyzed        | Dil Fac        |
|------------------------------|------------------|------------------|---------------|------|------|---|-----------------|-----------------|----------------|
| 1,4-Dioxane                  | 2.0              | U                | 2.0           | 0.86 | ug/L |   |                 | 11/27/22 20:58  | 1              |
| <b>Surrogate</b>             | <b>%Recovery</b> | <b>Qualifier</b> | <b>Limits</b> |      |      |   | <b>Prepared</b> | <b>Analyzed</b> | <b>Dil Fac</b> |
| 1,2-Dichloroethane-d4 (Surr) | 77               |                  | 66 - 120      |      |      |   |                 | 11/27/22 20:58  | 1              |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                      | Result           | Qualifier        | RL            | MDL  | Unit | D | Prepared        | Analyzed        | Dil Fac        |
|------------------------------|------------------|------------------|---------------|------|------|---|-----------------|-----------------|----------------|
| 1,1-Dichloroethene           | 1.0              | U                | 1.0           | 0.49 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| cis-1,2-Dichloroethene       | 1.0              | U                | 1.0           | 0.46 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| Tetrachloroethene            | 1.0              | U                | 1.0           | 0.44 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| trans-1,2-Dichloroethene     | 1.0              | U                | 1.0           | 0.51 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| Trichloroethene              | 1.0              | U                | 1.0           | 0.44 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| Vinyl chloride               | 1.0              | U                | 1.0           | 0.45 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| <b>Surrogate</b>             | <b>%Recovery</b> | <b>Qualifier</b> | <b>Limits</b> |      |      |   | <b>Prepared</b> | <b>Analyzed</b> | <b>Dil Fac</b> |
| 1,2-Dichloroethane-d4 (Surr) | 106              |                  | 62 - 137      |      |      |   |                 | 11/23/22 15:27  | 1              |
| 4-Bromofluorobenzene (Surr)  | 89               |                  | 56 - 136      |      |      |   |                 | 11/23/22 15:27  | 1              |
| Toluene-d8 (Surr)            | 104              |                  | 78 - 122      |      |      |   |                 | 11/23/22 15:27  | 1              |
| Dibromofluoromethane (Surr)  | 98               |                  | 73 - 120      |      |      |   |                 | 11/23/22 15:27  | 1              |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-229\_111422**

**Lab Sample ID: 240-176530-4**

**Date Collected: 11/14/22 12:15**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte     | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/27/22 21:24 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 77        |           | 66 - 120 |          | 11/27/22 21:24 | 1       |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 15:50 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 15:50 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:50 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 15:50 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:50 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 15:50 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 106       |           | 62 - 137 |          | 11/23/22 15:50 | 1       |
| 4-Bromofluorobenzene (Surr)  | 90        |           | 56 - 136 |          | 11/23/22 15:50 | 1       |
| Toluene-d8 (Surr)            | 102       |           | 78 - 122 |          | 11/23/22 15:50 | 1       |
| Dibromofluoromethane (Surr)  | 99        |           | 73 - 120 |          | 11/23/22 15:50 | 1       |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-229D\_111422**

**Lab Sample ID: 240-176530-5**

**Date Collected: 11/14/22 13:15**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte     | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/27/22 21:49 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 78        |           | 66 - 120 |          | 11/27/22 21:49 | 1       |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 16:13 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 16:13 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 16:13 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 16:13 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 16:13 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 16:13 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 104       |           | 62 - 137 |          | 11/23/22 16:13 | 1       |
| 4-Bromofluorobenzene (Surr)  | 87        |           | 56 - 136 |          | 11/23/22 16:13 | 1       |
| Toluene-d8 (Surr)            | 105       |           | 78 - 122 |          | 11/23/22 16:13 | 1       |
| Dibromofluoromethane (Surr)  | 96        |           | 73 - 120 |          | 11/23/22 16:13 | 1       |

# Surrogate Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Method: 8260D - Volatile Organic Compounds by GC/MS

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

| Lab Sample ID      | Client Sample ID       | Percent Surrogate Recovery (Acceptance Limits) |                 |                 |                  |
|--------------------|------------------------|--|-----------------|-----------------|------------------|
|                    |                        | DCA<br>(62-137)                                | BFB<br>(56-136) | TOL<br>(78-122) | DBFM<br>(73-120) |
| 240-176530-1       | TRIP BLANK_10          | 106  | 88              | 106             | 99               |
| 240-176530-2       | MW-228S_111422         | 106  | 91              | 109             | 97               |
| 240-176530-3       | MW-229S_111422         | 106  | 89              | 104             | 98               |
| 240-176530-4       | MW-229_111422          | 106  | 90              | 102             | 99               |
| 240-176530-5       | MW-229D_111422         | 104  | 87              | 105             | 96               |
| 240-176531-B-5 MS  | Matrix Spike           | 101  | 98              | 110             | 94               |
| 240-176531-B-5 MSD | Matrix Spike Duplicate | 100  | 107             | 104             | 95               |
| LCS 240-553308/5   | Lab Control Sample     | 100  | 103             | 112             | 93               |
| MB 240-553308/8    | Method Blank           | 103  | 89              | 103             | 95               |

#### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)  
BFB = 4-Bromofluorobenzene (Surr)  
TOL = Toluene-d8 (Surr)  
DBFM = Dibromofluoromethane (Surr)

## Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

| Lab Sample ID      | Client Sample ID       | DCA      |
|--------------------|------------------------|----------|
|                    |                        | (66-120) |
| 240-176530-2       | MW-228S_111422         | 79       |
| 240-176530-2 MS    | MW-228S_111422         | 79       |
| 240-176530-2 MSD   | MW-228S_111422         | 81       |
| 240-176530-3       | MW-229S_111422         | 77       |
| 240-176530-4       | MW-229_111422          | 77       |
| 240-176530-5       | MW-229D_111422         | 78       |
| 240-176634-I-5 MS  | Matrix Spike           | 80       |
| 240-176634-O-5 MSD | Matrix Spike Duplicate | 80       |
| LCS 240-553220/3   | Lab Control Sample     | 79       |
| LCS 240-553480/3   | Lab Control Sample     | 76       |
| MB 240-553220/5    | Method Blank           | 77       |
| MB 240-553480/4    | Method Blank           | 76       |

#### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

# QC Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Method: 8260D - Volatile Organic Compounds by GC/MS

**Lab Sample ID: MB 240-553308/8**  
**Matrix: Water**  
**Analysis Batch: 553308**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

| Analyte                  | MB     | MB        | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
|                          | Result | Qualifier |     |      |      |   |          |                |         |
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 11:34 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 11:34 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 11:34 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 11:34 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 11:34 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 11:34 | 1       |

| Surrogate                    | MB        | MB        | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
|                              | %Recovery | Qualifier |          |          |                |         |
| 1,2-Dichloroethane-d4 (Surr) | 103       |           | 62 - 137 |          | 11/23/22 11:34 | 1       |
| 4-Bromofluorobenzene (Surr)  | 89        |           | 56 - 136 |          | 11/23/22 11:34 | 1       |
| Toluene-d8 (Surr)            | 103       |           | 78 - 122 |          | 11/23/22 11:34 | 1       |
| Dibromofluoromethane (Surr)  | 95        |           | 73 - 120 |          | 11/23/22 11:34 | 1       |

**Lab Sample ID: LCS 240-553308/5**  
**Matrix: Water**  
**Analysis Batch: 553308**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

| Analyte                  | Spike Added | LCS    | LCS       | Unit | D | %Rec | %Rec Limits |
|--------------------------|-------------|--------|-----------|------|---|------|-------------|
|                          |             | Result | Qualifier |      |   |      |             |
| 1,1-Dichloroethene       | 20.0        | 16.6   |           | ug/L |   | 83   | 63 - 134    |
| cis-1,2-Dichloroethene   | 20.0        | 16.3   |           | ug/L |   | 81   | 77 - 123    |
| Tetrachloroethene        | 20.0        | 19.6   |           | ug/L |   | 98   | 76 - 123    |
| trans-1,2-Dichloroethene | 20.0        | 16.4   |           | ug/L |   | 82   | 75 - 124    |
| Trichloroethene          | 20.0        | 16.7   |           | ug/L |   | 84   | 70 - 122    |
| Vinyl chloride           | 20.0        | 16.2   |           | ug/L |   | 81   | 60 - 144    |

| Surrogate                    | LCS       | LCS       | Limits   |
|------------------------------|-----------|-----------|----------|
|                              | %Recovery | Qualifier |          |
| 1,2-Dichloroethane-d4 (Surr) | 100       |           | 62 - 137 |
| 4-Bromofluorobenzene (Surr)  | 103       |           | 56 - 136 |
| Toluene-d8 (Surr)            | 112       |           | 78 - 122 |
| Dibromofluoromethane (Surr)  | 93        |           | 73 - 120 |

**Lab Sample ID: 240-176531-B-5 MS**  
**Matrix: Water**  
**Analysis Batch: 553308**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

| Analyte                  | Sample | Sample    | Spike Added | MS     | MS        | Unit | D | %Rec | %Rec Limits |
|--------------------------|--------|-----------|-------------|--------|-----------|------|---|------|-------------|
|                          | Result | Qualifier |             | Result | Qualifier |      |   |      |             |
| 1,1-Dichloroethene       | 20     | U         | 400         | 298    |           | ug/L |   | 75   | 56 - 135    |
| cis-1,2-Dichloroethene   | 73     |           | 400         | 378    |           | ug/L |   | 76   | 66 - 128    |
| Tetrachloroethene        | 20     | U         | 400         | 337    |           | ug/L |   | 84   | 62 - 131    |
| trans-1,2-Dichloroethene | 20     | U         | 400         | 309    |           | ug/L |   | 77   | 56 - 136    |
| Trichloroethene          | 20     | U         | 400         | 299    |           | ug/L |   | 75   | 61 - 124    |
| Vinyl chloride           | 930    | F1        | 400         | 1140   |           | ug/L |   | 51   | 43 - 157    |

| Surrogate                    | MS        | MS        | Limits   |
|------------------------------|-----------|-----------|----------|
|                              | %Recovery | Qualifier |          |
| 1,2-Dichloroethane-d4 (Surr) | 101       |           | 62 - 137 |
| 4-Bromofluorobenzene (Surr)  | 98        |           | 56 - 136 |
| Toluene-d8 (Surr)            | 110       |           | 78 - 122 |

# QC Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Method: 8260D - Volatile Organic Compounds by GC/MS (Continued)

**Lab Sample ID: 240-176531-B-5 MS**  
**Matrix: Water**  
**Analysis Batch: 553308**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

|                             | MS        | MS        |          |
|-----------------------------|-----------|-----------|----------|
| Surrogate                   | %Recovery | Qualifier | Limits   |
| Dibromofluoromethane (Surr) | 94        |           | 73 - 120 |

**Lab Sample ID: 240-176531-B-5 MSD**  
**Matrix: Water**  
**Analysis Batch: 553308**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

| Analyte                  | Sample | Sample    | Spike | MSD    | MSD       | Unit | D | %Rec | %Rec     | RPD | RPD | Limit |
|--------------------------|--------|-----------|-------|--------|-----------|------|---|------|----------|-----|-----|-------|
|                          | Result | Qualifier |       | Result | Qualifier |      |   |      |          |     |     |       |
| 1,1-Dichloroethene       | 20     | U         | 400   | 304    |           | ug/L |   | 76   | 56 - 135 | 2   |     | 26    |
| cis-1,2-Dichloroethene   | 73     |           | 400   | 387    |           | ug/L |   | 78   | 66 - 128 | 2   |     | 14    |
| Tetrachloroethene        | 20     | U         | 400   | 352    |           | ug/L |   | 88   | 62 - 131 | 4   |     | 20    |
| trans-1,2-Dichloroethene | 20     | U         | 400   | 317    |           | ug/L |   | 79   | 56 - 136 | 2   |     | 15    |
| Trichloroethene          | 20     | U         | 400   | 316    |           | ug/L |   | 79   | 61 - 124 | 6   |     | 15    |
| Vinyl chloride           | 930    | F1        | 400   | 1090   | F1        | ug/L |   | 38   | 43 - 157 | 5   |     | 24    |

  

|                              | MSD       | MSD       |          |
|------------------------------|-----------|-----------|----------|
| Surrogate                    | %Recovery | Qualifier | Limits   |
| 1,2-Dichloroethane-d4 (Surr) | 100       |           | 62 - 137 |
| 4-Bromofluorobenzene (Surr)  | 107       |           | 56 - 136 |
| Toluene-d8 (Surr)            | 104       |           | 78 - 122 |
| Dibromofluoromethane (Surr)  | 95        |           | 73 - 120 |

## Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 240-553220/5**  
**Matrix: Water**  
**Analysis Batch: 553220**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

| Analyte     | MB     | MB        | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
|             | Result | Qualifier |     |      |      |   |          |                |         |
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/22/22 18:19 | 1       |

  

|                              | MB        | MB        |          |          |                |         |  |  |  |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|--|--|--|
| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |  |  |  |
| 1,2-Dichloroethane-d4 (Surr) | 77        |           | 66 - 120 |          | 11/22/22 18:19 | 1       |  |  |  |

**Lab Sample ID: LCS 240-553220/3**  
**Matrix: Water**  
**Analysis Batch: 553220**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

| Analyte     | Spike | LCS    | LCS       | Unit | D | %Rec | %Rec     | Limits |
|-------------|-------|--------|-----------|------|---|------|----------|--------|
|             |       | Result | Qualifier |      |   |      |          |        |
| 1,4-Dioxane | 10.0  | 9.37   |           | ug/L |   | 94   | 80 - 122 |        |

  

|                              | LCS       | LCS       |          |
|------------------------------|-----------|-----------|----------|
| Surrogate                    | %Recovery | Qualifier | Limits   |
| 1,2-Dichloroethane-d4 (Surr) | 79        |           | 66 - 120 |

**Lab Sample ID: 240-176530-2 MS**  
**Matrix: Water**  
**Analysis Batch: 553220**

**Client Sample ID: MW-228S\_111422**  
**Prep Type: Total/NA**

| Analyte     | Sample | Sample    | Spike | MS     | MS        | Unit | D | %Rec | %Rec     | Limits |
|-------------|--------|-----------|-------|--------|-----------|------|---|------|----------|--------|
|             | Result | Qualifier |       | Result | Qualifier |      |   |      |          |        |
| 1,4-Dioxane | 2.0    | U         | 10.0  | 10.0   |           | ug/L |   | 100  | 51 - 153 |        |

Eurofins Canton



# QC Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Method: 8260D SIM - Volatile Organic Compounds (GC/MS) (Continued)

|                              | MS        | MS        |          |
|------------------------------|-----------|-----------|----------|
| Surrogate                    | %Recovery | Qualifier | Limits   |
| 1,2-Dichloroethane-d4 (Surr) | 79        |           | 66 - 120 |

**Lab Sample ID: 240-176530-2 MSD**  
**Matrix: Water**  
**Analysis Batch: 553220**

**Client Sample ID: MW-228S\_111422**  
**Prep Type: Total/NA**

| Analyte     | Sample Result | Sample Qualifier | Spike Added | MSD Result | MSD Qualifier | Unit | D | %Rec | %Rec Limits | RPD | RPD Limit |
|-------------|---------------|------------------|-------------|------------|---------------|------|---|------|-------------|-----|-----------|
| 1,4-Dioxane | 2.0           | U                | 10.0        | 9.85       |               | ug/L |   | 98   | 51 - 153    | 2   | 16        |

|                              | MSD       | MSD       |          |
|------------------------------|-----------|-----------|----------|
| Surrogate                    | %Recovery | Qualifier | Limits   |
| 1,2-Dichloroethane-d4 (Surr) | 81        |           | 66 - 120 |

**Lab Sample ID: MB 240-553480/4**  
**Matrix: Water**  
**Analysis Batch: 553480**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

| Analyte     | MB Result | MB Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|-----------|--------------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0       | U            | 2.0 | 0.86 | ug/L |   |          | 11/27/22 19:42 | 1       |

|                              | MB        | MB        |          |          |                |         |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
| 1,2-Dichloroethane-d4 (Surr) | 76        |           | 66 - 120 |          | 11/27/22 19:42 | 1       |

**Lab Sample ID: LCS 240-553480/3**  
**Matrix: Water**  
**Analysis Batch: 553480**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

| Analyte     | Spike Added | LCS Result | LCS Qualifier | Unit | D | %Rec | %Rec Limits |
|-------------|-------------|------------|---------------|------|---|------|-------------|
| 1,4-Dioxane | 10.0        | 8.98       |               | ug/L |   | 90   | 80 - 122    |

|                              | LCS       | LCS       |          |
|------------------------------|-----------|-----------|----------|
| Surrogate                    | %Recovery | Qualifier | Limits   |
| 1,2-Dichloroethane-d4 (Surr) | 76        |           | 66 - 120 |

**Lab Sample ID: 240-176634-I-5 MS**  
**Matrix: Water**  
**Analysis Batch: 553480**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**

| Analyte     | Sample Result | Sample Qualifier | Spike Added | MS Result | MS Qualifier | Unit | D | %Rec | %Rec Limits |
|-------------|---------------|------------------|-------------|-----------|--------------|------|---|------|-------------|
| 1,4-Dioxane | 2.0           | U                | 10.0        | 10.2      |              | ug/L |   | 102  | 51 - 153    |

|                              | MS        | MS        |          |
|------------------------------|-----------|-----------|----------|
| Surrogate                    | %Recovery | Qualifier | Limits   |
| 1,2-Dichloroethane-d4 (Surr) | 80        |           | 66 - 120 |

**Lab Sample ID: 240-176634-O-5 MSD**  
**Matrix: Water**  
**Analysis Batch: 553480**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**

| Analyte     | Sample Result | Sample Qualifier | Spike Added | MSD Result | MSD Qualifier | Unit | D | %Rec | %Rec Limits | RPD | RPD Limit |
|-------------|---------------|------------------|-------------|------------|---------------|------|---|------|-------------|-----|-----------|
| 1,4-Dioxane | 2.0           | U                | 10.0        | 10.4       |               | ug/L |   | 104  | 51 - 153    | 2   | 16        |

# QC Sample Results

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Method: 8260D SIM - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 240-176634-O-5 MSD  
Matrix: Water  
Analysis Batch: 553480

Client Sample ID: Matrix Spike Duplicate  
Prep Type: Total/NA

| <u>Surrogate</u>             | <u>MSD</u><br><u>%Recovery</u> | <u>MSD</u><br><u>Qualifier</u> | <u>Limits</u> |
|------------------------------|--------------------------------|--------------------------------|---------------|
| 1,2-Dichloroethane-d4 (Surr) | 80                             |                                | 66 - 120      |

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

# QC Association Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## GC/MS VOA

### Analysis Batch: 553220

| Lab Sample ID    | Client Sample ID   | Prep Type | Matrix | Method    | Prep Batch |
|------------------|--------------------|-----------|--------|-----------|------------|
| 240-176530-2     | MW-228S_111422     | Total/NA  | Water  | 8260D SIM |            |
| MB 240-553220/5  | Method Blank       | Total/NA  | Water  | 8260D SIM |            |
| LCS 240-553220/3 | Lab Control Sample | Total/NA  | Water  | 8260D SIM |            |
| 240-176530-2 MS  | MW-228S_111422     | Total/NA  | Water  | 8260D SIM |            |
| 240-176530-2 MSD | MW-228S_111422     | Total/NA  | Water  | 8260D SIM |            |

### Analysis Batch: 553308

| Lab Sample ID      | Client Sample ID       | Prep Type | Matrix | Method | Prep Batch |
|--------------------|------------------------|-----------|--------|--------|------------|
| 240-176530-1       | TRIP BLANK_10          | Total/NA  | Water  | 8260D  |            |
| 240-176530-2       | MW-228S_111422         | Total/NA  | Water  | 8260D  |            |
| 240-176530-3       | MW-229S_111422         | Total/NA  | Water  | 8260D  |            |
| 240-176530-4       | MW-229_111422          | Total/NA  | Water  | 8260D  |            |
| 240-176530-5       | MW-229D_111422         | Total/NA  | Water  | 8260D  |            |
| MB 240-553308/8    | Method Blank           | Total/NA  | Water  | 8260D  |            |
| LCS 240-553308/5   | Lab Control Sample     | Total/NA  | Water  | 8260D  |            |
| 240-176531-B-5 MS  | Matrix Spike           | Total/NA  | Water  | 8260D  |            |
| 240-176531-B-5 MSD | Matrix Spike Duplicate | Total/NA  | Water  | 8260D  |            |

### Analysis Batch: 553480

| Lab Sample ID      | Client Sample ID       | Prep Type | Matrix | Method    | Prep Batch |
|--------------------|------------------------|-----------|--------|-----------|------------|
| 240-176530-3       | MW-229S_111422         | Total/NA  | Water  | 8260D SIM |            |
| 240-176530-4       | MW-229_111422          | Total/NA  | Water  | 8260D SIM |            |
| 240-176530-5       | MW-229D_111422         | Total/NA  | Water  | 8260D SIM |            |
| MB 240-553480/4    | Method Blank           | Total/NA  | Water  | 8260D SIM |            |
| LCS 240-553480/3   | Lab Control Sample     | Total/NA  | Water  | 8260D SIM |            |
| 240-176634-I-5 MS  | Matrix Spike           | Total/NA  | Water  | 8260D SIM |            |
| 240-176634-O-5 MSD | Matrix Spike Duplicate | Total/NA  | Water  | 8260D SIM |            |

# Lab Chronicle

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: TRIP BLANK\_10**

**Lab Sample ID: 240-176530-1**

Date Collected: 11/14/22 00:00

Matrix: Water

Date Received: 11/16/22 08:00

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab     | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|---------|----------------------|
| Total/NA  | Analysis   | 8260D        |     | 1               | 553308       | TJL1    | EET CAN | 11/23/22 14:40       |

**Client Sample ID: MW-228S\_111422**

**Lab Sample ID: 240-176530-2**

Date Collected: 11/14/22 10:10

Matrix: Water

Date Received: 11/16/22 08:00

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab     | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|---------|----------------------|
| Total/NA  | Analysis   | 8260D        |     | 1               | 553308       | TJL1    | EET CAN | 11/23/22 15:04       |
| Total/NA  | Analysis   | 8260D SIM    |     | 1               | 553220       | CS      | EET CAN | 11/23/22 03:11       |

**Client Sample ID: MW-229S\_111422**

**Lab Sample ID: 240-176530-3**

Date Collected: 11/14/22 11:20

Matrix: Water

Date Received: 11/16/22 08:00

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab     | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|---------|----------------------|
| Total/NA  | Analysis   | 8260D        |     | 1               | 553308       | TJL1    | EET CAN | 11/23/22 15:27       |
| Total/NA  | Analysis   | 8260D SIM    |     | 1               | 553480       | CS      | EET CAN | 11/27/22 20:58       |

**Client Sample ID: MW-229\_111422**

**Lab Sample ID: 240-176530-4**

Date Collected: 11/14/22 12:15

Matrix: Water

Date Received: 11/16/22 08:00

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab     | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|---------|----------------------|
| Total/NA  | Analysis   | 8260D        |     | 1               | 553308       | TJL1    | EET CAN | 11/23/22 15:50       |
| Total/NA  | Analysis   | 8260D SIM    |     | 1               | 553480       | CS      | EET CAN | 11/27/22 21:24       |

**Client Sample ID: MW-229D\_111422**

**Lab Sample ID: 240-176530-5**

Date Collected: 11/14/22 13:15

Matrix: Water

Date Received: 11/16/22 08:00

| Prep Type | Batch Type | Batch Method | Run | Dilution Factor | Batch Number | Analyst | Lab     | Prepared or Analyzed |
|-----------|------------|--------------|-----|-----------------|--------------|---------|---------|----------------------|
| Total/NA  | Analysis   | 8260D        |     | 1               | 553308       | TJL1    | EET CAN | 11/23/22 16:13       |
| Total/NA  | Analysis   | 8260D SIM    |     | 1               | 553480       | CS      | EET CAN | 11/27/22 21:49       |

**Laboratory References:**

EET CAN = Eurofins Canton, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

# Accreditation/Certification Summary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Laboratory: Eurofins Canton

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

| Authority             | Program | Identification Number | Expiration Date |
|-----------------------|---------|-----------------------|-----------------|
| California            | State   | 2927                  | 02-27-23        |
| Connecticut           | State   | PH-0590               | 12-31-23        |
| Florida               | NELAP   | E87225                | 06-30-23        |
| Georgia               | State   | 4062                  | 02-27-23        |
| Illinois              | NELAP   | 200004                | 07-31-23        |
| Iowa                  | State   | 421                   | 06-01-23        |
| Kentucky (UST)        | State   | 112225                | 02-27-23        |
| Kentucky (WW)         | State   | KY98016               | 12-31-22        |
| Minnesota             | NELAP   | 039-999-348           | 12-31-22        |
| Minnesota (Petrofund) | State   | 3506                  | 08-01-23        |
| New Jersey            | NELAP   | OH001                 | 06-30-23        |
| New York              | NELAP   | 10975                 | 04-01-23        |
| Ohio                  | State   | 8303                  | 02-27-23        |
| Ohio VAP              | State   | CL0024                | 02-27-23        |
| Oregon                | NELAP   | 4062                  | 02-27-23        |
| Pennsylvania          | NELAP   | 68-00340              | 08-31-23        |
| Texas                 | NELAP   | T104704517-22-17      | 08-31-23        |
| Virginia              | NELAP   | 460175                | 09-14-23        |
| Washington            | State   | C971                  | 01-12-23        |
| West Virginia DEP     | State   | 210                   | 12-31-22        |

TestAmerica Laboratory location: Brighton --- 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

39/39

|  |  |  |  |
|--|--|--|--|
| <b>Client Contact</b><br>Company Name: Arcadis<br>Address: 28550 Cabot Drive, Suite 500<br>City/State/Zip: Novi, MI, 48377<br>Phone: 248-994-2240  |  | <b>Regulatory program:</b><br><input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other  |  |
| <b>Client Project Manager:</b> Kris Hinskey<br>Telephone: 248-994-2240<br>Email: kristoffer.hinskey@arcadis.com  |  | <b>Lab Contact:</b> Mike DeMonico<br>Telephone: 310-497-9396   |  |
| <b>Project Name:</b> Ford LTP Off-Site<br>Project Number: 30146655-402.04<br>PO # 30146655-402.04  |  | <b>Analysis Turnaround Time</b><br>TAT if different from below:<br>3 weeks <input type="checkbox"/><br>2 weeks <input checked="" type="checkbox"/> <b>10 day</b><br>1 week <input type="checkbox"/><br>2 days <input type="checkbox"/><br>1 day <input type="checkbox"/> |  |
| <b>Sampler Name:</b><br>Sam Sukaria<br>Method of Shipment/Carrier:<br>Shipping/Tracking No:  |  | <b>Containers &amp; Preservatives</b><br>H2SO4 <input type="checkbox"/> HNO3 <input type="checkbox"/> HCl <input type="checkbox"/> NaOH <input type="checkbox"/> ZnAc <input type="checkbox"/> NaOH <input type="checkbox"/> Other:                                      |  |
| <b>Sample Identification</b><br>Sample Date      Sample Time   |  | <b>Matrix</b><br>Air <input type="checkbox"/> Sediment <input type="checkbox"/> Solid <input type="checkbox"/> Other:  |  |
| TRIP BLANK_10<br>MW-288-11472<br>MW-288-11472<br>MW-229-11472<br>MW-229-11472  |  | 1<br>6<br>6<br>6<br>6  |  |
| 1 Trip Blank<br>3 VOAs for 8260B<br>3 VOAs for 8260B SIM   |  | Filtered Sample (Y/N)<br>Composite C/Grab-G<br>1,1-DCE 8260B<br>Cis-1,2-DCE 8260B<br>Trans-1,2-DCE 8260B<br>PCE 8260B<br>TCE 8260B<br>Vinyl Chloride 8260B<br>1,4-Dioxane 8260B SIM  |  |
| Possible Hazard Identification<br><input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant   |  | Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)<br><input type="checkbox"/> Return to Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months  |  |
| Special Instructions/QC Requirements & Comments:<br>Sample Address: <b>Level 1st Row, 35416 York St Row</b><br>Submit all results through Cadena at jdemilio@cadematco.com. Cadena #E203831<br>Level IV Reporting requested. |  |  |  |
| Relinquished by: <b>Sam Sukaria</b>  |  | Received by: <b>Arcadis Cold Storage</b>   |  |
| Relinquished by: <b>[Signature]</b>  |  | Received by: <b>[Signature]</b>  |  |
| Relinquished by: <b>[Signature]</b>  |  | Received by: <b>[Signature]</b>  |  |
| Date/Time: 11/15/22 1530<br>Company: ARCADIS   |  | Date/Time: 11/15/22 0900<br>Company: ARCADIS   |  |
| Date/Time: 11/15/22 1145<br>Company: BEETA   |  | Date/Time: 11/15/22 0900<br>Company: BEETA   |  |
| Date/Time: 11/15/22 1530<br>Company: Arcadis   |  | Date/Time: 11/15/22 0900<br>Company: BEETA   |  |
| Date/Time: 11/15/22 1145<br>Company: BEETA   |  | Date/Time: 11/15/22 0900<br>Company: BEETA   |  |

©2000 TestAmerica Laboratories, Inc. All rights reserved.  
 TestAmerica is a registered trademark of TestAmerica Laboratories, Inc.



Eurofins - Canton Sample Receipt Form/Narrative

Login # : 116530

Barberton Facility

Client ARCADIS Site Name \_\_\_\_\_

Cooler unpacked by: Rachelle Haider

Cooler Received on 11-16-22 Opened on 11-16-22

FedEx: 1<sup>st</sup> Grd Exp UPS FAS Clipper Client Drop Off Eurofins Courier Other \_\_\_\_\_

Receipt After-hours: Drop-off Date/Time \_\_\_\_\_ Storage Location \_\_\_\_\_

Eurofins Cooler # TA Foam Box \_\_\_\_\_ Client Cooler \_\_\_\_\_ Box \_\_\_\_\_ Other \_\_\_\_\_

Packing material used: Bubble Wrap Foam Plastic Bag None Other \_\_\_\_\_

COOLANT: Water Ice Blue Ice Dry Ice Water None

1. Cooler temperature upon receipt  See Multiple Cooler Form  
IR GUN# IR-13 (CF +0.7 °C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C  
IR GUN #IR-15 (CF 0.0°C) Observed Cooler Temp. 3.9 °C Corrected Cooler Temp. 3.9 °C

2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 1  
-Were the seals on the outside of the cooler(s) signed & dated? Yes No  
-Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No NA  
-Were tamper/custody seals intact and uncompromised? Yes No NA

Tests that are not checked for pH by Receiving:  
VOAs  
Oil and Grease  
TOC

3. Shippers' packing slip attached to the cooler(s)? Yes No  
4. Did custody papers accompany the sample(s)? Yes No  
5. Were the custody papers relinquished & signed in the appropriate place? Yes No  
6. Was/were the person(s) who collected the samples clearly identified on the COC? Yes No  
7. Did all bottles arrive in good condition (Unbroken)? Yes No  
8. Could all bottle labels (ID/Date/Time) be reconciled with the COC? Yes No  
9. For each sample, does the COC specify preservatives (Y/N), # of containers (Y/N), and sample type of grab/comp (Y/N)? Yes No

10. Were correct bottle(s) used for the test(s) indicated? Yes No  
11. Sufficient quantity received to perform indicated analyses? Yes No  
12. Are these work share samples and all listed on the COC? Yes No

If yes, Questions 13-17 have been checked at the originating laboratory.

13. Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC286797  
14. Were VOAs on the COC? Yes No  
15. Were air bubbles >6 mm in any VOA vials? Yes NA Larger than this.  
16. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # Covered Yes No  
17. Was a LL Hg or Me Hg trip blank present? Yes No

Contacted PM \_\_\_\_\_ Date \_\_\_\_\_ by \_\_\_\_\_ via Verbal Voice Mail Other \_\_\_\_\_

Concerning \_\_\_\_\_

18. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES  additional next page

Samples processed by: \_\_\_\_\_

19. SAMPLE CONDITION

Sample(s) \_\_\_\_\_ were received after the recommended holding time had expired.  
Sample(s) \_\_\_\_\_ were received in a broken container.  
Sample(s) \_\_\_\_\_ were received with bubble >6 mm in diameter. (Notify PM)

20. SAMPLE PRESERVATION

Sample(s) \_\_\_\_\_ were further preserved in the laboratory.  
Time preserved: \_\_\_\_\_ Preservative(s) added/Lot number(s): \_\_\_\_\_

VOA Sample Preservation - Date/Time VOAs Frozen: \_\_\_\_\_

# DATA VERIFICATION REPORT



December 03, 2022

Kris Hinskey  
Arcadis Inc  
10559 Citation Ave  
Suite 100  
Brighton, MI 48116

CADENA project ID: E203631  
Project: Ford Livonia Transmission Project - OFF-SITE - Soil Gas and Groundwater  
Project number: 30146655.402.04 off-site  
Event Specific Scope of Work References: Sample COC  
Laboratory: Eurofins Environment Testing LLC - Barberton  
Laboratory submittal: 176530-1  
Sample date: 2022-11-14  
Report received by CADENA: 2022-12-02  
Initial Data Verification completed by CADENA: 2022-12-03  
Number of Samples:5  
Sample Matrices:Water  
Test Categories:GCMS VOC

**Please see attached criteria report or sample result/qualified analytical result summary for qualifier flags assigned to sample data.**

The following minor QC exceptions or missing information were noted:

GCMS VOC sample -005 MS or MSD recoveries but not both or RPD only were outliers for VINYL CHLORIDE so client sample results were not qualified based on this QC outlier alone.

Sample/MS/MSD Surrogate Recovery, Blank/LCS Surrogate Recovery, LCS/LCD Recovery, MS/MSD Recovery, MS/MSD RPD, Blank Contamination and Hold Time Exception were reviewed as part of our verification.

Data verification for the report specified above was completed using the Ford Motor Company Environmental Laboratory Technical Specification, the CADENA Standard Operating Procedure for the Verification of Environmental Analytical Data and the associated analytical methods as references for evaluating the batch QC, sample data and report content. The EPA National Functional Guidelines for validating organic and inorganic data were used as guidance when addressing out of control QC results and the associated data qualifiers.

The definitions of the qualifiers used for this data package are defined in the analytical report. CADENA valid qualifiers are defined in the table below. To view and download a PDF copy of the laboratory analytical report access the CADENA CLMS at <http://clms.cadenaco.com/index.cfm>.

Please contact me if you have any questions.

Sincerely,

Jim Tomalia

Project Scientist



CADENA Inc, 1099 Highland Drive, Suite E, Ann Arbor, MI 48108 517-819-0356

## CADENA Valid Qualifiers

| Valid Qualifiers | Description  |
|------------------|--|
| <                | Less than the reported concentration.  |
| >                | Greater than the reported concentration.   |
| B                | The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was greater than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the reported concentration. For Inorganic methods the sample concentration was greater than the RDL and less than 10x the blank concentration and is considered non-detect at the reported concentration. |
| E                | The analyte / Compound reported exceeds the calibration range and is considered estimated.   |
| EMPC             | Estimated Minimum Potential Contamination - Dioxin/Furan analyses only.  |
| J                | Indicates an estimated value. This flag is used either when estimating a concentration for a tentatively identified compound or when the data indicates the presence of an analyte / compound but the result is less than the sample Quantitation limit, but greater than zero. The flag is also used in data validation to indicate a reported value should be considered estimated due to associated quality assurance deficiencies.                     |
| J-               | The result is an estimated quantity, but the result may be biased low.   |
| JB               | NON-DETECT AT THE CONCENTRATION REPORTED AND ESTIMATED   |
| JH               | The sample result is considered estimated and is potentially biased high.  |
| JL               | The sample result is considered estimated and is potentially biased low.   |
| JUB              | NON-DETECT AT THE REPORTING LIMIT AND ESTIMATED  |
| NJ               | Tentatively identified compound with approximated concentration.   |
| R                | Indicates the value is considered to be unusable. (Note: The analyte / compound may or may not be present.)  |
| TNTC             | Too Numerous to Count - Asbestos and Microbiological Results.  |
| U                | Indicates that the analyte / compound was analyzed for, but not detected.  |
| UB               | The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was less than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the RDL. For Inorganic methods the sample concentration was less than the RDL and less than 10x the blank concentration and is considered non-detect at the RDL.   |
| UJ               | The analyte / compound was not detected above the reported sample Quantitation limit. However, the Quantitation limit is considered to be approximate due to associated quality assurance results and may or may not represent the actual limit of Quantitation to accurately and precisely report the analyte in the sample.  |

## Analytical Results Summary

CADENA Project ID: E203631

Laboratory: Eurofins Environment Testing LLC - Barberton

Laboratory Submittal: 176530-1

| Analyte                  | Cas No.  | Sample Name: TRIP BLANK_10 |       |       |       | Sample Name: MW-228S_111422 |       |       |       | Sample Name: MW-229S_111422 |       |       |       | Sample Name: MW-229_111422 |       |       |       | Sample Name: MW-229D_111422 |       |       |       |
|--------------------------|----------|----------------------------|-------|-------|-------|-----------------------------|-------|-------|-------|-----------------------------|-------|-------|-------|----------------------------|-------|-------|-------|-----------------------------|-------|-------|-------|
|                          |          | Result                     | Limit | Units | Valid | Result                      | Limit | Units | Valid | Result                      | Limit | Units | Valid | Result                     | Limit | Units | Valid | Result                      | Limit | Units | Valid |
| <b>GC/MS VOC</b>         |          |                            |       |       |       |                             |       |       |       |                             |       |       |       |                            |       |       |       |                             |       |       |       |
| <u>OSW-8260D</u>         |          |                            |       |       |       |                             |       |       |       |                             |       |       |       |                            |       |       |       |                             |       |       |       |
| 1,1-Dichloroethene       | 75-35-4  | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   |
| cis-1,2-Dichloroethene   | 156-59-2 | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   |
| Tetrachloroethene        | 127-18-4 | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   |
| trans-1,2-Dichloroethene | 156-60-5 | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   |
| Trichloroethene          | 79-01-6  | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   |
| Vinyl chloride           | 75-01-4  | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   | ND                         | 1.0   | ug/l  | ---   | ND                          | 1.0   | ug/l  | ---   |
| <u>OSW-8260DSIM</u>      |          |                            |       |       |       |                             |       |       |       |                             |       |       |       |                            |       |       |       |                             |       |       |       |
| 1,4-Dioxane              | 123-91-1 |                            |       |       |       | ND                          | 2.0   | ug/l  | ---   | ND                          | 2.0   | ug/l  | ---   | ND                         | 2.0   | ug/l  | ---   | ND                          | 2.0   | ug/l  | ---   |

# Ford Motor Company – Livonia Transmission Project

## DATA REVIEW

### Livonia, Michigan

Volatile Organic Compounds (VOC) Analysis

SDG # 240-176530-1

CADENA Verification Report: 2022-12-03

Analyses Performed By:

TestAmerica

North Canton, Ohio

Report # 47931R

Review Level: Tier III

Project: 30146655.402.02

---

## DATA REVIEW

### SUMMARY

This data quality assessment summarizes the review of Sample Delivery Group (SDG) # 240-176530-1 for samples collected in association with the Ford – Livonia, Michigan site. The review was conducted as a Tier III validation in addition to a verification/Tier II validation review performed by CADENA Inc. and included review of level IV laboratory data package completeness. Only elements of a Tier III validation effort (Tier III) include a detailed review of laboratory raw data to check for errors in calculation, calibration review, internal standard review and compound identification) and omitted deviations from the CADENA verification/Tier II report are documented in this report. Only analytical data associated with constituents of concern were reviewed for this validation. Field documentation was not included in this review. Included with this assessment are the validation annotated sample result sheets, and chain of custody. Analyses were performed on the following samples:

| Sample ID      | Lab ID       | Matrix | Sample Collection Date | Parent Sample | Analysis |         |
|----------------|--------------|--------|------------------------|---------------|----------|---------|
|                |              |        |                        |               | VOC      | VOC SIM |
| TRIP BLANK_10  | 240-176530-1 | Water  | 11/14/2022             |               | X        |         |
| MW-228S_111422 | 240-176530-2 | Water  | 11/14/2022             |               | X        | X       |
| MW-229S_111422 | 240-176530-3 | Water  | 11/14/2022             |               | X        | X       |
| MW-229_111422  | 240-176530-4 | Water  | 11/14/2022             |               | X        | X       |
| MW-229D_111422 | 240-176530-5 | Water  | 11/14/2022             |               | X        | X       |

## DATA REVIEW

### ANALYTICAL DATA PACKAGE DOCUMENTATION

The table below is the evaluation of the data package completeness.

| Items Reviewed   | Reported |     | Performance Acceptable |     | Not Required |
|--|----------|-----|------------------------|-----|--------------|
|  | No       | Yes | No                     | Yes |              |
| 1. Sample receipt condition  |          | X   |                        | X   |              |
| 2. Requested analyses and sample results                               |          | X   |                        | X   |              |
| 3. Master tracking list  |          | X   |                        | X   |              |
| 4. Methods of analysis   |          | X   |                        | X   |              |
| 5. Reporting limits  |          | X   |                        | X   |              |
| 6. Sample collection date  |          | X   |                        | X   |              |
| 7. Laboratory sample received date                                     |          | X   |                        | X   |              |
| 8. Sample preservation verification (as applicable)                    |          | X   |                        | X   |              |
| 9. Sample preparation/extraction/analysis dates                        |          | X   |                        | X   |              |
| 10. Fully executed Chain-of-Custody (COC) form                         |          | X   |                        | X   |              |
| 11. Narrative summary of Quality Assurance or sample problems provided |          | X   |                        | X   |              |
| 12. Data Package Completeness and Compliance                           |          | X   |                        | X   |              |

## DATA REVIEW

### ORGANIC ANALYSIS INTRODUCTION

Analyses were performed according to United States Environmental Protection Agency (USEPA) SW-846 Method 8260D and 8260D SIM. Data were reviewed in accordance with USEPA National Functional Guidelines for Organic Superfund Methods Data Review, EPA-540-R-20-005 November 2020 (with reference to the historical USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review, OSWER 9240.1-05A-P, October 1999, as appropriate).

The data review process is an evaluation of data on a technical basis rather than a determination of contract compliance. As such, the standards against which the data are being weighed may differ from those specified in the analytical method. It is assumed that the data package represents the best efforts of the laboratory and had already been subjected to adequate and sufficient quality review prior to submission.

During the review process, laboratory qualified, and unqualified data are verified against the supporting documentation. Based on this evaluation, qualifier codes may be added, deleted, or modified by the data reviewer. Results are qualified with the following codes in accordance with USEPA National Functional Guidelines:

- Concentration (C) Qualifiers
  - U The analyte was analyzed for but was not detected above the level of the reported sample quantitation limit.
  - B The compound has been found in the sample as well as its associated blank, its presence in the sample may be suspect.
- Quantitation (Q) Qualifiers
  - E The compound was quantitated above the calibration range.
  - D Concentration is based on a diluted sample analysis.
- Validation Qualifiers
  - J The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
  - UJ The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
  - J+ The result is an estimated quantity, but the result may be biased high.
  - J- The result is an estimated quantity, but the result may be biased low.
  - UB Analyte considered non-detect at the listed value due to associated blank contamination.
  - N The analysis indicates the presence of a compound for which there is presumptive evidence to make a tentative identification.
  - R The sample results are rejected.

Two facts should be noted by all data users. First, the "R" flag means that the associated value is unusable. In other words, due to significant quality control (QC) problems, the analysis is invalid and provides no information as to whether the compound is present or not. "R" values should not appear on data tables because they cannot be relied upon, even as a last resort. The second fact to keep in mind is that no compound concentration, even

## **DATA REVIEW**

if it has passed all QC tests, is guaranteed to be accurate. Strict QC serves to increase confidence in data, but any value potentially contains error.



## DATA REVIEW

### VOLATILE ORGANIC COMPOUND (VOC) ANALYSES

#### 1. Holding Times

The specified holding times for the following methods are presented in the following table.

| Method                 | Matrix | Holding Time                        | Preservation                    |
|------------------------|--------|-------------------------------------|---------------------------------|
| SW-846 8260D/8260D-SIM | Water  | 14 days from collection to analysis | Cool to < 6 °C; pH < 2 with HCl |

All samples were analyzed within the specified holding time criteria.

#### 2. Mass Spectrometer Tuning

Mass spectrometer performance was acceptable, and all analyses were performed within a 12-hour tune clock. System performance and column resolution were acceptable.

#### 3. Calibration

Satisfactory instrument calibration is established to ensure that the instrument is capable of producing acceptable quantitative data. An initial calibration demonstrates that the instrument is capable of acceptable performance at the beginning of an experimental sequence. The continuing calibration verifies that the instrument daily performance is satisfactory.

##### 3.1 Initial Calibration

The method specifies percent relative standard deviation (%RSD) and relative response factor (RRF) limits for select compounds only. A technical review of the data applies limits to all compounds with no exceptions.

All target compounds associated with the initial calibration standards must exhibit a %RSD less than the control limit (20%) or a correlation coefficient greater than 0.99 and an RRF value greater than control limit (0.05).

All compounds associated with the initial calibrations were within the specified control limits.

##### 3.2 Continuing Calibration

All target compounds associated with the continuing calibration standard must exhibit a percent difference (%D) less than the control limit (20%) and RRF value greater than control limit (0.05).

All compounds associated with the calibrations were within the specified control limits.

#### 4. Internal Standard Performance

Internal standard performance criteria ensure that the GC/MS sensitivity and response are stable during every sample analysis. The criteria require the internal standard compounds associated with the VOC exhibit area counts that are not greater than two times (+100%) or less than one-half (-50%) of the area counts of the associated continuing calibration standard.

All internal standard responses were within control limits.

#### 5. Field Duplicate Analysis

Field duplicate analysis is used to assess the overall precision of the field sampling procedures and analytical method. A control limit of 30% for water matrices is applied to the RPD between the parent sample and the field duplicate. In the instance when the parent and/or duplicate sample concentrations are less than or equal to 5 times the RL, a control limit of two times the RL is applied for water matrices.

A field duplicate sample is not collected for samples from this SDG.

#### 6. Compound Identification

Compounds are identified on the GC/MS by using the analytes relative retention time and ion spectra.

## **DATA REVIEW**

No compounds were detected in the samples within this SDG.

### **7. System Performance and Overall Assessment**

Overall system performance was acceptable. Other than for those deviations specifically mentioned in this review, the overall data quality is within the guidelines specified in the method.

**DATA REVIEW**

**DATA VALIDATION CHECKLIST FOR VOCs**

| VOCs: 8260D/8260D-SIM                                       | Reported |     | Performance Acceptable |     | Not Required |
|---|----------|-----|------------------------|-----|--------------|
|   | No       | Yes | No                     | Yes |              |
| <b>GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS)</b>         |          |     |                        |     |              |
| <b>Tier II Validation</b>                                   |          |     |                        |     |              |
| Holding times/Preservation                                  |          | X   |                        | X   |              |
| <b>Tier III Validation</b>                                  |          |     |                        |     |              |
| System performance and column resolution                    |          | X   |                        | X   |              |
| Initial calibration %RSDs                                   |          | X   |                        | X   |              |
| Continuing calibration RRFs                                 |          | X   |                        | X   |              |
| Continuing calibration %Ds                                  |          | X   |                        | X   |              |
| Instrument tune and performance check                       |          | X   |                        | X   |              |
| Ion abundance criteria for each instrument used             |          | X   |                        | X   |              |
| Field Duplicate RPD   | X        |     |                        |     | X            |
| Internal standard   |          | X   |                        | X   |              |
| Compound identification and quantitation                    |          |     |                        |     |              |
| A. Reconstructed ion chromatograms                          |          | X   |                        | X   |              |
| B. Quantitation Reports                                     |          | X   |                        | X   |              |
| C. RT of sample compounds within the established RT windows |          | X   |                        | X   |              |
| D. Transcription/calculation errors present                 |          | X   |                        | X   |              |
| E. Reporting limits adjusted to reflect sample dilutions    |          | X   |                        | X   |              |

Notes:

%RSD Relative standard deviation

%R Percent recovery

RPD Relative percent difference

%D Percent difference

## DATA REVIEW

VALIDATION PERFORMED BY: Hareesha Naik

SIGNATURE: HaliL

DATE: December 14, 2022

PEER REVIEW: Andrew Korycinski

DATE: December 17, 2022

**NO CORRECTIONS/QUALIFIERS ADDED  
TO SAMPLE ANALYSIS DATA SHEETS**

**CHAIN OF CUSTODY  
CORRECTED SAMPLE ANALYSIS DATA  
SHEETS**



TestAmerica Laboratory location: Brighton — 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|---------------------------------------|--|---|--|--------------------------------|--|---------|---|-------|--------|-------------|------|-----|------|-----------------------|--------------------|---------------|-------------------|-----------------------|--------------------|---------------|----------------------|-----------------------|--------------|-----------|----------------------|-----------------------|--|--------|--|--|---|
| Client Contact                        |  | Regulatory program: <input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other |  | TestAmerica Laboratories, Inc. |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
| Company Name: Arcadis                 |  | Client Project Manager: Kris Hinskey  |  | Site Contact: Christina Weaver |  |         | Lab Contact: Mike DelMonico   |       |        | COC No:     |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
| Address: 28550 Cabot Drive, Suite 500 |  | Telephone: 248-994-2240   |  | Telephone: 248-994-2293        |  |         | Telephone: 330-497-9396   |       |        | 1 of 1 COCs |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
| City/State/Zip: Novi, MI, 48377       |  | Email: kristoffer.hinskey@arcadis.com   |  | Analysis Turnaround Time       |  |         | Analyses  |       |        |             |      |     |      | For lab use only      |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
| Phone: 248-994-2240                   |  | Sampler Name: <i>Sym Sunkarta</i>   |  | FAT if different from below    |  |         | <input type="checkbox"/> 3 weeks<br><input checked="" type="checkbox"/> 2 weeks<br><input type="checkbox"/> 1 week<br><input type="checkbox"/> 2 days<br><input type="checkbox"/> 1 day |       |        |             |      |     |      | Walk-in client        |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
| Project Name: Ford LTP Off-Site       |  | Method of Shipment/Carrier:   |  | 10 day                         |  |         |   |       |        |             |      |     |      | Filtered Sample (Y/N) | Composite=C/Grab=G | 1,1-DCE 8260B | cis-1,2-DCE 8260B | Trans-1,2-DCE 8260B   | PCE 8260B          | TCE 8260B     | Vinyl Chloride 8260B | 1,4-Dioxane 8260B SIM | Lab sampling |           |                      |                       |  |        |  |  |   |
| Project Number: 30146655.402.04       |  | Shipping/Tracking No:   |  | Matrix                         |  |         | Containers & Preservatives  |       |        |             |      |     |      | Job/SDG No:           |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
| PO # 30146655.402.04                  |  |   |  | Air                            |  |         | H2SO4   |       |        | HNO3        |      |     | HCl  |                       |                    | NaOH          |                   |                       | ZnAc               |               |                      | NaOH                  |              |           | Unpres               |                       |  | Other: |  |  | Sample Specific Notes / Special Instructions: |
| Sample Identification                 |  | Sample Date   |  | Sample Time                    |  | Aqueous | Sediment  | Solid | Other: | H2SO4       | HNO3 | HCl | NaOH | ZnAc                  | NaOH               | Unpres        | Other:            | Filtered Sample (Y/N) | Composite=C/Grab=G | 1,1-DCE 8260B | cis-1,2-DCE 8260B    | Trans-1,2-DCE 8260B   | PCE 8260B    | TCE 8260B | Vinyl Chloride 8260B | 1,4-Dioxane 8260B SIM |  |        |  |  |   |
| TRIP BLANK_10                         |  | ---   |  | ---                            |  | 1       |   |       |        | 1           |      |     |      |                       |                    |               |                   | NG                    | X                  | X             | X                    | X                     | X            | X         |                      |                       | 1 Trip Blank                             |        |  |  |   |
| MW-228c-111422                        |  | 11/14/22  |  | 1010                           |  | 6       |   |       |        | 6           |      |     |      |                       |                    |               |                   | NG                    | X                  | X             | X                    | X                     | X            | X         |                      |                       | 3 VOAs for 8260B<br>3 VOAs for 8260B SIM |        |  |  |   |
| MW-228c-111422                        |  | 11/14/22  |  | 1120                           |  | 6       |   |       |        | 6           |      |     |      |                       |                    |               |                   | NG                    | X                  | X             | X                    | X                     | X            | X         |                      |                       |  |        |  |  |   |
| MW-229-111422                         |  | 11/14/22  |  | 1215                           |  | 6       |   |       |        | 6           |      |     |      |                       |                    |               |                   | MG                    | X                  | X             | X                    | X                     | X            | X         |                      |                       |  |        |  |  |   |
| MW-229D-111422                        |  | 11/14/22  |  | 1315                           |  | 6       |   |       |        | 6           |      |     |      |                       |                    |               |                   | MG                    | X                  | X             | X                    | X                     | X            | X         |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |
|                                       |  |   |  |                                |  |         |   |       |        |             |      |     |      |                       |                    |               |                   |                       |                    |               |                      |                       |              |           |                      |                       |  |        |  |  |   |

# Definitions/Glossary

Client: ARCADIS U.S., Inc.  
Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

## Qualifiers

### GC/MS VOA

| Qualifier | Qualifier Description                                    |
|-----------|--|
| F1        | MS and/or MSD recovery exceeds control limits.           |
| U         | Indicates the analyte was analyzed for but not detected. |

## Glossary

| Abbreviation   | These commonly used abbreviations may or may not be present in this report.                                 |
|----------------|---|
| α              | Listed under the "D" column to designate that the result is reported on a dry weight basis                  |
| %R             | Percent Recovery  |
| CFL            | Contains Free Liquid  |
| CFU            | Colony Forming Unit   |
| CNF            | Contains No Free Liquid   |
| DER            | Duplicate Error Ratio (normalized absolute difference)  |
| Dil Fac        | Dilution Factor   |
| DL             | Detection Limit (DoD/DOE)   |
| DL, RA, RE, IN | Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample |
| DLC            | Decision Level Concentration (Radiochemistry)   |
| EDL            | Estimated Detection Limit (Dioxin)  |
| LOD            | Limit of Detection (DoD/DOE)  |
| LOQ            | Limit of Quantitation (DoD/DOE)   |
| MCL            | EPA recommended "Maximum Contaminant Level"   |
| MDA            | Minimum Detectable Activity (Radiochemistry)  |
| MDC            | Minimum Detectable Concentration (Radiochemistry)   |
| MDL            | Method Detection Limit  |
| ML             | Minimum Level (Dioxin)  |
| MPN            | Most Probable Number  |
| MQL            | Method Quantitation Limit   |
| NC             | Not Calculated  |
| ND             | Not Detected at the reporting limit (or MDL or EDL if shown)  |
| NEG            | Negative / Absent   |
| POS            | Positive / Present  |
| PQL            | Practical Quantitation Limit  |
| PRES           | Presumptive   |
| QC             | Quality Control   |
| RER            | Relative Error Ratio (Radiochemistry)   |
| RL             | Reporting Limit or Requested Limit (Radiochemistry)   |
| RPD            | Relative Percent Difference, a measure of the relative difference between two points                        |
| TEF            | Toxicity Equivalent Factor (Dioxin)   |
| TEQ            | Toxicity Equivalent Quotient (Dioxin)   |
| TNTC           | Too Numerous To Count   |



# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: TRIP BLANK\_10**

**Lab Sample ID: 240-176530-1**

**Date Collected: 11/14/22 00:00**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 14:40 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 14:40 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 14:40 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 14:40 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 14:40 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 14:40 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 106       |           | 62 - 137 |          | 11/23/22 14:40 | 1       |
| 4-Bromofluorobenzene (Surr)  | 88        |           | 56 - 136 |          | 11/23/22 14:40 | 1       |
| Toluene-d8 (Surr)            | 106       |           | 78 - 122 |          | 11/23/22 14:40 | 1       |
| Dibromofluoromethane (Surr)  | 99        |           | 73 - 120 |          | 11/23/22 14:40 | 1       |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-228S\_111422**

**Lab Sample ID: 240-176530-2**

**Date Collected: 11/14/22 10:10**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte     | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/23/22 03:11 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 79        |           | 66 - 120 |          | 11/23/22 03:11 | 1       |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 15:04 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 15:04 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:04 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 15:04 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:04 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 15:04 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 106       |           | 62 - 137 |          | 11/23/22 15:04 | 1       |
| 4-Bromofluorobenzene (Surr)  | 91        |           | 56 - 136 |          | 11/23/22 15:04 | 1       |
| Toluene-d8 (Surr)            | 109       |           | 78 - 122 |          | 11/23/22 15:04 | 1       |
| Dibromofluoromethane (Surr)  | 97        |           | 73 - 120 |          | 11/23/22 15:04 | 1       |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-229S\_111422**

**Lab Sample ID: 240-176530-3**

**Date Collected: 11/14/22 11:20**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte                      | Result           | Qualifier        | RL            | MDL  | Unit | D | Prepared        | Analyzed        | Dil Fac        |
|------------------------------|------------------|------------------|---------------|------|------|---|-----------------|-----------------|----------------|
| 1,4-Dioxane                  | 2.0              | U                | 2.0           | 0.86 | ug/L |   |                 | 11/27/22 20:58  | 1              |
| <b>Surrogate</b>             | <b>%Recovery</b> | <b>Qualifier</b> | <b>Limits</b> |      |      |   | <b>Prepared</b> | <b>Analyzed</b> | <b>Dil Fac</b> |
| 1,2-Dichloroethane-d4 (Surr) | 77               |                  | 66 - 120      |      |      |   |                 | 11/27/22 20:58  | 1              |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                      | Result           | Qualifier        | RL            | MDL  | Unit | D | Prepared        | Analyzed        | Dil Fac        |
|------------------------------|------------------|------------------|---------------|------|------|---|-----------------|-----------------|----------------|
| 1,1-Dichloroethene           | 1.0              | U                | 1.0           | 0.49 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| cis-1,2-Dichloroethene       | 1.0              | U                | 1.0           | 0.46 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| Tetrachloroethene            | 1.0              | U                | 1.0           | 0.44 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| trans-1,2-Dichloroethene     | 1.0              | U                | 1.0           | 0.51 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| Trichloroethene              | 1.0              | U                | 1.0           | 0.44 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| Vinyl chloride               | 1.0              | U                | 1.0           | 0.45 | ug/L |   |                 | 11/23/22 15:27  | 1              |
| <b>Surrogate</b>             | <b>%Recovery</b> | <b>Qualifier</b> | <b>Limits</b> |      |      |   | <b>Prepared</b> | <b>Analyzed</b> | <b>Dil Fac</b> |
| 1,2-Dichloroethane-d4 (Surr) | 106              |                  | 62 - 137      |      |      |   |                 | 11/23/22 15:27  | 1              |
| 4-Bromofluorobenzene (Surr)  | 89               |                  | 56 - 136      |      |      |   |                 | 11/23/22 15:27  | 1              |
| Toluene-d8 (Surr)            | 104              |                  | 78 - 122      |      |      |   |                 | 11/23/22 15:27  | 1              |
| Dibromofluoromethane (Surr)  | 98               |                  | 73 - 120      |      |      |   |                 | 11/23/22 15:27  | 1              |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-229\_111422**

**Lab Sample ID: 240-176530-4**

**Date Collected: 11/14/22 12:15**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte     | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/27/22 21:24 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 77        |           | 66 - 120 |          | 11/27/22 21:24 | 1       |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 15:50 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 15:50 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:50 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 15:50 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 15:50 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 15:50 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 106       |           | 62 - 137 |          | 11/23/22 15:50 | 1       |
| 4-Bromofluorobenzene (Surr)  | 90        |           | 56 - 136 |          | 11/23/22 15:50 | 1       |
| Toluene-d8 (Surr)            | 102       |           | 78 - 122 |          | 11/23/22 15:50 | 1       |
| Dibromofluoromethane (Surr)  | 99        |           | 73 - 120 |          | 11/23/22 15:50 | 1       |

# Client Sample Results

Client: ARCADIS U.S., Inc.  
 Project/Site: Ford LTP - Off Site

Job ID: 240-176530-1

**Client Sample ID: MW-229D\_111422**

**Lab Sample ID: 240-176530-5**

**Date Collected: 11/14/22 13:15**

**Matrix: Water**

**Date Received: 11/16/22 08:00**

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

| Analyte     | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|-------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,4-Dioxane | 2.0    | U         | 2.0 | 0.86 | ug/L |   |          | 11/27/22 21:49 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 78        |           | 66 - 120 |          | 11/27/22 21:49 | 1       |

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

| Analyte                  | Result | Qualifier | RL  | MDL  | Unit | D | Prepared | Analyzed       | Dil Fac |
|--------------------------|--------|-----------|-----|------|------|---|----------|----------------|---------|
| 1,1-Dichloroethene       | 1.0    | U         | 1.0 | 0.49 | ug/L |   |          | 11/23/22 16:13 | 1       |
| cis-1,2-Dichloroethene   | 1.0    | U         | 1.0 | 0.46 | ug/L |   |          | 11/23/22 16:13 | 1       |
| Tetrachloroethene        | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 16:13 | 1       |
| trans-1,2-Dichloroethene | 1.0    | U         | 1.0 | 0.51 | ug/L |   |          | 11/23/22 16:13 | 1       |
| Trichloroethene          | 1.0    | U         | 1.0 | 0.44 | ug/L |   |          | 11/23/22 16:13 | 1       |
| Vinyl chloride           | 1.0    | U         | 1.0 | 0.45 | ug/L |   |          | 11/23/22 16:13 | 1       |

| Surrogate                    | %Recovery | Qualifier | Limits   | Prepared | Analyzed       | Dil Fac |
|------------------------------|-----------|-----------|----------|----------|----------------|---------|
| 1,2-Dichloroethane-d4 (Surr) | 104       |           | 62 - 137 |          | 11/23/22 16:13 | 1       |
| 4-Bromofluorobenzene (Surr)  | 87        |           | 56 - 136 |          | 11/23/22 16:13 | 1       |
| Toluene-d8 (Surr)            | 105       |           | 78 - 122 |          | 11/23/22 16:13 | 1       |
| Dibromofluoromethane (Surr)  | 96        |           | 73 - 120 |          | 11/23/22 16:13 | 1       |