

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Kristoffer Hinskey  
ARCADIS US Inc  
28550 Cabot Drive  
Suite 500  
Novi, Michigan 48377

Generated 5/24/2023 9:37:16 PM

## JOB DESCRIPTION

Ford LTP - Off Site

## JOB NUMBER

240-185407-1

# Eurofins Cleveland

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Environment Testing North Central, LLC Project Manager.

## Authorization



Generated  
5/24/2023 9:37:16 PM

Authorized for release by  
Michael DeMonico, Project Manager I  
[Michael.DeMonico@et.eurofinsus.com](mailto:Michael.DeMonico@et.eurofinsus.com)  
(330)497-9396



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Method Summary . . . . .	6
Sample Summary . . . . .	7
Detection Summary . . . . .	8
Client Sample Results . . . . .	9
Surrogate Summary . . . . .	14
QC Sample Results . . . . .	15
QC Association Summary . . . . .	17
Lab Chronicle . . . . .	18
Certification Summary . . . . .	19
Chain of Custody . . . . .	20
Receipt Checklists . . . . .	24

# Definitions/Glossary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Qualifiers

### GC/MS VOA

Qualifier	Qualifier Description
U	Indicates the analyte was analyzed for but not detected.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

---

**Job ID: 240-185407-1**

---

**Laboratory: Eurofins Cleveland**

---

**Narrative**

**Job Narrative**  
**240-185407-1**

**Receipt**

The samples were received on 5/16/2023 9:45 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 1.7°C and 1.8°C

**GC/MS VOA**

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15

# Method Summary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

Method	Method Description	Protocol	Laboratory
8260D	Volatile Organic Compounds by GC/MS	SW846	EET EDI
8260D SIM	Volatile Organic Compounds (GC/MS)	SW846	EET EDI
5030C	Purge and Trap	SW846	EET EDI

**Protocol References:**

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

**Laboratory References:**

EET EDI = Eurofins Edison, 777 New Durham Road, Edison, NJ 08817, TEL (732)549-3900

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15

# Sample Summary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
240-185407-1	TRIP BLANK_136	Water	05/11/23 00:00	05/16/23 09:45
240-185407-2	MW-79D_051123	Water	05/11/23 12:55	05/16/23 09:45
240-185407-3	DUP-12	Water	05/11/23 00:00	05/16/23 09:45
240-185407-4	MW-79SR_051123	Water	05/11/23 14:20	05/16/23 09:45
240-185407-5	MW-141S_051123	Water	05/11/23 11:25	05/16/23 09:45

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

# Detection Summary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Client Sample ID: TRIP BLANK\_136

Lab Sample ID: 240-185407-1

No Detections.

## Client Sample ID: MW-79D\_051123

Lab Sample ID: 240-185407-2

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Vinyl chloride	1.8		1.0	0.45	ug/L	1		8260D	Total/NA

## Client Sample ID: DUP-12

Lab Sample ID: 240-185407-3

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Vinyl chloride	2.2		1.0	0.45	ug/L	1		8260D	Total/NA

## Client Sample ID: MW-79SR\_051123

Lab Sample ID: 240-185407-4

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	2.6		2.0	0.86	ug/L	1		8260D SIM	Total/NA

## Client Sample ID: MW-141S\_051123

Lab Sample ID: 240-185407-5

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
1,4-Dioxane	2.0		2.0	0.86	ug/L	1		8260D SIM	Total/NA

This Detection Summary does not include radiochemical test results.

Eurofins Cleveland



# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: TRIP BLANK\_136**

**Lab Sample ID: 240-185407-1**

Date Collected: 05/11/23 00:00

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/23/23 21:43	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/23/23 21:43	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/23/23 21:43	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/23/23 21:43	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/23/23 21:43	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/23/23 21:43	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	87		70 - 128		05/23/23 21:43	1
Dibromofluoromethane (Surr)	94		77 - 124		05/23/23 21:43	1
Toluene-d8 (Surr)	86		80 - 120		05/23/23 21:43	1
4-Bromofluorobenzene	87		76 - 120		05/23/23 21:43	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: MW-79D\_051123**

**Lab Sample ID: 240-185407-2**

Date Collected: 05/11/23 12:55

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/21/23 04:13	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	99		75 - 133					05/21/23 04:13	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 00:05	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 00:05	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:05	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 00:05	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:05	1
<b>Vinyl chloride</b>	<b>1.8</b>		1.0	0.45	ug/L			05/24/23 00:05	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		70 - 128					05/24/23 00:05	1
Dibromofluoromethane (Surr)	97		77 - 124					05/24/23 00:05	1
Toluene-d8 (Surr)	85		80 - 120					05/24/23 00:05	1
4-Bromofluorobenzene	86		76 - 120					05/24/23 00:05	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: DUP-12**

**Lab Sample ID: 240-185407-3**

Date Collected: 05/11/23 00:00

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/21/23 04:35	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	100		75 - 133					05/21/23 04:35	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 00:25	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 00:25	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:25	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 00:25	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:25	1
<b>Vinyl chloride</b>	<b>2.2</b>		1.0	0.45	ug/L			05/24/23 00:25	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	92		70 - 128					05/24/23 00:25	1
Dibromofluoromethane (Surr)	99		77 - 124					05/24/23 00:25	1
Toluene-d8 (Surr)	85		80 - 120					05/24/23 00:25	1
4-Bromofluorobenzene	84		76 - 120					05/24/23 00:25	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: MW-79SR\_051123**

**Lab Sample ID: 240-185407-4**

Date Collected: 05/11/23 14:20

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.6		2.0	0.86	ug/L			05/21/23 04:56	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	100		75 - 133					05/21/23 04:56	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 00:46	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 00:46	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:46	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 00:46	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:46	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/24/23 00:46	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		70 - 128					05/24/23 00:46	1
Dibromofluoromethane (Surr)	99		77 - 124					05/24/23 00:46	1
Toluene-d8 (Surr)	84		80 - 120					05/24/23 00:46	1
4-Bromofluorobenzene	85		76 - 120					05/24/23 00:46	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: MW-141S\_051123**

**Lab Sample ID: 240-185407-5**

Date Collected: 05/11/23 11:25

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0		2.0	0.86	ug/L			05/21/23 05:18	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	98		75 - 133					05/21/23 05:18	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 01:06	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 01:06	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 01:06	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 01:06	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 01:06	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/24/23 01:06	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	90		70 - 128					05/24/23 01:06	1
Dibromofluoromethane (Surr)	99		77 - 124					05/24/23 01:06	1
Toluene-d8 (Surr)	83		80 - 120					05/24/23 01:06	1
4-Bromofluorobenzene	86		76 - 120					05/24/23 01:06	1

# Surrogate Summary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Method: 8260D - Volatile Organic Compounds by GC/MS

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA	DBFM	TOL	BFB
		(70-128)	(77-124)	(80-120)	(76-120)
240-185407-1	TRIP BLANK_136	87	94	86	87
240-185407-2	MW-79D_051123	91	97	85	86
240-185407-3	DUP-12	92	99	85	84
240-185407-4	MW-79SR_051123	91	99	84	85
240-185407-5	MW-141S_051123	90	99	83	86
LCS 460-910908/3	Lab Control Sample	86	90	90	96
LCS 460-910908/4	Lab Control Sample Dup	86	90	91	92
MB 460-910908/9	Method Blank	85	94	84	91

### Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)  
DBFM = Dibromofluoromethane (Surr)  
TOL = Toluene-d8 (Surr)  
BFB = 4-Bromofluorobenzene

## Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB
		(75-133)
240-185407-2	MW-79D_051123	99
240-185407-3	DUP-12	100
240-185407-4	MW-79SR_051123	100
240-185407-5	MW-141S_051123	98
LCS 460-910494/3	Lab Control Sample	100
LCS 460-910494/4	Lab Control Sample Dup	97
MB 460-910494/7	Method Blank	97

### Surrogate Legend

BFB = 4-Bromofluorobenzene

# QC Sample Results

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Method: 8260D - Volatile Organic Compounds by GC/MS

**Lab Sample ID: MB 460-910908/9**  
**Matrix: Water**  
**Analysis Batch: 910908**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/23/23 20:22	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/23/23 20:22	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/23/23 20:22	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/23/23 20:22	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/23/23 20:22	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			05/23/23 20:22	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
1,2-Dichloroethane-d4 (Surr)	85		70 - 128		05/23/23 20:22	1
Dibromofluoromethane (Surr)	94		77 - 124		05/23/23 20:22	1
Toluene-d8 (Surr)	84		80 - 120		05/23/23 20:22	1
4-Bromofluorobenzene	91		76 - 120		05/23/23 20:22	1

**Lab Sample ID: LCS 460-910908/3**  
**Matrix: Water**  
**Analysis Batch: 910908**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
1,1-Dichloroethene	20.0	17.3		ug/L		87	68 - 133
cis-1,2-Dichloroethene	20.0	16.6		ug/L		83	78 - 121
Tetrachloroethene	20.0	18.7		ug/L		93	70 - 127
trans-1,2-Dichloroethene	20.0	16.8		ug/L		84	74 - 126
Trichloroethene	20.0	17.8		ug/L		89	71 - 121
Vinyl chloride	20.0	15.8		ug/L		79	55 - 144

Surrogate	LCS	LCS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	86		70 - 128
Dibromofluoromethane (Surr)	90		77 - 124
Toluene-d8 (Surr)	90		80 - 120
4-Bromofluorobenzene	96		76 - 120

**Lab Sample ID: LCSD 460-910908/4**  
**Matrix: Water**  
**Analysis Batch: 910908**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCSD	LCSD	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
		Result	Qualifier						
1,1-Dichloroethene	20.0	18.6		ug/L		93	68 - 133	7	30
cis-1,2-Dichloroethene	20.0	18.0		ug/L		90	78 - 121	8	30
Tetrachloroethene	20.0	20.5		ug/L		102	70 - 127	9	30
trans-1,2-Dichloroethene	20.0	18.6		ug/L		93	74 - 126	10	30
Trichloroethene	20.0	19.0		ug/L		95	71 - 121	7	30
Vinyl chloride	20.0	17.1		ug/L		86	55 - 144	8	30

Surrogate	LCSD	LCSD	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	86		70 - 128
Dibromofluoromethane (Surr)	90		77 - 124
Toluene-d8 (Surr)	91		80 - 120

Eurofins Cleveland

# QC Sample Results

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Method: 8260D - Volatile Organic Compounds by GC/MS (Continued)

Lab Sample ID: LCSD 460-910908/4  
Matrix: Water  
Analysis Batch: 910908

Client Sample ID: Lab Control Sample Dup  
Prep Type: Total/NA

	LCSD	LCSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene	92		76 - 120

## Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 460-910494/7  
Matrix: Water  
Analysis Batch: 910494

Client Sample ID: Method Blank  
Prep Type: Total/NA

Analyte	MB	MB							
	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			05/20/23 22:49	1

Surrogate	MB	MB						
	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac		
4-Bromofluorobenzene	97		75 - 133		05/20/23 22:49	1		

Lab Sample ID: LCS 460-910494/3  
Matrix: Water  
Analysis Batch: 910494

Client Sample ID: Lab Control Sample  
Prep Type: Total/NA

Analyte		Spike		LCS	LCS				%Rec
		Added	Result	Qualifier	Unit	D	%Rec	Limits	
1,4-Dioxane		5.00	4.92		ug/L		98	57 - 124	

	LCS	LCS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene	100		75 - 133

Lab Sample ID: LCSD 460-910494/4  
Matrix: Water  
Analysis Batch: 910494

Client Sample ID: Lab Control Sample Dup  
Prep Type: Total/NA

Analyte		Spike		LCSD	LCSD				%Rec		RPD
		Added	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit	
1,4-Dioxane		5.00	5.62		ug/L		112	57 - 124	13	30	

	LCSD	LCSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene	97		75 - 133



# QC Association Summary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## GC/MS VOA

### Analysis Batch: 910494

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-185407-2	MW-79D_051123	Total/NA	Water	8260D SIM	
240-185407-3	DUP-12	Total/NA	Water	8260D SIM	
240-185407-4	MW-79SR_051123	Total/NA	Water	8260D SIM	
240-185407-5	MW-141S_051123	Total/NA	Water	8260D SIM	
MB 460-910494/7	Method Blank	Total/NA	Water	8260D SIM	
LCS 460-910494/3	Lab Control Sample	Total/NA	Water	8260D SIM	
LCSD 460-910494/4	Lab Control Sample Dup	Total/NA	Water	8260D SIM	

### Analysis Batch: 910908

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-185407-1	TRIP BLANK_136	Total/NA	Water	8260D	
240-185407-2	MW-79D_051123	Total/NA	Water	8260D	
240-185407-3	DUP-12	Total/NA	Water	8260D	
240-185407-4	MW-79SR_051123	Total/NA	Water	8260D	
240-185407-5	MW-141S_051123	Total/NA	Water	8260D	
MB 460-910908/9	Method Blank	Total/NA	Water	8260D	
LCS 460-910908/3	Lab Control Sample	Total/NA	Water	8260D	
LCSD 460-910908/4	Lab Control Sample Dup	Total/NA	Water	8260D	

# Lab Chronicle

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Client Sample ID: TRIP BLANK\_136

Lab Sample ID: 240-185407-1

Date Collected: 05/11/23 00:00

Matrix: Water

Date Received: 05/16/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	910908	CJM	EET EDI	05/23/23 21:43

## Client Sample ID: MW-79D\_051123

Lab Sample ID: 240-185407-2

Date Collected: 05/11/23 12:55

Matrix: Water

Date Received: 05/16/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	910908	CJM	EET EDI	05/24/23 00:05
Total/NA	Analysis	8260D SIM		1	910494	KLB	EET EDI	05/21/23 04:13

## Client Sample ID: DUP-12

Lab Sample ID: 240-185407-3

Date Collected: 05/11/23 00:00

Matrix: Water

Date Received: 05/16/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	910908	CJM	EET EDI	05/24/23 00:25
Total/NA	Analysis	8260D SIM		1	910494	KLB	EET EDI	05/21/23 04:35

## Client Sample ID: MW-79SR\_051123

Lab Sample ID: 240-185407-4

Date Collected: 05/11/23 14:20

Matrix: Water

Date Received: 05/16/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	910908	CJM	EET EDI	05/24/23 00:46
Total/NA	Analysis	8260D SIM		1	910494	KLB	EET EDI	05/21/23 04:56

## Client Sample ID: MW-141S\_051123

Lab Sample ID: 240-185407-5

Date Collected: 05/11/23 11:25

Matrix: Water

Date Received: 05/16/23 09:45

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	910908	CJM	EET EDI	05/24/23 01:06
Total/NA	Analysis	8260D SIM		1	910494	KLB	EET EDI	05/21/23 05:18

### Laboratory References:

EET EDI = Eurofins Edison, 777 New Durham Road, Edison, NJ 08817, TEL (732)549-3900

# Accreditation/Certification Summary

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

## Laboratory: Eurofins Edison

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Connecticut	State	PH-0818	01-30-24
DE Haz. Subst. Cleanup Act (HSCA)	State	N/A	01-01-24
Georgia	State	12028 (NJ)	06-30-23
Massachusetts	State	M-NJ312	06-30-23
New Jersey	NELAP	12028	06-30-23
New York	NELAP	11452	04-01-24
Pennsylvania	NELAP	68-00522	03-01-24
Rhode Island	State	LAO00376	12-30-23
USDA	US Federal Programs	P330-20-00244	11-03-23

TestAmerica Laboratory location: Brighton --- 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

Regulatory program:  DW  NPDES  RCRA  Other

Client Contact: Arcadis  
 Company Name: Arcadis  
 Address: 28550 Cabot Drive, Suite 500  
 City/State/Zip: Novi, MI, 48377  
 Phone: 248-994-2240  
 Project Name: Ford LTP Off-Site  
 Project Number: 30167538-402.04  
 PO # 30167538-402.04

Client Project Manager: Kris Hinskey  
 Telephone: 248-994-2240  
 Email: kristoffer.hinskey@arcadis.com

Site Contact: Christina Weaver  
 Telephone: 248-994-2240

Lab Contact: Mike DelMontico  
 Telephone: 330-497-9396

TestAmerica Laboratories, Inc.  
 COC No: 1 of 1 COCs

Sampler Name: *Kent Kruper*  
 Method of Shipment/Carrier:  
 Shipping/Tracking No:

Analysis Turnaround Time  
 TAT if different from below  
 3 weeks  
 2 weeks  
 1 week  
 2 days  
 1 day

Sample Identification	Sample Date	Sample Time	Matrix				Containers & Preservatives				Filtered Sample (Y/N)	Composite C/Grab G	Analyses					Sample Specific Notes / Special Instructions:						
			Air	Aqueous	Sediment	Solid	Other:	H2SO4	HNO3	HCl			NaOH	ZnAc	NaOH	Uppres	Other:		1,1-DCE 8260B	CS-1,2-DCE 8260B	Trans-1,2-DCE 8260B	PCE 8260B	TCE 8260B	Vinyl Chloride 8260B
TRIP BLANK_ 136	5/11/23	---	1																					1 Trip Blank
MW-790-051123	5/11/23	1255	6																					3 VOAs for 8260B 3 VOAs for 8260B SIM
DUP-12	5/11/23	---	6																					
MW-795R-051123	5/11/23	1420	6																					
MW-1415-051123	5/11/23	11:25	6																					



Possible Hazard Identification  
 Non-Hazard  Flammable  Skin Irritant  Unknown  Poison B  Return to Client  Disposal By Lab  Archive For \_\_\_\_\_ Months

Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)

Special Instructions/QC Requirements & Comments:  
 Sample Address: BOSTON POST ROW  
 Submit all results through Cadena at jtomalia@cadenaco.com. Cadena #E203631  
 Level IV Reporting requested.

Relinquished by: *Kent Kruper*  
 Date/Time: 5/11/23 1534  
 Company: Arcadis

Received by: *Novi Cold Storage*  
 Date/Time: 5/15/23/0825  
 Company: ARCADIS

Relinquished by: *[Signature]*  
 Date/Time: 5/15/23 0900  
 Company: BETA

Received in Laboratory by: *[Signature]*  
 Date/Time: 5/16/23 945  
 Company: BETA



**Eurofins - Canton Sample Receipt Form/Narrative**  
**Barberton Facility**


Login #: 185407

Client Arcadis Site Name \_\_\_\_\_ Cooler unpacked by: Rachelle Hardet  
Cooler Received on 5/16/23 Opened on 5/16/23  
FedEx: 1<sup>st</sup> Grd  Exp  UPS  FAS  Clipper Client Drop Off Eurofins Courier Other \_\_\_\_\_

Receipt After-hours: Drop-off Date/Time \_\_\_\_\_ Storage Location \_\_\_\_\_

Eurofins Cooler # EC Foam Box  Client Cooler  Box  Other \_\_\_\_\_  
Packing material used:  Bubble Wrap  Foam  Plastic Bag  None Other \_\_\_\_\_  
COOLANT:  Wet Ice  Blue Ice  Dry Ice  Water  None

1. Cooler temperature upon receipt  See Multiple Cooler Form  
IR GUN # 13 (CF +2 °C) Observed Cooler Temp. \_\_\_\_\_ °C Corrected Cooler Temp. \_\_\_\_\_ °C

- 2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity 1  
-Were the seals on the outside of the cooler(s) signed & dated?  Yes  No  NA  
-Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)?  Yes  No  NA  
-Were tamper/custody seals intact and uncompromised?  Yes  No  NA
- 3. Shippers' packing slip attached to the cooler(s)?  Yes  No
- 4. Did custody papers accompany the sample(s)?  Yes  No
- 5. Were the custody papers relinquished & signed in the appropriate place?  Yes  No
- 6. Was/were the person(s) who collected the samples clearly identified on the COC?  Yes  No
- 7. Did all bottles arrive in good condition (Unbroken)?  Yes  No
- 8. Could all bottle labels (ID/Date/Time) be reconciled with the COC?  Yes  No
- 9. For each sample, does the COC specify preservatives (Y/N), # of containers (Y/N), and sample type of grab/comp (Y/N)?  Yes  No
- 10. Were correct bottle(s) used for the test(s) indicated?  Yes  No
- 11. Sufficient quantity received to perform indicated analyses?  Yes  No
- 12. Are these work share samples and all listed on the COC?  Yes  No  
If yes, Questions 13-17 have been checked at the originating laboratory.
- 13. Were all preserved sample(s) at the correct pH upon receipt?  Yes  No  NA pH Strip Lot# HC208070
- 14. Were VOAs on the COC?  Yes  No  NA 16/23
- 15. Were air bubbles >6 mm in any VOA vials?  Yes  No  NA LN  
 Larger than this.
- 16. Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # \_\_\_\_\_  Yes  No
- 17. Was a LL Hg or Me Hg trip blank present? \_\_\_\_\_  Yes  No

Tests that are not checked for pH by Receiving:  
VOAs  
Oil and Grease  
TOC

Contacted PM \_\_\_\_\_ Date \_\_\_\_\_ by \_\_\_\_\_ via Verbal Voice Mail Other \_\_\_\_\_  
Concerning \_\_\_\_\_

**18. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES**  additional next page Samples processed by: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**19. SAMPLE CONDITION**  
Sample(s) \_\_\_\_\_ were received after the recommended holding time had expired.  
Sample(s) \_\_\_\_\_ were received in a broken container.  
Sample(s) \_\_\_\_\_ were received with bubble >6 mm in diameter. (Notify PM)

**20. SAMPLE PRESERVATION**  
Sample(s) \_\_\_\_\_ were further preserved in the laboratory.  
Time preserved: \_\_\_\_\_ Preservative(s) added/Lot number(s): \_\_\_\_\_  
VOA Sample Preservation - Date/Time VOAs Frozen: \_\_\_\_\_

Eurofins - Canton Sample Receipt Multiple Cooler Form					
Cooler Description (Circle)	IR Gun # (Circle)	Observed Temp °C	Corrected Temp °C	Coolant (Circle)	
<u>EC</u> Client Box Other	IR GUN #: <u>13</u>	<u>1.6</u>	<u>1.8</u>	<u>Wet</u> Ice Water None	Blue Ice Dry Ice
<u>EC</u> Client Box Other	IR GUN #: <u>13</u>	<u>1.5</u>	<u>1.7</u>	<u>Wet</u> Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice
EC Client Box Other	IR GUN #: _____			Wet Ice Water None	Blue Ice Dry Ice

See Temperature Excursion Form



# Chain of Custody Record



**Client Information (Sub Contract Lab)**

Company: Eurofins Environment Testing Northeast  
Address: 777 New Durham Road  
City: Edison  
State, Zip: NJ, 08817  
Phone: 732-549-3900(Tel) 732-549-3679(Fax)  
Email:   
Project Name: Ford LTP Off Site  
Site:   
Project #: 24015353  
SSOW#:   
Due Date Requested: 5/29/2023  
TAT Requested (days):   
PO #:   
WO #:   
Project #:   
SSOW#:

Sampler: DelMonico, Michael  
Phone: Michael.DelMonico@et.eurofins.com  
Company: Eurofins Environment Testing Northeast  
Address: 777 New Durham Road  
City: Edison  
State, Zip: NJ, 08817  
Phone: 732-549-3900(Tel) 732-549-3679(Fax)  
Email:   
Project Name: Ford LTP Off Site  
Site:   
Project #: 24015353  
SSOW#:   
Due Date Requested: 5/29/2023  
TAT Requested (days):   
PO #:   
WO #:   
Project #:   
SSOW#:

Carrier Tracking No(s): 240-168174-1  
Page: Page 1 of 1  
Job #: 240-185407-1  
Preservation Codes:  
A HCL  
B NaOH  
C Zn Acetate  
D Nitric Acid  
E NaHSO4  
F MeOH  
G Amchlor  
H Ascorbic Acid  
I Ice  
J DI Water  
K EDTA  
L EDA  
Other  
M Hexane  
N None  
O ASNO2  
P Na2O4S  
Q Na2SO3  
R Na2S2O3  
S H2SO4  
T TSP Dodecahydrate  
U Acetone  
V MCAA  
W pH 4-5  
Y Trizma  
Z other (Specify)

Sample Identification - Client ID (Lab ID)	Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Matrix (Water, Spill, On-water, BTL, Acid)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	8260D/5030C (MOD) VOCs (Short List)	8260D SIM/6030C	Analysis Requested		Total Number of Containers	Special Instructions/Note
									Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)		
TRIP BLANK_136 (240-185407-1)	5/11/23	Eastern		Water	X		X				1	
MW-79D_051123 (240-185407-2)	5/11/23	12:55 Eastern		Water	X		X				6	
DUP-12 (240-185407-3)	5/11/23	Eastern		Water	X		X				6	
MW-79SR_051123 (240-185407-4)	5/11/23	14:20 Eastern		Water	X		X				6	
MW-141S_051123 (240-185407-5)	5/11/23	11:25 Eastern		Water	X		X				6	

**Possible Hazard Identification**

Unconfirmed  Return To Client  Disposal By Lab  Archive For  Months

Special Instructions/QC Requirements:

Primary Deliverable Rank: 2

Empty Kit Relinquished by: *Michelle Avelis* Date: *5/23/2023 12:40*  
 Relinquished by: *Michelle Avelis* Date: *5/23/2023 12:40*  
 Relinquished by: *Bhavishya* Date: *5/23/2023 10:40*  
 Relinquished by: *TRIGLITE LLC* Date: *5/24/2023*

Custody Seals Intact: Custody Seal No.   
 Yes  No

Cooler Temperature(s) °C and Other Remarks:

## Login Sample Receipt Checklist

Client: ARCADIS US Inc

Job Number: 240-185407-1

**Login Number: 185407**

**List Number: 2**

**Creator: Armbruster, Chris**

**List Source: Eurofins Edison**

**List Creation: 05/17/23 12:25 PM**

Question	Answer	Comment
Radioactivity wasn't checked or is <=/ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	





# DATA VERIFICATION REPORT



May 30, 2023

Kris Hinskey  
Arcadis Inc  
10559 Citation Ave  
Suite 100  
Brighton, MI 48116

CADENA project ID: E203631

Project: Ford Livonia Transmission Project - OFF-SITE - Soil Gas and Groundwater

Project number: 30167538.402.04 off-site

Event Specific Scope of Work References: Sample COC

Laboratory: Eurofins Environment Testing LLC - Cleveland

Laboratory submittal: 185407-1

Sample date: 2023-05-11

Report received by CADENA: 2023-05-30

Initial Data Verification completed by CADENA: 2023-05-30

Number of Samples:5

Sample Matrices:Water

Test Categories:GCMS VOC

**Please see attached criteria report or sample result/qualified analytical result summary for qualifier flags assigned to sample data.**

There were no significant QC anomalies or exceptions to report.

Sample/MS/MSD Surrogate Recovery, Blank/LCS Surrogate Recovery, LCS/LCD Recovery, LCS/LCD RPD, Blank Contamination and Hold Time Exception were reviewed as part of our verification.

Data verification for the report specified above was completed using the Ford Motor Company Environmental Laboratory Technical Specification, the CADENA Standard Operating Procedure for the Verification of Environmental Analytical Data and the associated analytical methods as references for evaluating the batch QC, sample data and report content. The EPA National Functional Guidelines for validating organic and inorganic data were used as guidance when addressing out of control QC results and the associated data qualifiers.

The definitions of the qualifiers used for this data package are defined in the analytical report. CADENA valid qualifiers are defined in the table below. To view and download a PDF copy of the laboratory analytical report access the CADENA CLMS at <http://clms.cadenaco.com/index.cfm>.

Please contact me if you have any questions.

Sincerely,

Jim Tomalia

Project Scientist

## CADENA Valid Qualifiers

Valid Qualifiers	Description
<	Less than the reported concentration.
>	Greater than the reported concentration.
B	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was greater than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the reported concentration. For Inorganic methods the sample concentration was greater than the RDL and less than 10x the blank concentration and is considered non-detect at the reported concentration.
E	The analyte / Compound reported exceeds the calibration range and is considered estimated.
EMPC	Estimated Minimum Potential Contamination - Dioxin/Furan analyses only.
J	Indicates an estimated value. This flag is used either when estimating a concentration for a tentatively identified compound or when the data indicates the presence of an analyte / compound but the result is less than the sample Quantitation limit, but greater than zero. The flag is also used in data validation to indicate a reported value should be considered estimated due to associated quality assurance deficiencies.
J-	The result is an estimated quantity, but the result may be biased low.
JB	NON-DETECT AT THE CONCENTRATION REPORTED AND ESTIMATED
JH	The sample result is considered estimated and is potentially biased high.
JL	The sample result is considered estimated and is potentially biased low.
JUB	NON-DETECT AT THE REPORTING LIMIT AND ESTIMATED
NJ	Tentatively identified compound with approximated concentration.
R	Indicates the value is considered to be unusable. (Note: The analyte / compound may or may not be present.)
TNTC	Too Numerous to Count - Asbestos and Microbiological Results.
U	Indicates that the analyte / compound was analyzed for, but not detected.
UB	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was less than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the RDL. For Inorganic methods the sample concentration was less than the RDL and less than 10x the blank concentration and is considered non-detect at the RDL.
UJ	The analyte / compound was not detected above the reported sample Quantitation limit. However, the Quantitation limit is considered to be approximate due to associated quality assurance results and may or may not represent the actual limit of Quantitation to accurately and precisely report the analyte in the sample.

## Analytical Results Summary

CADENA Project ID: E203631

Laboratory: Eurofins Environment Testing LLC - Cleveland

Laboratory Submittal: 185407-1

Analyte	Cas No.	Sample Name: TRIP BLANK_136				Sample Name: MW-79D_051123				Sample Name: DUP-12				Sample Name: MW-79SR_051123				Sample Name: MW-141S_051123			
		Result	Limit	Units	Valid	Result	Limit	Units	Valid	Result	Limit	Units	Valid	Result	Limit	Units	Valid	Result	Limit	Units	Valid
GC/MS VOC																					
<u>OSW-8260D</u>																					
1,1-Dichloroethene	75-35-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---
cis-1,2-Dichloroethene	156-59-2	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---
Tetrachloroethene	127-18-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---
trans-1,2-Dichloroethene	156-60-5	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---
Trichloroethene	79-01-6	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---
Vinyl chloride	75-01-4	ND	1.0	ug/l	---	1.8	1.0	ug/l	---	2.2	1.0	ug/l	---	ND	1.0	ug/l	---	ND	1.0	ug/l	---
<u>OSW-8260DSIM</u>																					
1,4-Dioxane	123-91-1					ND	2.0	ug/l	---	ND	2.0	ug/l	---	2.6	2.0	ug/l	---	2.0	2.0	ug/l	---

# Ford Motor Company – Livonia Transmission Project

## Data Review

### Livonia, Michigan

Volatile Organic Compounds (VOC) Analysis

SDG # 240-185407-1

CADENA Verification Report: 2023-05-30

Analyses Performed By:

Eurofins

North Canton, Ohio

Report # 49941R

Review Level: Tier III

Project: 30167538.402.02

## DATA REVIEW

### SUMMARY

This data quality assessment summarizes the review of Sample Delivery Group (SDG) # 240-185407-1 for samples collected in association with the Ford – Livonia, Michigan site. The review was conducted as a Tier III validation in addition to a verification/Tier II validation review performed by CADENA Inc. and included review of level IV laboratory data package completeness. Only elements of a Tier III validation effort (Tier III) include a detailed review of laboratory raw data to check for errors in calculation, calibration review, internal standard review and compound identification) and omitted deviations from the CADENA verification/Tier II report are documented in this report. Only analytical data associated with constituents of concern were reviewed for this validation. Field documentation was not included in this review. Included with this assessment are the validation annotated sample result sheets, and chain of custody. Analyses were performed on the following samples:

Sample ID	Lab ID	Matrix	Sample Collection Date	Parent Sample	Analysis	
					VOC	VOC SIM
TRIP BLANK_136	240-185407-1	Water	05/11/23		X	
MW-79D_051123	240-185407-2	Water	05/11/23		X	X
DUP-12	240-185407-3	Water	05/11/23	MW-79D_051123	X	X
MW-79SR_051123	240-185407-4	Water	05/11/23		X	X
MW-141S_051123	240-185407-5	Water	05/11/23		X	X

## DATA REVIEW

### ANALYTICAL DATA PACKAGE DOCUMENTATION

The table below is the evaluation of the data package completeness.

Items Reviewed	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
1. Sample receipt condition		X		X	
2. Requested analyses and sample results		X		X	
3. Master tracking list		X		X	
4. Methods of analysis		X		X	
5. Reporting limits		X		X	
6. Sample collection date		X		X	
7. Laboratory sample received date		X		X	
8. Sample preservation verification (as applicable)		X		X	
9. Sample preparation/extraction/analysis dates		X		X	
10. Fully executed Chain-of-Custody (COC) form		X		X	
11. Narrative summary of Quality Assurance or sample problems provided		X		X	
12. Data Package Completeness and Compliance		X		X	

## DATA REVIEW

### ORGANIC ANALYSIS INTRODUCTION

Analyses were performed according to United States Environmental Protection Agency (USEPA) SW-846 Method 8260D and 8260D SIM. Data were reviewed in accordance with USEPA National Functional Guidelines for Organic Superfund Methods Data Review, EPA 540-R-20-005, November 2020 (with reference to the historical USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review, OSWER 9240.1-05A-P, October 1999), as appropriate.

The data review process is an evaluation of data on a technical basis rather than a determination of contract compliance. As such, the standards against which the data are being weighed may differ from those specified in the analytical method. It is assumed that the data package represents the best efforts of the laboratory and had already been subjected to adequate and sufficient quality review prior to submission.

During the review process, laboratory qualified and unqualified data are verified against the supporting documentation. Based on this evaluation, qualifier codes may be added, deleted, or modified by the data reviewer. Results are qualified with the following codes in accordance with USEPA National Functional Guidelines:

- Concentration (C) Qualifiers
  - U The analyte was analyzed for but was not detected above the level of the reported sample quantitation limit.
  - B The compound has been found in the sample as well as its associated blank, its presence in the sample may be suspect.
- Quantitation (Q) Qualifiers
  - E The compound was quantitated above the calibration range.
  - D Concentration is based on a diluted sample analysis.
- Validation Qualifiers
  - J The result is an estimated quantity. The associated numerical value is the approximate concentration of the analyte in the sample.
  - UJ The analyte was analyzed for but was not detected. The reported quantitation limit is approximate and may be inaccurate or imprecise.
  - UB Analyte considered non-detect at the listed value due to associated blank contamination.
  - R The sample results are rejected.

Two facts should be noted by all data users. First, the "R" flag means that the associated value is unusable. In other words, due to significant quality control (QC) problems, the analysis is invalid and provides no information as to whether the compound is present or not. "R" values should not appear on data tables because they cannot be relied upon, even as a last resort. The second fact to keep in mind is that no compound concentration, even if it has passed all QC tests, is guaranteed to be accurate. Strict QC serves to increase confidence in data but any value potentially contains error.

## DATA REVIEW

### VOLATILE ORGANIC COMPOUND (VOC) ANALYSES

#### 1. Holding Times

The specified holding times for the following methods are presented in the following table.

Method	Matrix	Holding Time	Preservation
SW-846 8260D/8260D-SIM	Water	14 days from collection to analysis	Cool to < 6 °C; pH < 2 with HCl

All samples were analyzed within the specified holding time criteria.

#### 2. Mass Spectrometer Tuning

Mass spectrometer performance was acceptable and all analyses were performed within a 12-hour tune clock.

System performance and column resolution were acceptable.

#### 3. Calibration

Satisfactory instrument calibration is established to ensure that the instrument is capable of producing acceptable quantitative data. An initial calibration demonstrates that the instrument is capable of acceptable performance at the beginning of an experimental sequence. The continuing calibration verifies that the instrument daily performance is satisfactory.

##### 3.1 Initial Calibration

The method specifies percent relative standard deviation (%RSD) and relative response factor (RRF) limits for select compounds only. A technical review of the data applies limits to all compounds with no exceptions.

All target compounds associated with the initial calibration standards must exhibit a %RSD less than the control limit (20%) or a correlation coefficient greater than 0.99 and an RRF value greater than control limit (0.05).

All compounds associated with the initial calibrations were within the specified control limits.

##### 3.2 Continuing Calibration

All target compounds associated with the continuing calibration standard must exhibit a percent difference (%D) less than the control limit (20%) and RRF value greater than control limit (0.05).

All compounds associated with the calibrations were within the specified control limits, with the exception of the compounds presented in the following table.

Sample ID	Initial / Continuing	Compound	Criteria
TRIP BLANK_136 MW-79D_051123 DUP-12 MW-79SR_051123 MW-141S_051123	Initial Calibration Verification %D	Vinyl chloride	+25.9%
MW-79D_051123 DUP-12 MW-79SR_051123 MW-141S_051123	Initial Calibration Verification %D	1,4-Dioxane	+28.1%

The criteria used to evaluate the initial and continuing calibration are presented in the following table. In the case of a calibration deviation, the sample results are qualified.

Initial/Continuing	Criteria	Sample Result	Qualification
	RRF <0.05	Non-detect	R



## DATA REVIEW

Initial/Continuing	Criteria	Sample Result	Qualification
Initial and Continuing Calibration	RRF <0.01 <sup>1</sup>	Detect	J
		Non-detect	R
	Detect	J	
	RRF >0.05 or RRF >0.01 <sup>1</sup>	Non-detect	No Action
		Detect	
Initial Calibration	%RSD > 20% or a correlation coefficient <0.99	Non-detect	UJ
		Detect	J
	%RSD > 90%	Non-detect	R
		Detect	J
Continuing Calibration	%D >20% (increase in sensitivity)	Non-detect	UJ
		Detect	J
	%D >20% (decrease in sensitivity)	Non-detect	UJ
		Detect	J
	%D > 90% (increase/decrease in sensitivity)	Non-detect	R
		Detect	J

Note:

<sup>1</sup>RRF of 0.01 only applies to compounds which are typically poor responding compounds

### 4. Internal Standard Performance

Internal standard performance criteria ensure that the GC/MS sensitivity and response are stable during every sample analysis. The criteria require the internal standard compounds associated with the VOC exhibit area counts that are not greater than two times (+100%) or less than one-half (-50%) of the area counts of the associated continuing calibration standard.

All internal standard responses were within control limits.

### 5. Field Duplicate Analysis

Field duplicate analysis is used to assess the overall precision of the field sampling procedures and analytical method. A control limit of 30% for water matrices is applied to the RPD between the parent sample and the field duplicate. In the instance when the parent and/or duplicate sample concentrations are less than or equal to 5 times the RL, a control limit of two times the RL is applied for water matrices.

Results for duplicate samples are summarized in the following table.

Sample ID/Duplicate ID	Compound	Sample Result	Duplicate Result	RPD
MW-79D_051123/DUP-12	Vinyl chloride	1.8	2.2	AC

Notes:

AC - Acceptable

The calculated RPDs between the parent sample and field duplicate were acceptable.

### 6. Compound Identification

Compounds are identified on the GC/MS by using the analytes relative retention time and ion spectra.

All identified compounds met the specified criteria.

## **DATA REVIEW**

### **7. System Performance and Overall Assessment**

Overall system performance was acceptable. Other than for those deviations specifically mentioned in this review, the overall data quality is within the guidelines specified in the method.

## DATA REVIEW

### DATA VALIDATION CHECKLIST FOR VOCs

VOCs: 8260D/8260D-SIM	Reported		Performance Acceptable		Not Required
	No	Yes	No	Yes	
<b>GAS CHROMATOGRAPHY/MASS SPECTROMETRY (GC/MS)</b>					
<b>Tier II Validation</b>					
Holding times/Preservation		X		X	
<b>Tier III Validation</b>					
System performance and column resolution		X		X	
Initial calibration %RSDs		X		X	
Continuing calibration RRFs		X		X	
Continuing calibration %Ds		X	X		
Instrument tune and performance check		X		X	
Ion abundance criteria for each instrument used		X		X	
Field Duplicate RPD		X		X	
Internal standard		X		X	
Compound identification and quantitation					
A. Reconstructed ion chromatograms		X		X	
B. Quantitation Reports		X		X	
C. RT of sample compounds within the established RT windows		X		X	
D. Transcription/calculation errors present		X		X	
E. Reporting limits adjusted to reflect sample dilutions		X		X	

Notes:

%RSD Relative standard deviation

%R Percent recovery

RPD Relative percent difference

%D Percent difference

## DATA REVIEW

VALIDATION PERFORMED BY: Dilip Kumar

SIGNATURE: 

---

DATE: June 16, 2023

---

PEER REVIEW: Andrew Korycinski

DATE: June 19, 2023

---

# **NO CORRECTIONS/QUALIFIERS ADDED TO SAMPLE ANALYSIS DATA SHEETS**



**CHAIN OF CUSTODY  
CORRECTED SAMPLE ANALYSIS DATA  
SHEETS**



# Chain of Custody Record

# MICHIGAN

# TestAmerica

TestAmerica Laboratory location: Brighton --- 10448 Citation Drive, Suite 200 / Brighton, MI 48116 / 810-229-2763

190

THE LEADER IN ENVIRONMENTAL TESTING

<b>Client Contact</b>		<b>Regulatory program:</b> <input type="checkbox"/> DW <input type="checkbox"/> NPDES <input type="checkbox"/> RCRA <input type="checkbox"/> Other				TestAmerica Laboratories, Inc.	
Company Name: Arcadis		Client Project Manager: Kris Hinskey		Site Contact: Christina Weaver		Lab Contact: Mike DelMonico	
Address: 28550 Cabot Drive, Suite 500		Telephone: 248-994-2240		Telephone: 248-994-2240		Telephone: 330-497-9396	
City/State/Zip: Novi, MI, 48377		Email: kristoffer.hinskey@arcadis.com		<b>Analysis Turnaround Time</b>		1 of 1 COCs	
Phone: 248-994-2240		Sampler Name: <i>Kent Kurper</i>		TAT if different from below		For lab use only	
Project Name: Ford LTP Off-Site		Method of Shipment/Carrier:		<input type="checkbox"/> 3 weeks <input checked="" type="checkbox"/> 2 weeks <input type="checkbox"/> 1 week <input type="checkbox"/> 2 days <input type="checkbox"/> 1 day			
Project Number: 30167538.402.04		Shipping/Tracking No:					
PO # 30167538.402.04						Walk-in client Lab sampling Job/SDG No:	

Sample Identification	Sample Date	Sample Time	Matrix					Containers & Preservatives								Filtered Sample (Y/N)	Composite=C / Grab=G	Analyses										Sample Specific Notes / Special Instructions								
			Air	Aqueous	Sediment	Solid	Other:	H2SO4	HNO3	HCl	NaOH	ZnAc	NaOH	Uppres	Other:			1,1-DCE 8260B	cis-1,2-DCE 8260B	Trans-1,2-DCE 8260B	PCE 8260B	TCE 8260B	Vinyl Chloride 8260B	1,4-Dioxane 8260B SIM												
TRIP BLANK_ 136	5/11/23	---	1							1																										1 Trip Blank
MW-79N-051123	5/11/23	1255	6							6																										3 VOAs for 8260B 3 VOAs for 8260B SIM
DUP-12	5/11/23	---	6							6																										
MW-79SR-051123	5/11/23	1420	6							6																										
MW-141S-051123	5/11/23	11:25	6							6																										



240-185407 Chain of Custody

<b>Possible Hazard Identification</b>				<b>Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</b>			
<input checked="" type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown				<input type="checkbox"/> Return to Client <input checked="" type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For <input type="checkbox"/> Months			

**Special Instructions/QC Requirements & Comments:**  
 Sample Address: BOSTON POST ROW  
 Submit all results through Cadena at jtomalia@cadenaco.com. Cadena #E203631  
 Level IV Reporting requested.

Relinquished by: <i>Kent Kurper</i>	Company: <i>Arcadis</i>	Date/Time: <i>5/11/23 1534</i>	Received by: <i>Novi Cold Storage</i>	Company: <i>Arcadis</i>	Date/Time: <i>5/11/23 1534</i>
Relinquished by: <i>Christina Weaver</i>	Company: <i>ARCADIS</i>	Date/Time: <i>5/15/23 / 0825</i>	Received by: <i>[Signature]</i>	Company: <i>[Signature]</i>	Date/Time: <i>5/15/23 / 0825</i>
Relinquished by: <i>[Signature]</i>	Company: <i>ETA</i>	Date/Time: <i>5/16/23 0900</i>	Received in Laboratory by: <i>[Signature]</i>	Company: <i>ETINC</i>	Date/Time: <i>5/16/23 945</i>

Page 20 of 24

5/24/2023

©2008 TestAmerica Laboratories, Inc. All rights reserved. TestAmerica & Design™ are trademarks of TestAmerica Laboratories, Inc.



**Eurofins Cleveland**

180 S. Van Buren Avenue  
 Barberton OH 44203  
 Phone: 330-497-9396 Fax: 330-497-0772

**Chain of Custody Record**



Environment Test'ng

<b>Client Information (Sub Contract Lab)</b>		Sampler: DeIMonico, Michael		Lab PM: DeIMonico, Michael		Carrier Tracking No(s):		COC No: 240-168174.1																																																	
Client Contact: Shipping/Receiving		Phone:		E-Mail: Michael.DeIMonico@et.eurofinsus.com		State of Origin: Michigan		Page: Page 1 of 1																																																	
Company: Eurofins Environment Testing Northeast		Address: 777 New Durham Road		Due Date Requested: 5/29/2023		Accreditations Required (See note):		Job #: 240-185407-1																																																	
City: Edison		State, Zip: NJ, 08817		TAT Requested (days):		<table border="1"> <tr> <th colspan="12">Analysis Requested</th> </tr> <tr> <td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td> </tr> <tr> <td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td> </tr> <tr> <td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td> </tr> </table>		Analysis Requested																																																<b>Preservation Codes:</b> A HCL M Hexane B NaOH N None C Zn Acetate O AsNaO2 D Nitric Acid P Na2O4S E NaHSO4 Q Na2SO3 F MeOH R Na2S2O3 G Amchlor S H2SO4 H Ascorbic Acid T TSP Dodecahydrate I Ice U Acetone J DI Water V MCAA K EDTA W pH 4-5 L EDA Y Trizma Z other (specify)	
Analysis Requested																																																									
Project Name: Ford LTP Off Site		Project #: 24015353		SSOW#:		Field Filtered Sample (Yes or No)		Perform: MS/MSD (Yes or No)		8260D/6030C (MOD) VOCs (Short List)		8260D_SIM/6030C		Total Number of Containers																																											
Site:		Sample Date		Sample Time		Sample Type (C=comp, G=grab)		Matrix (W=water, S=solid, O=water/oil, BT=Tissue, A=Air)		Preservation Code:		Special Instructions/Note																																													
Sample Identification - Client ID (Lab ID)		TRIP BLANK_136 (240-185407-1)		5/11/23 Eastern		Water				X		1																																													
MW-79D_051123 (240-185407-2)		5/11/23 Eastern		12:55 Eastern		Water				X X		6																																													
DUP-12 (240-185407-3)		5/11/23 Eastern		Eastern		Water				X X		6																																													
MW-79SR_051123 (240-185407-4)		5/11/23 Eastern		14:20 Eastern		Water				X X		6																																													
MW-141S_051123 (240-185407-5)		5/11/23 Eastern		11:25 Eastern		Water				X X		6																																													

Page 23 of 24

5/24/2023

Note: Since laboratory accreditations are subject to change, Eurofins Environment Testing North Central, LLC places the ownership of method, analyte & accreditation compliance upon our subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/tests/matrix being analyzed, the samples must be shipped back to the Eurofins Environment Testing North Central, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Environment Testing North Central, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Environment Testing North Central, LLC.

<b>Possible Hazard Identification</b>		<b>Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</b>	
Unconfirmed		<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months	
Deliverable Requested: I, II, III, IV Other (specify)		Special Instructions/QC Requirements.	

Empty Kit Relinquished by: <i>Rachelle</i>		Date: 5/23/23		Time: 12:40		Method of Shipment: FedEx	
Relinquished by: <i>Rachelle</i>		Date/Time: 5/23/23 12:40		Company: EETNC		Received by: <i>Bhavisha</i>	
Relinquished by:		Date/Time:		Company:		Received by:	
Relinquished by:		Date/Time:		Company:		Received by:	

Custody Seals Intact: <input type="checkbox"/> Yes <input type="checkbox"/> No	Custody Seal No.	Cooler Temperature(s) °C and Other Remarks: <i>TRIP LIT ILLIC</i>
--	------------------	---





# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: TRIP BLANK\_136**

**Lab Sample ID: 240-185407-1**

Date Collected: 05/11/23 00:00

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/23/23 21:43	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/23/23 21:43	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/23/23 21:43	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/23/23 21:43	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/23/23 21:43	1
Vinyl chloride	1.0	U JJ	1.0	0.45	ug/L			05/23/23 21:43	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	87		70 - 128		05/23/23 21:43	1
Dibromofluoromethane (Surr)	94		77 - 124		05/23/23 21:43	1
Toluene-d8 (Surr)	86		80 - 120		05/23/23 21:43	1
4-Bromofluorobenzene	87		76 - 120		05/23/23 21:43	1

# Client Sample Results

Client: ARCADIS US Inc  
Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: MW-79D\_051123**

**Lab Sample ID: 240-185407-2**

Date Collected: 05/11/23 12:55

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	UJ	2.0	0.86	ug/L			05/21/23 04:13	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	99		75 - 133					05/21/23 04:13	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 00:05	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 00:05	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:05	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 00:05	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:05	1
<b>Vinyl chloride</b>	<b>1.8</b>	<b>J</b>	1.0	0.45	ug/L			05/24/23 00:05	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		70 - 128					05/24/23 00:05	1
Dibromofluoromethane (Surr)	97		77 - 124					05/24/23 00:05	1
Toluene-d8 (Surr)	85		80 - 120					05/24/23 00:05	1
4-Bromofluorobenzene	86		76 - 120					05/24/23 00:05	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: DUP-12**

**Lab Sample ID: 240-185407-3**

Date Collected: 05/11/23 00:00

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	UJ	2.0	0.86	ug/L			05/21/23 04:35	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	100		75 - 133					05/21/23 04:35	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 00:25	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 00:25	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:25	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 00:25	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:25	1
<b>Vinyl chloride</b>	<b>2.2</b>	<b>J</b>	1.0	0.45	ug/L			05/24/23 00:25	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	92		70 - 128					05/24/23 00:25	1
Dibromofluoromethane (Surr)	99		77 - 124					05/24/23 00:25	1
Toluene-d8 (Surr)	85		80 - 120					05/24/23 00:25	1
4-Bromofluorobenzene	84		76 - 120					05/24/23 00:25	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: MW-79SR\_051123**

**Lab Sample ID: 240-185407-4**

Date Collected: 05/11/23 14:20

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.6	J	2.0	0.86	ug/L			05/21/23 04:56	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	100		75 - 133					05/21/23 04:56	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 00:46	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 00:46	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:46	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 00:46	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 00:46	1
Vinyl chloride	1.0	UUJ	1.0	0.45	ug/L			05/24/23 00:46	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	91		70 - 128					05/24/23 00:46	1
Dibromofluoromethane (Surr)	99		77 - 124					05/24/23 00:46	1
Toluene-d8 (Surr)	84		80 - 120					05/24/23 00:46	1
4-Bromofluorobenzene	85		76 - 120					05/24/23 00:46	1

# Client Sample Results

Client: ARCADIS US Inc  
 Project/Site: Ford LTP - Off Site

Job ID: 240-185407-1

**Client Sample ID: MW-141S\_051123**

**Lab Sample ID: 240-185407-5**

Date Collected: 05/11/23 11:25

Matrix: Water

Date Received: 05/16/23 09:45

**Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	J	2.0	0.86	ug/L			05/21/23 05:18	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene	98		75 - 133					05/21/23 05:18	1

**Method: SW846 8260D - Volatile Organic Compounds by GC/MS**

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			05/24/23 01:06	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			05/24/23 01:06	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 01:06	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			05/24/23 01:06	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			05/24/23 01:06	1
Vinyl chloride	1.0	UU	1.0	0.45	ug/L			05/24/23 01:06	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	90		70 - 128					05/24/23 01:06	1
Dibromofluoromethane (Surr)	99		77 - 124					05/24/23 01:06	1
Toluene-d8 (Surr)	83		80 - 120					05/24/23 01:06	1
4-Bromofluorobenzene	86		76 - 120					05/24/23 01:06	1