

ANALYTICAL REPORT

PREPARED FOR

Attn: Ms. Megan Meckley
Arcadis US Inc.
28550 Cabot Drive
Suite 500
Novi, Michigan 48377

Generated 11/25/2025 9:06:56 PM

JOB DESCRIPTION

Ford LTP

JOB NUMBER

240-238061-1

Eurofins Cleveland

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Environment Testing North Central, LLC Project Manager.

Authorization



Generated
11/25/2025 9:06:56 PM

Authorized for release by
Michael DeMonico, Project Manager I
Michael.DeMonico@et.eurofinsus.com
(330)966-9783



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Method Summary	6
Sample Summary	7
Detection Summary	8
Client Sample Results	9
Surrogate Summary	11
QC Sample Results	12
QC Association Summary	16
Lab Chronicle	17
Certification Summary	18
Chain of Custody	19

Definitions/Glossary

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Qualifiers

GC/MS VOA

Qualifier	Qualifier Description
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: Arcadis US Inc.
Project: Ford LTP

Job ID: 240-238061-1

Job ID: 240-238061-1

Eurofins Cleveland

Job Narrative 240-238061-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 11/19/2025 8:00 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.8°C.

GC/MS VOA

Method 8260D: Sample (240-237730-A-4), (240-237730-A-4 MS) and (240-237730-A-4 MSD) had to be re-run multiple times because of (sample matrix problems / instrument problems), and this consumed a large amount of sample. Therefore, it was necessary to use the vial opened for preservation confirmation for the final analysis

Method 8260D: The method requirement for no headspace was not met. The container (s) used for reanalysis of the following samples contained headspace: (240-237730-A-4), (240-237730-A-4 MS) and (240-237730-A-4 MSD). The sample container was received with headspace.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Cleveland

Method Summary

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Method	Method Description	Protocol	Laboratory
8260D	Volatile Organic Compounds by GC/MS	SW846	EET CLE
8260D SIM	Volatile Organic Compounds (GC/MS)	SW846	EET CLE
5030C	Purge and Trap	SW846	EET CLE

Protocol References:

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

EET CLE = Eurofins Cleveland, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Sample Summary

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
240-238061-1	TRIP BLANK_83	Water	11/14/25 00:00	11/19/25 08:00	Michigan
240-238061-2	MW-211S_111425	Water	11/14/25 11:10	11/19/25 08:00	Michigan

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

Detection Summary

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Client Sample ID: TRIP BLANK_83

Lab Sample ID: 240-238061-1

No Detections.

Client Sample ID: MW-211S_111425

Lab Sample ID: 240-238061-2

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14

This Detection Summary does not include radiochemical test results.

Eurofins Cleveland

Client Sample Results

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Client Sample ID: TRIP BLANK_83

Lab Sample ID: 240-238061-1

Date Collected: 11/14/25 00:00

Matrix: Water

Date Received: 11/19/25 08:00

Method: SW846 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			11/21/25 13:35	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			11/21/25 13:35	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			11/21/25 13:35	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			11/21/25 13:35	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			11/21/25 13:35	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			11/21/25 13:35	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	124		62 - 137		11/21/25 13:35	1
4-Bromofluorobenzene (Surr)	87		56 - 136		11/21/25 13:35	1
Toluene-d8 (Surr)	100		78 - 122		11/21/25 13:35	1
Dibromofluoromethane (Surr)	117		73 - 120		11/21/25 13:35	1

Client Sample Results

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Client Sample ID: MW-211S_111425

Lab Sample ID: 240-238061-2

Date Collected: 11/14/25 11:10

Matrix: Water

Date Received: 11/19/25 08:00

Method: SW846 8260D SIM - Volatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			11/24/25 14:45	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	112		64 - 136					11/24/25 14:45	1

Method: SW846 8260D - Volatile Organic Compounds by GC/MS

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			11/22/25 12:18	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			11/22/25 12:18	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			11/22/25 12:18	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			11/22/25 12:18	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			11/22/25 12:18	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			11/22/25 12:18	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	101		62 - 137					11/22/25 12:18	1
4-Bromofluorobenzene (Surr)	104		56 - 136					11/22/25 12:18	1
Toluene-d8 (Surr)	101		78 - 122					11/22/25 12:18	1
Dibromofluoromethane (Surr)	104		73 - 120					11/22/25 12:18	1

Surrogate Summary

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Method: 8260D - Volatile Organic Compounds by GC/MS

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA	BFB	TOL	DBFM
		(62-137)	(56-136)	(78-122)	(73-120)
240-237730-A-4 MS	Matrix Spike	97	108	105	102
240-237730-A-4 MSD	Matrix Spike Duplicate	97	108	103	103
240-237996-A-5 MS	Matrix Spike	110	102	101	105
240-237996-A-5 MSD	Matrix Spike Duplicate	112	102	102	105
240-238061-1	TRIP BLANK_83	124	87	100	117
240-238061-2	MW-211S_111425	101	104	101	104
LCS 240-681556/5	Lab Control Sample	109	104	106	107
LCS 240-681710/5	Lab Control Sample	93	104	102	102
MB 240-681556/9	Method Blank	116	85	96	114
MB 240-681710/10	Method Blank	102	104	102	106

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)
BFB = 4-Bromofluorobenzene (Surr)
TOL = Toluene-d8 (Surr)
DBFM = Dibromofluoromethane (Surr)

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	DCA
		(64-136)
240-238061-2	MW-211S_111425	112
240-238067-E-4 MS	Matrix Spike	110
240-238067-E-4 MSD	Matrix Spike Duplicate	109
LCS 240-681852/5	Lab Control Sample	96
MB 240-681852/7	Method Blank	96

Surrogate Legend

DCA = 1,2-Dichloroethane-d4 (Surr)

QC Sample Results

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Method: 8260D - Volatile Organic Compounds by GC/MS

Lab Sample ID: MB 240-681556/9

Matrix: Water

Analysis Batch: 681556

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			11/21/25 12:48	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			11/21/25 12:48	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			11/21/25 12:48	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			11/21/25 12:48	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			11/21/25 12:48	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			11/21/25 12:48	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
1,2-Dichloroethane-d4 (Surr)	116		62 - 137		11/21/25 12:48	1
4-Bromofluorobenzene (Surr)	85		56 - 136		11/21/25 12:48	1
Toluene-d8 (Surr)	96		78 - 122		11/21/25 12:48	1
Dibromofluoromethane (Surr)	114		73 - 120		11/21/25 12:48	1

Lab Sample ID: LCS 240-681556/5

Matrix: Water

Analysis Batch: 681556

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
1,1-Dichloroethene	25.0	24.8		ug/L		99	63 - 134
cis-1,2-Dichloroethene	25.0	24.1		ug/L		96	77 - 123
Tetrachloroethene	25.0	24.7		ug/L		99	76 - 123
trans-1,2-Dichloroethene	25.0	24.7		ug/L		99	75 - 124
Trichloroethene	25.0	24.3		ug/L		97	70 - 122
Vinyl chloride	25.0	25.1		ug/L		101	60 - 144

Surrogate	LCS	LCS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	109		62 - 137
4-Bromofluorobenzene (Surr)	104		56 - 136
Toluene-d8 (Surr)	106		78 - 122
Dibromofluoromethane (Surr)	107		73 - 120

Lab Sample ID: 240-237996-A-5 MS

Matrix: Water

Analysis Batch: 681556

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Analyte	Sample	Sample	Spike Added	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier		Result	Qualifier				
Tetrachloroethene	1.0	U	25.0	19.5		ug/L		78	62 - 131

Surrogate	MS	MS	Limits
	%Recovery	Qualifier	
1,2-Dichloroethane-d4 (Surr)	110		62 - 137
4-Bromofluorobenzene (Surr)	102		56 - 136
Toluene-d8 (Surr)	101		78 - 122
Dibromofluoromethane (Surr)	105		73 - 120

QC Sample Results

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Method: 8260D - Volatile Organic Compounds by GC/MS (Continued)

Lab Sample ID: 240-237996-A-5 MSD

Matrix: Water

Analysis Batch: 681556

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Tetrachloroethene	1.0	U	25.0	20.2		ug/L		81	62 - 131	4	20
Surrogate	%Recovery	Qualifier	Limits								
1,2-Dichloroethane-d4 (Surr)	112		62 - 137								
4-Bromofluorobenzene (Surr)	102		56 - 136								
Toluene-d8 (Surr)	102		78 - 122								
Dibromofluoromethane (Surr)	105		73 - 120								

Lab Sample ID: MB 240-681710/10

Matrix: Water

Analysis Batch: 681710

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
1,1-Dichloroethene	1.0	U	1.0	0.49	ug/L			11/22/25 11:54	1
cis-1,2-Dichloroethene	1.0	U	1.0	0.46	ug/L			11/22/25 11:54	1
Tetrachloroethene	1.0	U	1.0	0.44	ug/L			11/22/25 11:54	1
trans-1,2-Dichloroethene	1.0	U	1.0	0.51	ug/L			11/22/25 11:54	1
Trichloroethene	1.0	U	1.0	0.44	ug/L			11/22/25 11:54	1
Vinyl chloride	1.0	U	1.0	0.45	ug/L			11/22/25 11:54	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
1,2-Dichloroethane-d4 (Surr)	102		62 - 137					11/22/25 11:54	1
4-Bromofluorobenzene (Surr)	104		56 - 136					11/22/25 11:54	1
Toluene-d8 (Surr)	102		78 - 122					11/22/25 11:54	1
Dibromofluoromethane (Surr)	106		73 - 120					11/22/25 11:54	1

Lab Sample ID: LCS 240-681710/5

Matrix: Water

Analysis Batch: 681710

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1,1-Dichloroethene	25.0	27.8		ug/L		111	63 - 134
cis-1,2-Dichloroethene	25.0	25.4		ug/L		102	77 - 123
Tetrachloroethene	25.0	23.6		ug/L		94	76 - 123
trans-1,2-Dichloroethene	25.0	26.5		ug/L		106	75 - 124
Trichloroethene	25.0	24.7		ug/L		99	70 - 122
Vinyl chloride	25.0	26.3		ug/L		105	60 - 144
Surrogate	%Recovery	Qualifier	Limits				
1,2-Dichloroethane-d4 (Surr)	93		62 - 137				
4-Bromofluorobenzene (Surr)	104		56 - 136				
Toluene-d8 (Surr)	102		78 - 122				
Dibromofluoromethane (Surr)	102		73 - 120				

QC Sample Results

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Method: 8260D - Volatile Organic Compounds by GC/MS (Continued)

Lab Sample ID: 240-237730-A-4 MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 681710

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec	Limits	
	Result	Qualifier	Added	Result	Qualifier						
1,1-Dichloroethene	68		125	204		ug/L		109		56 - 135	
cis-1,2-Dichloroethene	5.0	U	125	129		ug/L		103		66 - 128	
Tetrachloroethene	10		125	130		ug/L		96		62 - 131	
trans-1,2-Dichloroethene	5.0	U	125	133		ug/L		107		56 - 136	
Trichloroethene	6.3		125	129		ug/L		98		61 - 124	
Vinyl chloride	5.0	U	125	138		ug/L		110		43 - 157	
MS MS											
Surrogate	%Recovery	Qualifier	Limits								
1,2-Dichloroethane-d4 (Surr)	97		62 - 137								
4-Bromofluorobenzene (Surr)	108		56 - 136								
Toluene-d8 (Surr)	105		78 - 122								
Dibromofluoromethane (Surr)	102		73 - 120								

Lab Sample ID: 240-237730-A-4 MSD

Client Sample ID: Matrix Spike Duplicate

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 681710

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	Limits	RPD	RPD
	Result	Qualifier	Added	Result	Qualifier						RPD	Limit
1,1-Dichloroethene	68		125	198		ug/L		104		56 - 135	3	26
cis-1,2-Dichloroethene	5.0	U	125	128		ug/L		103		66 - 128	0	14
Tetrachloroethene	10		125	121		ug/L		88		62 - 131	8	20
trans-1,2-Dichloroethene	5.0	U	125	131		ug/L		105		56 - 136	2	15
Trichloroethene	6.3		125	124		ug/L		94		61 - 124	4	15
Vinyl chloride	5.0	U	125	138		ug/L		110		43 - 157	0	24
MSD MSD												
Surrogate	%Recovery	Qualifier	Limits									
1,2-Dichloroethane-d4 (Surr)	97		62 - 137									
4-Bromofluorobenzene (Surr)	108		56 - 136									
Toluene-d8 (Surr)	103		78 - 122									
Dibromofluoromethane (Surr)	103		73 - 120									

Method: 8260D SIM - Volatile Organic Compounds (GC/MS)

Lab Sample ID: MB 240-681852/7

Client Sample ID: Method Blank

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 681852

Analyte	MB	MB	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier							
1,4-Dioxane	2.0	U	2.0	0.86	ug/L			11/24/25 12:00	1
MB MB									
Surrogate	%Recovery	Qualifier	Limits						
1,2-Dichloroethane-d4 (Surr)	96		64 - 136						

QC Sample Results

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Method: 8260D SIM - Volatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 240-681852/5

Matrix: Water

Analysis Batch: 681852

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1,4-Dioxane	10.0	8.30		ug/L		83	68 - 120
Surrogate	%Recovery	LCS Qualifier	Limits				
1,2-Dichloroethane-d4 (Surr)	96		64 - 136				

Lab Sample ID: 240-238067-E-4 MS

Matrix: Water

Analysis Batch: 681852

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1,4-Dioxane	2.0	U	10.0	8.51		ug/L		85	45 - 145
Surrogate	%Recovery	MS Qualifier	Limits						
1,2-Dichloroethane-d4 (Surr)	110		64 - 136						

Lab Sample ID: 240-238067-E-4 MSD

Matrix: Water

Analysis Batch: 681852

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1,4-Dioxane	2.0	U	10.0	8.67		ug/L		87	45 - 145	2	19
Surrogate	%Recovery	MSD Qualifier	Limits								
1,2-Dichloroethane-d4 (Surr)	109		64 - 136								

QC Association Summary

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

GC/MS VOA

Analysis Batch: 681556

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-238061-1	TRIP BLANK_83	Total/NA	Water	8260D	
MB 240-681556/9	Method Blank	Total/NA	Water	8260D	
LCS 240-681556/5	Lab Control Sample	Total/NA	Water	8260D	
240-237996-A-5 MS	Matrix Spike	Total/NA	Water	8260D	
240-237996-A-5 MSD	Matrix Spike Duplicate	Total/NA	Water	8260D	

Analysis Batch: 681710

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-238061-2	MW-211S_111425	Total/NA	Water	8260D	
MB 240-681710/10	Method Blank	Total/NA	Water	8260D	
LCS 240-681710/5	Lab Control Sample	Total/NA	Water	8260D	
240-237730-A-4 MS	Matrix Spike	Total/NA	Water	8260D	
240-237730-A-4 MSD	Matrix Spike Duplicate	Total/NA	Water	8260D	

Analysis Batch: 681852

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
240-238061-2	MW-211S_111425	Total/NA	Water	8260D SIM	
MB 240-681852/7	Method Blank	Total/NA	Water	8260D SIM	
LCS 240-681852/5	Lab Control Sample	Total/NA	Water	8260D SIM	
240-238067-E-4 MS	Matrix Spike	Total/NA	Water	8260D SIM	
240-238067-E-4 MSD	Matrix Spike Duplicate	Total/NA	Water	8260D SIM	

Lab Chronicle

Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Client Sample ID: TRIP BLANK_83

Lab Sample ID: 240-238061-1

Date Collected: 11/14/25 00:00

Matrix: Water

Date Received: 11/19/25 08:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	681556	MS	EET CLE	11/21/25 13:35

Client Sample ID: MW-211S_111425

Lab Sample ID: 240-238061-2

Date Collected: 11/14/25 11:10

Matrix: Water

Date Received: 11/19/25 08:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8260D		1	681710	MS	EET CLE	11/22/25 12:18
Total/NA	Analysis	8260D SIM		1	681852	R5XG	EET CLE	11/24/25 14:45

Laboratory References:

EET CLE = Eurofins Cleveland, 180 S. Van Buren Avenue, Barberton, OH 44203, TEL (330)497-9396

Accreditation/Certification Summary


Client: Arcadis US Inc.
Project/Site: Ford LTP

Job ID: 240-238061-1

Laboratory: Eurofins Cleveland

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Connecticut	State	PH-0806	09-30-26
Georgia	State	4062	02-27-26
Illinois	NELAP	200004	08-31-26
Iowa	State	421	06-01-27
Kansas	NELAP	E-10336	01-31-26
Kentucky (UST)	State	112225	02-28-26
Kentucky (WW)	State	KY98016	12-31-25
Minnesota	NELAP	039-999-348	12-31-25
New Hampshire	NELAP	2250	09-30-26
New Jersey	NELAP	OH001	06-30-26
New York	NELAP	10975	04-01-26
North Dakota	State	R-244	02-27-26
Ohio	State	8303	02-27-26
Ohio VAP	State	ORELAP 4062	02-28-26
Oregon	NELAP	4062	02-27-26
Pennsylvania	NELAP	68-00340	08-31-26
Texas	NELAP	T104704517	08-31-26
US Fish & Wildlife	US Federal Programs	A26406	02-28-26
USDA	US Federal Programs	P330-18-00281	01-05-27
Virginia	NELAP	460175	09-30-26
West Virginia DEP	State	210	12-31-25
Wisconsin	State	399167560	08-31-26



Eurofins - Cleveland Sample Receipt Form/Narrative Login # _____
 Barberton Facility

Client Aradis Site Name _____ Cooler unpacked by: W Martin

Cooler Received on 1119/25 Opened on 1119/25

FedEx, 1st Grd Exp UPS FAS W Spot Client Drop Off Eurofins Courier Other _____

Receipt After-hours, Drop-off Date/Time _____ Storage Location _____

Eurofins Cooler # EC Foam Box _____ Client Cooler _____ Box _____ Other _____
 Packing material used. Bubble Wrap Foam Plastic Bag None Other _____
 COOLANT: W Ice Blue Ice Dry Ice Water None _____
 1 Cooler temperature upon receipt See Multiple Cooler Form

IR GUN # 13 (CF 10.6 °C) Observed Cooler Temp. 0.2 °C Corrected Cooler Temp 0.8 °C

2. Were tamper/custody seals on the outside of the cooler(s)? If Yes Quantity _____
- Were the seals on the outside of the cooler(s) signed & dated? Yes No NA
 - Were tamper/custody seals on the bottle(s) or bottle kits (LLHg/MeHg)? Yes No NA
 - Were tamper/custody seals intact and uncompromised? Yes No NA
- 3 Shippers' packing slip attached to the cooler(s)? Yes No
- 4 Did custody papers accompany the sample(s)? Yes No
- 5 Were the custody papers relinquished & signed in the appropriate place? Yes No
- 6 Was/were the person(s) who collected the samples clearly identified on the COC? Yes No
- 7 Did all bottles arrive in good condition (Unbroken)? Yes No
- 8 Could all bottle labels (ID/Date/Time) be reconciled with the COC? Yes No
- 9 For each sample, does the COC specify preservatives (Y/N), # of containers (Y/N), and sample type of grab/comp (Y/N)? Yes No
- 10 Were correct bottle(s) used for the test(s) indicated? Yes No
- 11 Sufficient quantity received to perform indicated analyses? Yes No
12. Are these work share samples and all listed on the COC? Yes No

Tests that are not checked for pH by Receiving:
 VOAs
 Oil and Grease
 TOC

- If yes, Questions 13-17 have been checked at the originating laboratory
- 13 Were all preserved sample(s) at the correct pH upon receipt? Yes No NA pH Strip Lot# HC567196
- 14 Were VOAs on the COC? Yes No
- 15 Were air bubbles >6 mm in any VOA vials? Larger than this Yes NA
- 16 Was a VOA trip blank present in the cooler(s)? Trip Blank Lot # NA Yes No
- 17 Was a LL Hg or Me Hg trip blank present? Yes No

Contacted PM _____ Date _____ by _____ via Verbal Voice Mail Other _____
 Concerning _____

18. CHAIN OF CUSTODY & SAMPLE DISCREPANCIES additional next page Labeled by: _____
Labels Verified by _____

19. SAMPLE CONDITION

Sample(s) _____ were received after the recommended holding time had expired

Sample(s) _____ were received in a broken container

Sample(s) _____ were received with bubble >6 mm in diameter (Notify PM)

20 SAMPLE PRESERVATION

Sample(s) _____ were further preserved in the laboratory

Time preserved. _____ Preservative(s) added/Lot number(s) _____

VOA Sample Preservation - Date/Time VOAs Frozen _____



11/19/2025

Login Container Summary Report

240-238061

Temperature readings

11/25/2025

<u>Client Sample ID</u>	<u>Lab ID</u>	<u>Container Type</u>	<u>Container Preservation</u>		
			<u>pH</u>	<u>Temp</u>	<u>Added Lot Number</u>
TRIP BLANK_83	240-238061-A-1	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____
MW-211S_111425	240-238061-A-2	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____
MW-211S_111425	240-238061-B-2	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____
MW-211S_111425	240-238061-C-2	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____
MW-211S_111425	240-238061-D-2	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____
MW-211S_111425	240-238061-E-2	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____
MW-211S_111425	240-238061-F-2	Voa Vial 40ml - Hydrochloric Acid	_____	_____	_____

DATA VERIFICATION REPORT



November 26, 2025

Megan Meckley
Arcadis
28550 Cabot Drive
Suite 500
Novi, MI US 48377

CADENA project ID: E203728
Project: Ford Livonia Transmission Plant - ON-SITE Soil Gas, Ground Water and Soil
Project number: 30251157.401.04 LTP
Event Specific Scope of Work References: Sample COC
Laboratory: Eurofins Environment Testing LLC - Cleveland
Laboratory submittal: 238061-1
Sample date: 2025-11-14
Report received by CADENA: 2025-11-26
Initial Data Verification completed by CADENA: 2025-11-26
Number of Samples:2
Sample Matrices:Water
Test Categories:GCMS VOC
Please see attached criteria report or sample result/qualified analytical result summary for qualifier flags assigned to sample data.

The following minor QC exceptions or missing information were noted:

GCMS VOC QC batch MS/MSD recovery outliers were not determined using a client sample so qualification was not required based on these sample-specific QC outliers.

Sample/MS/MSD Surrogate Recovery, Blank/LCS Surrogate Recovery, LCS/LCD Recovery, Blank Contamination and Hold Time Exception were reviewed as part of our verification.

Data verification for the report specified above was completed using the Ford Motor Company Environmental Laboratory Technical Specification, the CADENA Standard Operating Procedure for the Verification of Environmental Analytical Data and the associated analytical methods as references for evaluating the batch QC, sample data and report content. The EPA National Functional Guidelines for validating organic and inorganic data were used as guidance when addressing out of control QC results and the associated data qualifiers.

The definitions of the qualifiers used for this data package are defined in the analytical report. CADENA valid qualifiers are defined in the table below. To view and download a PDF copy of the laboratory analytical report access the CADENA CLMS at <http://clms.cadenaco.com/index.cfm>.

Please contact me if you have any questions.

Sincerely,

Jim Tomalia

Project Scientist

CADENA Inc, 1099 Highland Drive, Suite E, Ann Arbor, MI 48108 517-819-0356

CADENA Valid Qualifiers

Valid Qualifiers	Description
<	Less than the reported concentration.
>	Greater than the reported concentration.
B	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was greater than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the reported concentration. For Inorganic methods the sample concentration was greater than the RDL and less than 10x the blank concentration and is considered non-detect at the reported concentration.
E	The analyte / Compound reported exceeds the calibration range and is considered estimated.
EMPC	Estimated Minimum Potential Contamination - Dioxin/Furan analyses only.
J	Indicates an estimated value. This flag is used either when estimating a concentration for a tentatively identified compound or when the data indicates the presence of an analyte / compound but the result is less than the sample Quantitation limit, but greater than zero. The flag is also used in data validation to indicate a reported value should be considered estimated due to associated quality assurance deficiencies.
J-	The result is an estimated quantity, but the result may be biased low.
JB	NON-DETECT AT THE CONCENTRATION REPORTED AND ESTIMATED
JH	The sample result is considered estimated and is potentially biased high.
JL	The sample result is considered estimated and is potentially biased low.
JUB	NON-DETECT AT THE REPORTING LIMIT AND ESTIMATED
NJ	Tentatively identified compound with approximated concentration.
R	Indicates the value is considered to be unusable. (Note: The analyte / compound may or may not be present.)
TNTC	Too Numerous to Count - Asbestos and Microbiological Results.
U	Indicates that the analyte / compound was analyzed for, but not detected.
UB	The analyte / compound was detected in the associated blank. For Organic methods the sample concentration was less than the RDL and less than 5x (or 10x for common lab contaminants) the blank concentration and is considered non-detect at the RDL. For Inorganic methods the sample concentration was less than the RDL and less than 10x the blank concentration and is considered non-detect at the RDL.
UJ	The analyte / compound was not detected above the reported sample Quantitation limit. However, the Quantitation limit is considered to be approximate due to associated quality assurance results and may or may not represent the actual limit of Quantitation to accurately and precisely report the analyte in the sample.

Analytical Results Summary

CADENA Project ID: E203728

Laboratory: Eurofins Environment Testing LLC - Cleveland

Laboratory Submittal: 238061-1

Sample Name:	TRIP BLANK_83	MW-211S_111425
Lab Sample ID:	2402380611	2402380612
Sample Date:	11/14/2025	11/14/2025

Analyte	Cas No.	Report		Units	Valid Qualifier	Report		Units	Valid Qualifier
		Result	Limit			Result	Limit		
GC/MS VOC									
<u>OSW-8260D</u>									
1,1-Dichloroethene	75-35-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---
cis-1,2-Dichloroethene	156-59-2	ND	1.0	ug/l	---	ND	1.0	ug/l	---
Tetrachloroethene	127-18-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---
trans-1,2-Dichloroethene	156-60-5	ND	1.0	ug/l	---	ND	1.0	ug/l	---
Trichloroethene	79-01-6	ND	1.0	ug/l	---	ND	1.0	ug/l	---
Vinyl chloride	75-01-4	ND	1.0	ug/l	---	ND	1.0	ug/l	---
<u>OSW-8260DSIM</u>									
1,4-Dioxane	123-91-1					ND	2.0	ug/l	---